


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	THQH - Varkom
3	Name of Equipment with Make, Model and Serial Number	Oxygen Concentrator QJ-2410L2
4	Equipment ID & Barcode	0521303, 899444
5	Date of purchase / Year of manufacture / Installation Date	3/08/2021
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	NO
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	10-01-2024
9	Action taken	Serviced sieve bed, main board & compressors
10	Present status of the equipment (Fully damaged / partially damaged)	fully damaged
11	Recommendations for repair (required service details)	needed new sieve beds, compressors and main board.
12	Cost of spares (specify parts and cost)	spare parts not available

13	Asset Value	OEM not available
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	The spare parts and OEM supports are not available. According to close S.3.14.2, this machine recommended for BER.
17	Name & Signature of CYRIX Authority	Munish R. Naik 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - not mentioned  
 Cost of spares - not mentioned  
 Life period - 2 year 5 months

RBER criteria not satisfied as per the guidelines


CONSULTANT BIOMEDICAL ENGINEER  
 NATIONAL HEALTH MISSION  
 KOTTAYAM  
 8/2/24  
 Signature of JC BM (NHM)

PREETHEE S. RAJAN

Taluk Head Quarters Hospital  
 VAIKOM PIN - 686 141

Date

Seal

  
 Superintendent  
 Taluk Head Quarters Hospital  
 Vaikom Pin - 686 141  
 Superintendent / Medical Officer (I/c)








O2 101) RENTAL

	Rd	used	149
1/8/2021 Rd from store.	1		1
1/13/21 Rd from store	1		2
1/1/2022 - TRONK to Wd2 (P no 27)		1	1
1/5 from Wd 2	1		2
12/1/22 Rd from store	1		3
29/1/22 TRONK to CSKAC		1	2
Rd from CSKAC	1		3
		2- Wd 1- 3	2- Wd 1- 3
4/7/22 Rd from store	3		6
6/11/2022 to Wincable brook		1	5
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			
29/3/21 - Rd from CSKAC			



Oxygen concentrator

Model	O2-2V10Lx
Flow	10L/min
Purity	93% ± 3%
Pressure	0.01 - 0.1 MPa
Power	~100W
Weight	~10kg
Dimensions	~400x300x1000mm

QC PASSED



REDMI NOTE 11 PRO+ 5G  
TEAM EAGLE

18/01/2024 14:28



18/01/2024 14:27

REDMI NOTE 11 PRO+ 5G  
TEAMEAGLE



18/01/2024 14:27

REDMI NOTE 11 PRO+ 5G  
TEAMAGLE

**CYRIX** BIOLOGICAL EQUIPMENT MAINTENANCE PROGRAM (BEP)  
TOLL FREE NO: 1-800-425-7666  
BAR CODE - (8004-800615229)  
0621303



18/01/2024





REDMI NOTE 11 PRO+ 5G  
TEAMEAGLE

18/01/2024 14:28