

13	Asset Value	N/A.
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not available.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	Ups supplied along with Haematology analyzer. Machine aged upto 5 years. Main board not available. Economically not repairable. So recommending BER.
17	Name & Signature of CYRIX Authority	Nidhish Roy <u>NBS</u>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - Not mentioned
 Cost of spare - Not mentioned
 Life period - 5 years 5 months
 As per the BER guidelines cost of spare and percentage not available. Cyrix reported that spare not available.


Signature of JC, BM (NHM)

Machine not useful condition

Sudhanot. M.S.
 Lab Technician
 Sudhanot. M.S.

Date: 16/1/24

Signature of P.H.C. Mutholy
 Superintendent / Medical Officer (i/c)



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	PHC Mutholy
3	Name of Equipment with Make, Model and Serial Number	Ups of Haematology Analyzer Make: i'ball Model: Nisantar Ups 6211 Sn: 1800330000191
4	Equipment ID & Barcode	0546279
5	Date of purchase / Year of manufacture /Installation Date	11-09-2018
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	06-01-2024, 89316
9	Action taken	Serviced. Checked with new ups battery.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Machine not power ON Charging circuit of main board complaint. Output not getting during power line
12	Cost of spares (specify parts and cost)	Spares not available.



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 170741

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>Phc Mutholy</u>	Call Registration Date : <u>06-01-2024</u>
Address <u>Mutholy</u>	Caller ID : <u>89316</u>
<u>Kottayam</u>	Date of Visit : <u>06-01-2024</u>
Ph : <u>8078080462</u>	Asset No. : <u>0546079</u>
	EQPT Name : <u>UPS of Haematology Analyzer</u>
	Manufacture <u>Ball</u> Model : <u>Nixator</u>
	S. No. <u>180093000191</u> Dept. <u>Laboratory</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not power ON properly, Low battery back up, Battery charging circuit complaint, No output voltage

Action Taken : Checked. Machine checked with new battery. Battery didn't charging and output not get due to charger part of main board complaint. Board unrepairable. Required a new board. Board not available so that unserviceable.

Completed Date : 16-01-2024 Time : 11:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nidhish Roy</u>	<u>16-01-2024</u>	<u>10:45am</u>	<u>11:30am</u>

Customer Remark <u>Machine not in usable condition.</u>	Completed <input checked="" type="checkbox"/> Pending
Service Engineer Name : <u>Nidhish</u> Signature : <u>Nas</u> Date : <u>16-01-2024</u> Contact Number : <u>9070575296</u>	Customer Name : <u>Sudhanoj M S</u> Signature : <u>Sudhanoj M S</u> Date : <u>16/1/24</u> Contact Number : <u>8078080462</u> Designation : <u>Lab Technician</u> Hospital Seal :



Sudhanoj M S Medical Officer I/C
P.H.C. Mutholy

47

HAEMATOLOGY ANALYSER

Date	ITEM	Receipt	issue	Balance
09/9/18	<p>Received from 2500 Page of store Dated 11-09-2018 (RS 294000/+18% tax) Net Amt RS - 346920.00/- BLOOD MIXER - Received from store with Haematology Analyser. UPS - ALONG WITH HAEMATO LOGY ANALYSER.</p>	<p>IDI Registered 01</p>	<p>01</p>	<p>NIL</p>

Subtotal

