



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 173559

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC purathur</u></p> <p>Address <u>purathur</u> <u>malappuram</u></p> <p>Ph : <u>9847313900</u></p>	<p>Call Registration Date : <u>24-01-2024</u></p> <p>Caller ID : <u>92611</u></p> <p>Date of Visit : <u>29-01-2024</u></p> <p>Asset No. : <u>1031123</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>Anita Industrial</u> Model : <u>EIKONEB</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
-------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Fumes not getting

Action Taken : checked the equipment found problem with machine motor piston wear and broken and need to replace Nebuliser motor with piston for further - checking of complaint. Enquired for spare with OEM

Completed Date : 29-1-2024 Time : 1:40pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Nebuliser motor with-</u>	1		
2. <u>- piston</u>			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Ameenulla Cx</u>	<u>29-1-24</u>	<u>12:00pm</u>	<u>1:40pm</u>

Customer Remark Completed Pending



Sabina
MEDICAL OFFICER,
COMMUNITY HEALTH CENTRE
PURATHUR
MALAPPURAM P.O. - 676 561


<p>Service Engineer Name : <u>Ameenulla Cx</u></p> <p>Signature : <i>Ameenulla</i></p> <p>Date : <u>29-1-2024</u></p> <p>Contact Number : <u>9901944433</u></p>	<p>Customer Name : <u>Sabina</u></p> <p>Signature : <i>Sabina</i></p> <p>Date : <u>29/1/24</u></p> <p>Contact Number : <u>9847313900</u></p> <p>Designation : <u>Nursing officer</u></p> <p>Hospital Seal :</p>
---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Malappuram
2	Name of Hospital	CHC parathur
3	Name of Equipment with Make, Model and Serial Number	Name: Nebuliser Make: Anita Industries Model: EIKOMB EL-710 SN: NA
4	Equipment ID & Barcode	1031123
5	Date of purchase / Year of manufacture / Installation Date	20/04/2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	24/01/2024
9	Action taken	checked the nebuliser found problems with motor piston wear and broken and need to replace nebuliser motor for further checking of complaint enquired for spare
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Need To Replace Nebuliser motor with piston
12	Cost of spares (specify parts and cost)	spare cost admit received. since spare are not available in market

13	Asset Value	2999.36 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX service Report attached
16	Reasons for recommending the equipment as BER	checked the nebuliser found Nebuliser pressure pump motor piston broken. Equipment installed - on 20/04/2021 and aged up to 2+ years, spare cost didn't received since spare are not available in market Hence recommended for condemnation
17	Name & Signature of CYRIX Authority	Ameerulla. CK 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked with the nebulizer piston broken and wear. Beyond economic repair.

Jr. Consultant Biomedical Engineer
NHM Malappuram



Signature of JC BM (NHM)

Maximum usage life attained.
Recommended for condemnation.

Date



Signature of
Superintendent / Medical Officer (i/c)

MEDICAL OFFICE
COMMUNITY HEALTH CENTRE
PIRATHUR
MUTIALA P.O. - 676 561

DEPARTMENT OF HEALTH SERVICES

ISSUE NOTE

Sheets No. 01
 Total No Sheets one

I.N.No 003/2021-22
 Date 20/4/21

Sir,

To CHC PURATHUR

with reference to indent No.....dated.....I am Forwarding herewith the following medicines. Please acknowledge and return the original copy duly signed

District Medical Stores


Store Superintendent

Sl.No	Particulars	Qty	Unit	Rate	Total	Mfr	Batch No	Expiry
1.	Fumigation dispenser	01		27, 001.50	01			
2.	Fire extinguisher CO2 type	06		8556.00	06			
3.	Air conditioner 1 ton	02		31804 +69	02			
4.	Otoscope	01		14,000	01			
5.	Refrigerator Model B (with stabilizer)	01		515 8.00	01			
6.	Water purifier with Hot & cold (Aqua guard)	01		125	01			
7.	Crack cart							
8.	Nebulizer			36	2			
9.	Infrared thermometer	04		11	04			
10.	Mobile spot light	01		100	01			
11.	V-guard 400 stabilizer	02		zero	02			

Reg. costly equip.
 vol. 1/1/19
 Reg. Instrument
 vol. 2/1/19
 Reg. costly equip.
 vol. 3/1/19
 Reg. costly equip.
 vol. 4/1/19
 Reg. costly equip.
 vol. 1/1/19
 Reg. costly equip.
 vol. 1/1/19
 Reg. Instrument
 vol. 2/1/19
 Reg. Instrument
 vol. 3/1/19
 Reg. costly equip.
 vol. 4/1/19
 Reg. Instrument
 vol. 5/1/19

The above article have been duly received and found correct in good condition and brought into stock

Pharmacist / Store keeper

Sumayya




Medical Officer in Charge / Superintendent

PURATHUR
 WITTANNUR P.O. 676 561

Stock taken in Reg. costly equipments & Reg. For instrument
 pages concerned.

NEBULIZER.

03.01.20	Inv No: 01/2020 dt: 03.01.20	Rd:- CHC Vettom		
	C Block Panchayat Sooti/20 (2019-20)			
	Philips	2	4600/-	2
03.01.20	Issued to H/Nurse.		2	NIL
	S.S. P.No: 87			
20/4/2021	Rd from CHC Thavaneer, inv. no: 003/2021-2022 dt 20/4/21			
	(ELKO model- EL-710)	2	2999.36	2
20/5/2021	ISSUED to H/N			
		2		NIL.

