


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	LITSI KOTTATHARA
3	Name of Equipment with Make, Model and Serial Number	ELETROLYTE ANALYZER MAKE: CORNLEY MODEL: AFT 300 SLNO: DT05B13060025
4	Equipment ID & Barcode	0913284
5	Date of purchase / Year of manufacture /Installation Date	12/9/2013
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No Amc / Camc
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	10/10/2023
9	Action taken	checked and found that multiplexer damaged and main board faulty
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	1,12,395/-

13	Asset Value	95238 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	118 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX
16	Reasons for recommending the equipment as BER	checked and found that multiplexer damaged and main board faulty. Equipment installed on 12/19/2013 and aged up to 10 year. Comparing lifespan and repairing cost equipment is beyond economical. both criteria for BER met as per tender dates 3/14/
17	Name & Signature of CYRIX Authority	JREGJTHC 


\*Not mandatory #Based on the period of life and value as per the BER guidelines



\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER.



  
Signature of JC BM (NHM)

Date	 Seal	<p style="text-align: center;">           Signature of          Superintendent / Medical Officer (i/c)          Govt. Tribal Speciality Hospital,          Kottathara (P.O), Attappady       </p>
------	---	---



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 144443

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>GOV. TRIBAL SPECIALITY HOSPITAL KOTTATHARA</u>	Call Registration Date : <u>10/10/2023</u>
Address <u>KOTTATHARA</u>	Caller ID : <u>73986</u>
<u>PALAKKAD</u>	Date of Visit : <u>23/10/2023</u>
Ph : <u>0994615467</u>	Asset No. : <u>0913284</u>
	EQPT Name : <u>Electrolyte Analyzer</u>
	Manufacture <u>CORNLEY</u> Model : <u>AET 300</u>
	S. No. <u>DJ05B1306008</u> Dept. <u>Laboratory</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

Problem Identified : Equipment Shows 'No calibrant A' error

Action Taken : checked and found that multiplexer damaged and main board faulty. Equipment installed on 12/9/2013 and aged up to 10 year. Comparing lifespan and repairing costs equipment is beyond economical. As per the KMSCL Tender clause 5.3.14.1, Machine beyond economical repair, so we recommended for condemnation.

Completed  Date : ..... Time : ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SREETHA C</u>	<u>23/10/2023</u>	<u>11:00</u>	<u>12:30</u>

Customer Remark  Completed  Pending

Service Engineer Name : SREETHA C Customer Name : Superintendent  
Signature : [Signature] Date : 23/10/2023  
Date : 23/10/2023 Contact Number : 9072522058  
Contact Number : 9072522058 Designation : Superintendent  
Hospital Seal : Gov. Tribal Speciality Hospital Kottathara (P.O), Attappady







---

**REQUIRED QUOTATION # 0913284 ELECTROLYTE ANALYZER @ GTSH KOTTATHARA , PALAKKAD**

2 messages

---

**Akhilesh P** <akhileshcyrix@gmail.com>

Mon, 16 Oct, 2023 at 8:10 am

To: aspenkochi@gmail.com, Service.RDG <service@rdgc.in>

Cc: binto@rdgc.in, rajaram.k@rdgc.in, service support <servicesupport@rdgc.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>

Dear sir,

Please share the price and stock availability of

1. REFERENCE ELETRODE
2. NA+ , K+ ELETRODE
3. MULTIPLEXER 6 WAY
4. CONTROL BOARD

ELECTROLYTEANALYZER @ GTSH KOTTATHARA , PALAKKAD

MAKE : ASPEN

MODEL : AFT 300

S/N : DJ05B13060005

Thanks & Regards

AKHILESH P

Biomedical Engineer | Kannur

Bemp-Kerala

Mobile : +91 7593847154

Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD

30/641B, Petta, Punithura

Ernakulam- 682038

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

---

**service support** <servicesupport@rdgc.in>

Tue, 17 Oct, 2023 at 12:16 pm

To: Akhilesh P <akhileshcyrix@gmail.com>, aspenkochi@gmail.com

Cc: binto@rdgc.in, rajaram.k@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>, Service.RDG <service@rdgc.in>

Dear sir,

Please find the attached P.I. & share the P.O. along with payment details for billing.

Thanks & Regards

Hari Ram

[Quoted text hidden]



# Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



Proforma Invoice No.: RDPL/S/AD/20231016/23-24		Date : 17-10-2023			
To, Cyrix Healthcare Pvt. Ltd. 30/641B, Petta, Punithura Poonithura, Ernakulam-682038, India		Ref. : By Mail Date : 17-10-2023  Ref.GTSH KOTTATHARA, PALAKKAD			
Sr. No	Cat No.	Description	Price Each Ex- Work	Qty	Total (In Rs.)
		(Instrument Model: Electrolyte Analyzer) (Model :AFT-300) (S/NO. DJ05B13060005)			
A		<u>Spare &amp; Consumables Procurement Details:</u>			
(i)		Multiplexer (6 Way)	24,750.00	1	24,750.00
(ii)		Main board	70,500.00	1	70,500.00
		<b>Total</b>			<b>95,250.00</b>
		<b>(+) GST @ 18%</b>			<b>17,145.00</b>
		<b>Total: One Lakh Twelve Thousand Three Hundred Ninety Five Only.</b>			<b>1,12,395.00</b>

**Terms & Conditions : { Validity: 90 Days } , {Delivery : 3-4 Weeks}, {Payment Terms : 100% Advance }**

<b>Bank Account Details :-</b>	
Name of the Firm : Aspen Diagnostic Pvt. Ltd.	Aspen Diagnostics Pvt. Ltd.
Bank Name : ICICI Bank Limited	
Account No. : 016005003239	
Account Type : Current Account	
MICR Code : 110229029	
IFSC Code : ICIC0000160	
Authorized Signatory	



## Electrolyte Analyzer

Type **AFT-300**

POWER SUPPLY: 100-240V~

SN **DJ05B13060005**

FREQUENCY : 50/60Hz

RATED POWER : 100VA



**2013.06.08**



### Caution:

100-240V~ inside, turn off the power before opening the back cover



Meizhou Cornley Hi-Tech Co., Ltd  
Nanshan Industrial, Baigong, Meixian, Guangdong  
Province - P.R.China - 514765  
TEL: +86-755-86330808 . FAX: +86-755-86330882  
Email: sales@cornley.com Http://www.cornley.com



Renault Petersen Limited  
Couching House  
Couching Street, Watlington  
Oxfordshire OX49 5PX

**CYRIX**<sup>TM</sup>  
HEALTH CARE PVT LTD

**BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (BEMP)**



**TOLL FREE NO: 1800 - 425-7669**

**BAR CODE - (8004-890615225)  
0913284**

