



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 176679

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>CHC MAYYIL</u></p> <p>Address : <u>Mayyil</u> <u>KANNUR</u></p> <p>Ph : <u>7907969533</u></p>	<p>Call Registration Date : <u>30/12/2023</u></p> <p>Caller ID : <u>88056</u></p> <p>Date of Visit : <u>01/01/2024</u></p> <p>Asset No. : <u>1331063</u></p> <p>EQPT Name : <u>HEMATOLOGY ANALYZER</u></p> <p>Manufacture : <u>PROKAN</u> Model : <u>PE 6000</u></p> <p>S. No. : <u>A600616030006TEH-10</u> Dept. : <u>LABORATORY</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : White display and bath leakage

Action Taken : Checked the machine and found that white display - Showing, also found bath leakage and valve not working. Engineded OEM for quotation.

Completed Date : 09/01/2024 Time : 11.00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sayanth</u>	<u>09/01/2024</u>	<u>9.30 AM</u>	<u>11.00 AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Sayanth</u></p> <p>Signature : </p> <p>Date : <u>09/01/2024</u></p> <p>Contact Number : <u>8606345195</u></p>	<p>Customer Name : <u>Namitha.M</u></p> <p>Signature : </p> <p>Date : <u>9/1/24</u></p> <p>Contact Number : <u>953901995</u></p> <p>Designation : <u>Lab Technician</u></p> <p>Hospital Seal :</p>
---	--




Handwritten signature and text in Malayalam script.

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC MAYYIL
3	Name of Equipment with Make, Model and Serial Number	HEMATOLOGY ANALYZER Make: PROKAN Model: PE-6000 SN: A6006160300067EH10
4	Equipment ID/ Barcode	1331063 #88056
5	Date of purchase/ Year of manufacture/Installation Date	14-12-2016
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 30-12-2023
9	Action taken	Checked the machine and found that White display showing, also found bath leakage and valve not working , enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Switching power supply-26250/- 2 way valve -18450/- CCU Board dom card supportable-59500/- WBC Chamber- 35250/- TOTAL =139,450+GST@18% =164,551/-

13	Asset Value	199,083/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	82.6545%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & OEM quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that White display showing, also found bath leakage and valve not working. Enquired OEM for quotation. Equipment installed on 14-12-2016 aged up to 7 years. Repairing cost is 82.6545% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	SARANG K M 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that bath leakage & valve not working. Aged up to to 7 years. its service cost 82% above as per OEM'S quotation. so recommended for BER as per BER Protocol.



Signature of JC BM (NHM)

10/01/2024

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Date: 11.1.2024





Signature of
Superintendent/Medical Officer (i/c)

മെഡിക്കൽ ഓഫീസർ
സാമൂഹികാരോഗ്യ കേന്ദ്രം
മയിൽ



Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



ISO 9001:2015
CERTIFIED COMPANY

Proforma Invoice No.: ADPL/S/RD/20240104/23-24 Date :04.01.2024

To,
Cyrix Healthcare Pvt. Ltd.
Ernakulam, First Floor, Petta Junction
30/641B, Poonithura, kerala

Ref. : By Mail.
Date :04.01.2024
1331063 CHC MAYYIL
, KANNUR

Sr. No	Description	Part No.	Price Each Ex-Work	Qty	Total (In Rs.)
	(Instrument Model: Hematology Analyzer) (Model : PE-6000) (S/No. A6006160300067EH10)				
A	Spares Procurement Details:				
(i)	Switching Power Supply	PE00-013	26,250.00	1	26,250.00
(ii)	2-Way Valve	PE00-005	18,450.00	1	18,450.00
(iii)	CCU Board Dom Card Supportable	PE60-010	59,500.00	1	59,500.00
(iv)	WBC Chamber	PE60-003	35,250.00	1	35,250.00
Total					1,39,450.00
(+ GST @ 18%					25,101.00
Total: One Lakh Sixty Four Thousand Five Hundred Fifty One Only.					1,64,551.00

Terms & Conditions : { Validity: 90 Days } { Delivery : 3-4 Weeks } { Payment Terms : 100% Advance }

Bank Account Details :-

Name of the Firm : Aspen Diagnostic Pvt. Ltd.	<p>Rajaram Krishnamurthy National Service Manager</p>
Bank Name : ICICI Bank Limited	
Account No. : 016005003239	
Account Type : Current Account	
MICR Code : 110229029	
IFSC Code : ICIC0000160	



FULLY AUTO HEMATOLOGY ANALYZER

MODEL: PE-6000
SERIAL NUMBER: A6006160300067EH 10
 100-240V- 50/60Hz ≤150VA




SHENZHEN PROKAN ELECTRONICS INC.
 No.1002,Bld. 4,Fantasia MIC Plaza,Pengji
 Pioneering Park,Nanshan,Shenzhen,518067 P.R.China
 Factory Add.:The 2nd floor, Yanda tech-park, FengHuangGang,
 XIXiang, Bao'An district, Shenzhen,518126 P.R.China
 Shanghai International Holding Corp.GmbH (Europe)
 Eiffelstraße 80. 20537 Hamburg, Germany





Handwritten signature
 മെഡിക്കൽ ഓഫീസർ
 സാമൂഹികാരോഗ്യ കേന്ദ്രം
 മന്ദിരം

HAEMATOLOGY ANALYSER

Date

14/12/16

Invoice RDPL/1064/16-17

Rapid diagnostics. From KMSCL.

Total

1 - Haematology Analyser - 2,14,900 ~~Amount~~ - 2,14,900

2 - UPS with back up mod - 1 NOS.

3. Diluent 20L for PE6 - 2 bottle

4. Rinse 10L for PE6 - 1 bottle

5. Rinse 1L " " 6000 - 1 bottle

6. Lyse 500 ml for PE6 - 2 bottle

7. Lyse 200 ml for PE60 - 1 bottle

8. Probe clean 100 ml For - 1 NOS

9. E-2 cleaner 100 ml For - 1 NOS

10. Diluent 1L for PE6 - 3 bottle

Grand Total - 2,25,645.00



മെഡിക്കൽ ഓഫീസർ
സാമൂഹികാരോഗ്യ കേന്ദ്രം
മുളിൽ



REQUIRED QUOTATION # 1331063 HEMATOLOGY ANALYZER @ CHC MAYYIL , KANNUR

7 messages

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 3 Jan, 2024 at 9:32 am

To: Service.RDG <service@rdgc.in>, service support <servicesupport@rdgc.in>

Cc: rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Dear sir,

Please share the price and stock availability of

1. Switched Mode Power Supply
2. Wbc bath
3. 2 way valve
4. Cpu board

HEMATOLOGY ANALYZER @ CHC MAYYIL , KANNUR

MAKE : RAPID

MODEL : PE 6000

S/N : A6006160300067EH10

Thanks & Regards

AKHILESH P

SPECIALIST ENGINEER - LAB NORTH ZONE

Bemp-Kerala

Mobile : +91 7593847154

Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD

30/641B, Petta, Punithura

Ernakulam- 682038

www.cyrixhealthcare.com

Service.RDG <service@rdgc.in>

Wed, 3 Jan, 2024 at 1:49 pm

To: Akhilesh P <akhileshcyrix@gmail.com>

Cc: rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>, service support <servicesupport@rdgc.in>

Dear Sir,

Kindly share the picture of CPU Board for confirmation.

Thanks & Regards,

Jitendra.

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 3 Jan, 2024 at 4:09 pm

To: Service.RDG <service@rdgc.in>

Cc: rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>, service support <servicesupport@rdgc.in>

Dear sir

Please find the attached picture for more clarification.

Thanks & Regards

AKHILESH P
SPECIALIST ENGINEER - LAB NORTH ZONE
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

Wed, 3 Jan, 2024 at 4:55 pm

Service.RDG <service@rdgc.in>

To: Akhilesh P <akhileshcyrix@gmail.com>, service support <servicesupport@rdgc.in>

Cc: rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Dear Sir,

Please find the below prices, all are available in stock.

@Hari, please prepare the P.I. accordingly & share.

<u>Part No.</u>	<u>Part Name</u>	<u>Unit Price</u>
· PE00-013	Switching Power Supply	26250
· PE60-003	WBC Chamber	35250
· PE00-005	2-Way Valve	18450
· PE60-010	CCU Board Dom Card Supportable)	59500

Note: GST 18% Extra.

Thanks & Regards,

Jitendra

[Quoted text hidden]

Wed, 3 Jan, 2024 at 9:31 pm

Akhilesh P <akhileshcyrix@gmail.com>

To: Service.RDG <service@rdgc.in>

Cc: service support <servicesupport@rdgc.in>, rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Dear sir

Kindly share pi in letter head with seal and signature is mandatory

Thanks & Regards

AKHILESH P
SPECIALIST ENGINEER - LAB NORTH ZONE
Bemp-Kerala

Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Thu, 4 Jan, 2024 at 1:08 pm

To: Service.RDG <service@rdgc.in>

Cc: service support <servicesupport@rdgc.in>, rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Dear sir

Kindly update the status

Thanks & Regards

AKHILESH P
SPECIALIST ENGINEER - LAB NORTH ZONE
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

Service.RDG <service@rdgc.in>

Thu, 4 Jan, 2024 at 6:05 pm

To: Akhilesh P <akhileshcyrix@gmail.com>

Cc: service support <servicesupport@rdgc.in>, rajaram.k@rdgc.in, binto@rdgc.in, Blesson Jose <Zm2.klbemp@cyrix.in>, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>

Dear Sir,

Please find the attached P.I. & share the P.O. along payment details for billing.

[Quoted text hidden]