



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 148530

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC ORKKATTERI</u></p> <p>Address <u>KOZHIKODE</u></p> <p>Ph : <u>7907093654</u></p>	<p>Call Registration Date : <u>13-12-2023</u></p> <p>Caller ID : <u>85553</u></p> <p>Date of Visit : <u>14-12-2023</u></p> <p>Asset No. : <u>1130777</u></p> <p>EQPT Name : <u>Centrifuge</u></p> <p>Manufacture <u>Remi</u> Model : <u>R4C</u></p> <p>S. No. <u>AELO-589</u> Dept. <u>Lab</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Centrifuge rotation problem.

Action Taken : checked and found transformer, electric motor and rheostate defective. Need to replace these spares for further checking and working condition of the equipment. Enquired the spares.

Completed Date : 25-01-24 Time : 12pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Abijith. A</u>	<u>25-01-24</u>	<u>11-30am</u>	<u>12pm</u>

Customer Remark Completed Pending

Service Engineer Name : Abijith. A
Signature :
Date : 25-01-2024
Contact Number : 7593847



Recommending for condemnation

Customer Name : Vanika.TK
Signature :
Date : 29/1/24
Contact Number : 9496387595
Designation : Lab Technician
Hospital Seal :


മോഡറൻ ജിപി - 036015 NA
 ഫോൺ: 0496 - 2547

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	CHC ORKKATTERI
3	Name of Equipment with Make, Model and Serial Number	Equipment :Centrifuge Make : Remi Model :R4C SN : AELO-589
4	Equipment ID & Barcode	85553 & 1130777
5	Date of purchase / Year of manufacture /Installation Date	17-06-2011
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	13-12-2023 (Toll free)
9	Action taken	Checked and found transformer,electric motor and rheostat defective.Enquired the spare with vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	1)Transformer – Rs-971 2)Electric motor – Rs-6443 3)Rheostate – Rs-1624 Total – Rs-9038

13	Asset Value	Rs . 9651/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	93%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found transformer ,electric motor and rheostat defective. Equipment out lived 12 + years. Repairing cost of the equipment is 93% both criteria met, so recommended equipment for condemnation.
17	Name & Signature of CYRIX Authority	ABIJITH.A 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that transformer, electric motor & rheostat defective. Equipment installed on 17/12/2011. Repairing cost of equipment is 93% of asset value. So recommended equipment for condemnation.

Consultant Bio Medical Engineer
National Health Mission
Kozhikode


Signature of JC BM (NHM)

Date 29/1/24

Vanika.T.k. 


 Seal
 മലപ്പുറം ആരോഗ്യകേന്ദ്രം
 നാഷണൽ ആരോഗ്യകേന്ദ്രം
 ഓർക്കാട്ടൂരി, നടകര
 ടെലിഫോൺ: 0496 - 2547182

Signature of
 Superintendent / Medical Officer (i/c)

 Dr. Usha

**REGARDING QUOTATION - 85553 -
1130777 - CHC ORKKATTERI -
CENTRIFUGE**

From: Abijith A abijithacyrix@gmail.com

To: Service@Beston Industries

service@bestonindustries.com,

inlabservice@gmail.com

inlabservice@gmail.com, inlabclt@gmail.com

inlabclt@gmail.com

Cc: Blesson Jose zm2.klbemp@cyrix.in,

Assistant Manager Purchase

ampurchase@cyrix.in

Sent: Thursday, 4 January, 4:03 pm

Dear Sir,

Please send quotation for the
mentioned spare parts of Centrifuge Remi-R4C

1)Motor

2)Transformer

3)Rheostat

--

Thanks & Regards,

Abijith A

District Incharge

Mob-7593847141

abijithacyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin

-38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR

PRADESH



Abijith A <abijithacyrix@gmail.com>

QUOTATION CUM PROFORMA INVOICE FOR CYRIX HEALTH CARE DT 24-1-2024

instservicekochi <instservicekochi@remigroup.com>

Wed, Jan 24, 2024 at 2:21 PM

To: abijithacyrix@gmail.com

Cc: raman@remilabworld.com, kochisales.hcd@remigroup.com, kochi@remigroup.com

Dear Sir,


Greetings from REMI GROUP.

As requested attached quote for your reference. **PLEASE RELEASE THE PAYEMENT IN ADVANCE .**

Regards,

Rebitha

0484-2401361

 **Engineer Quotation-50 cyrix pdf sign.pdf**
366K

REMI SALES & ENGINEERING LIMITED

REMI GROUP

70/1750,5th Cross Road , Link Park ,Pottakuzhi Road, Pachalam P.O, Kochi-682012
Phone : 0484-2401361,Email Id.:instservicekochi@remigroup.com/service@remilabworld.com

Corporate Off : Remi House, 3rd floor, 11 Cama Indl. Estate, Goregaon (East) Mumbai 400 063. INDIA
• Board: +91 22 4058 9888 SALES SUPPORT NO.8879448682• Email Id.: sales@remilabworld.com

QUOTATION CUM PROFORMA INVOICE

QN :50

Complaint No.	Customer Name & Address.:					Quotation No	PRINTED		
Dated	24/1/2024	Cyrix Health care pvt ltd.					Date	24/1/2024	
ENGINEER NAME	Vipin	2nd Floor, Silver Plaza, Mavoor Rd, Near Moffusil Bus Stand, Arayidathupalam, Kozhikode, Kerala 673004					REF	Phone	
REF.	Phone	7593847141					DATE	24/1/2024	
Email	abijithacyrix@gmail.com	Abijith							
Model no.	R-4C	SR NO. /MFG DT.	AELC-589/01/2011						
Sr No.	Description	Part code	HSN/SAC	Rate Rs.	Qty	Amount Rs. (A)	GST % (B)	GST Amt (C)	Total Amt (A+C)
1	TRANSFORMER (NRT)	L22000019	850490	823	1	823	18%	148	971
2	ELECTRIC MOTOR (RT)	ML25000140	850140	5460	1	5460	18%	983	6443
3	RHEOSTATE 1.2K 100W(NRT)	L22000003	401699	1376	1	1376	18%	248	1624
								TOTAL (D)	9038
CALIBRATION CHARGES		0	0	0	0	0	0	0	0
SERVICE CHARGES		0	0	0	0	0	0	0	0
								TOTAL (E)	9038
Rupees in Words:		Rs. Nine thousand thirty eight only.							9038

Terms & Conditions:

- Spares are suitable only to use with REMI instruments & not for any other make.
- We will accept return of spares only if they are unused and not damaged. At our discretion, the returned spare will be replaced. The difference in prices, if any, will have to be borne by you.
- To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
- 100 % Advance along with your order confirmation
- Payment can be made by Cheque/ NEFT in our Bank account. For Cash Payment you must collect Receipt from recipient of cash failing which REMI will not be held responsible for any discrepancy.
- Courier/ transportation charges are payable at destination & not included in our price.
- Validity.** The quotation is valid for 30 days from the date of this quotation / Proforma Invoice
- Subject to Ernakulam Jurisdiction only

REMARKS : PAYMENT 100% ADVANCE

For REMI SALES & ENGINEERING LTD


 Authorized Signatory

BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

- | | | |
|---------------------------|----------------------------------|-------------------------------|
| • A/C NO: 35699154671 | • Bank Name: STATE BANK OF INDIA | • BRANCH: ELAMAKKARA |
| • IFS CODE: SBIN0013223 | • MICR CODE:682002053 | • SWIFT CODE: SBININBB395 |
| • GST NO: 32AAACR0421M1ZS | • PAN NO: AAACR0421M | • CIN NO. U31100MH198PL022314 |



മെഡിക്കൽ ഓഫീസർ
 സാമൂഹിക ആരോഗ്യകേന്ദ്രം
 - ഓർക്കാട്ടേരി, വടകര
 കോഴിക്കോട് ജില്ല - 673 501
 ഫോൺ: 0496 - 2547182



Dr-12849.00