

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	W & C HOSPITAL PALAKKAD
3	Name of Equipment with Make, Model and Serial Number	Fetal Doppler make - Perfecta model - FD 620C SN - F620C017000330
4	Equipment ID/ Barcode	81795/0912442
5	Date of purchase/ Year of manufacture/Installation Date	02/07/2018
6	Warranty details	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	23/11/2023 (Toll Free)
9	Action taken	checked the equipment found main board, speaker doppler probe with cable faulty
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Details of service required (Recommendations for repair)	Need to replace main board, speaker & doppler probe with cable
12	Cost of spares (specify parts and cost)	1) main board - 1746 2) speaker - 472 3) Doppler probe with cable set } - 1357 Total - 3575/-


13	Asset Value	4000/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	89%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix report and spare quotation attached
16	Reasons for recommending the equipment as BER	checked the equipment found main board, speaker & doppler probe with cable damaged. Equipment installed on 02/07/18 and aged up to 5+ year. Repairing cost is 89% of asset value. so we recommended for condemnation
17	Signature of CYRIX Authority	LIBIL LAL <i>libillal</i>

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Main board, speaker and doppler probe with cable set damaged hence recommended for condemnation.




Joshi
BME NHM
11/12/2023

Signature of JC BM (NHM)

Consultant
NHM (Arogya Keralam), Palakkad

Forwarding to
Super. Revenue CS
SND

Date



[Signature]
11/12/2023

Signature of
Superintendent/Medical Officer (i/c)

GOVT. WAC Hospital
Palakkad



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 165840

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility	W.C Hospital	Call Registration Date :	23/11/2023
Address	Palakkad	Caller ID :	81795
Ph :	9495170195	Date of Visit :	24/11/2023
		Asset No. :	0912442
		EQPT Name :	Fetal Doppler
		Manufacture	Perfecxa Model : FD-820C
		S. No.	F620C017000330 Dept. Antenatal ward

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Speaker not working, Doppler probe with cable damaged, main board faulty*

Action Taken : *Checked the equipment and found problem with main board, doppler probe with cable and speaker. Need to replace these faulty spare for further checking. Enquired for spare.*

Completed Date : *11/12/23* Time : *12:30pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>LIBIL LALIL</i>	<i>11/12/23</i>	<i>12:00pm</i>	<i>12:30pm</i>



Customer Remark Completed Pending



11/12/2023
SUPERINTENDENT
Govt. W&C Hospital
Palakkad

Service Engineer Name : *LIBIL LALIL*
Signature : *[Signature]*
Date : *11/12/2023*
Contact Number : *9961933355*

Customer Name : *REENA-C-3*
Signature : *[Signature]*
Date : *11/12/2023*
Contact Number : *9495170195*
Designation : *SNB*
Hospital Seal :



REENA C-S
 SUPERINTENDENT
 Govt. W&C Hospital
 Palakkad
 REENA C-S
 SNO

MEDADE HEALTHCARE

Engineered For Healthcare

Name:	Cyrix Healthcare Pvt Ltd	Date:	08-12-2023
Address:	Poonithura, Ernakulam	Qtn#	MHK/QTN/23-24/092
Phone:		Salesman:	ARJUN
		Mob#	8129102919

QUOTATION

SL No	Description	Price/Unit	Gst	Qty	Total (incl tax)
2	Perfexa Fetal Doppler FD-620C				
	Mainboard	₹ 1,480.00	18%	1	₹ 1,746.40
	Speaker	₹ 400.00	18%	1	₹ 472.00
	Probe with cable set	₹ 1,150.00	18%	1	₹ 1,357.00
TOTAL					₹ 3,575.40

Terms and Conditions

1. Payment: 50% ADVANCE: 50% AGAINST DELIVERY
2. Quotation validity: 30 days.
3. Delivery time : 2-4 weeks

For Medade healthcare



ARJUN.R

Address : KM32/1475 A4,Ground floor , Saaff Arcade, New bus stand link road, Koyilandy-673305

Phone: 8129102919, 7012112424, Email: sales.medade@gmail.com

Foetal Doppler.

18.12	Rd From Store CRd From Store BPL Limited No-35-261 milkamam Road Koommathai, Colappally, eachim Pincode, 682024, Phone - 0484-2558418/4014406 Invoice NO EK01FG1427 11-12 Dtd. 31.1.2012. SIO NO A S/5817/2011 DT 3/10/2011 - Bill for RS 4000.03. (RS 4806.00). Issued to wd.	1		1
12/13	Transfer in from LW (Stock book page no: 263) Rd from Yermangy Surgicent's Bill no. 310 dtcd 7/11/2013 Issued to ward Rs 8750/-	1		2
1-2014	Tire out to unseizable Reg.			
1-2015	Tire out to unseizable Reg.			
3-1-2017	Condemned As per order no B/3196/2016/DT 13-1-17 As per order reduced stock Book			2 Nil @ As
2.7.18	Rd from Nic Store (LP) Issued to nic	1		1
20.1.20	Rd from Store LIP	1		2 2
20.1.20	Issued to ward	1		NIL 2



SUPERINTENDENT
Govt. W&C H. Palakkad
REENA-C.S
SNO

Fwd: Quotation

From: Libil Lal L libilcyrix@gmail.com
To: Blesson Jose zm2.klbemp@cyrix.in
Sent: Thursday, 14 March, 9:41 am

Sir

For your reference
PFA quotation

Thanks & Regards

LIBIL LAL L
Sr Biomedical Engineer
Palakkad
Mob - 9961933355
Mail - libilcyrix@gmail.com
Cyrix Health Care Pvt Ltd

----- Forwarded message -----

From: **Medade healthcare**
<sales.medade@gmail.com>
Date: Thu, Mar 14, 2024, 9:40 AM
Subject: Quotation
To: <libilcyrix@gmail.com>

Dear Libil,

Please find the attached quotation for the foetal
Doppler spares.

Thanks and regards,

MEDADE HEALTHCARE

**KM 32/1475 A4 Ground, Saaff Arcade,
New Bus Stand Link Road Koyilandy,
Koyilandy - 673305,**

Kozhikode, Kerala

Tel: +9181 2910 2919

+9170 1211 2424