



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 154724

ISO 13485 : 2012 & ISO 9001:2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ... <u>CHC Mattannur</u> ...</p> <p>Address ... <u>Mattannur</u> ...</p> <p style="padding-left: 40px;"><u>Kannur</u></p> <p>Ph : <u>9446505266</u></p>	<p>Call Registration Date : <u>20/10/23</u></p> <p>Caller ID : <u>75989</u></p> <p>Date of Visit : <u>25/10/23</u></p> <p>Asset No. : <u>1330286</u></p> <p>EQPT Name : <u>BP apparatus</u></p> <p>Manufacture <u>Beurer</u> Model : <u>BM29</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : not working

Action Taken : checked the equipment found equipment is not working found problem with motor main board and pressure sensor. Equipment installed on 3/10/18 equipment aged more than 5 years also comparing spare cost and machine cost beyond same. so we are recommending machine for

Completed Date : 25/10/23 Time : 10:35 Condemnation Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Avinash</u>	<u>25/10/23</u>	<u>10.00</u>	<u>10.35</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Avinash</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>25/10/23</u></p> <p>Contact Number : <u>6364064948</u></p>	<p>Customer Name : <u>[Name]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>25/10/23</u></p> <p>Contact Number : <u>9446505266</u></p> <p>Designation : <u>SNO</u></p> <p>Hospital Seal :</p>
---	---




[Signature]
മെഡിക്കൽ ഓഫീസർ
സംസ്ഥാന ആരോഗ്യ കേന്ദ്രം

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kannur
2	Name of Hospital	CHC Mattannur
3	Name of Equipment with Make, Model and Serial Number	BP apparatus Make - Beurer model - BM29 SerialNo - NA
4	Equipment ID & Barcode	1330286 / 75989
5	Date of purchase / Year of manufacture / Installation Date	3/10/18
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	20/10/23
9	Action taken	checked the equipment found mainboard motor defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	NOT Recommended
12	Cost of spares (specify parts and cost)	

13	Asset Value	2200
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	100%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service Report attached
16	Reasons for recommending the equipment as BER	checked the equipment found equipment is not working found problem with motor mainboard and pressure sensor equipment installed on 31/10/2023 equipment aged more than 5 years also comparing spare cost and machine cost beyond same. so we are recommending this machine for condemnation.
17	Name & Signature of CYRIX Authority	Avinagh T 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that motor, mainboard & pressure sensor defective. Aged up to 5 years. its service cost economically not feasible. So recommended for BER as per BER Protocol.


27/10/2023
Rachin
Signature of JC BM (NHM)
CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Regeene Meltaw s/o
Rajan


Date 25/10/23


Signature of
Superintendent Medical Officer (i/c)



AR MEDICARE

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL: armedicare07@gmail.com
MOB: 7994654108, 9446057350

OUR REF : AR/QTN/112-2023
DATE : 25/10/2023

TO
CYRIX HEALTHCARE
ERNAKULAM

Respected Sir,

SUB: QUOTATION FOR DIGITAL BP APPARATUS PARTS

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

In case of any further clarifications kindly feel free to contact us.

Thanking you and assuring you of our best attention and engineering services at all times.

Waiting for your earliest reply.

Yours truly,

For AR MEDICARE,
KANNUR.



TRUST US WITH YOUR NEEDS

OUR REF : AR/QTN/112-2023
DATE : 25/10/2023

SLNO:	DESCRIPTION	TOTAL(INCL.ALL TAXES)
01	BEURER DIGITAL BP APPARATUS MAINBOARD	1200/-
02	PRESSURE SENSOR	400/-
03	MOTOR	600/-
	TOTAL	2,200/-

TERMS & CONDITIONS OF SALE

Payment Terms : Complete Payment Within 30 Days

Delivery : Within 4 Days from the date of purchase order .



AR MEDICARE

TRUST US WITH YOUR NEEDS



**Fwd: QUOTATION MATTANUR BP
1330286**

From: Sarang KM sarangcyrix@gmail.com

To: Blesson Jose zm2.klbemp@cyrix.in

Sent: Wednesday, 6 March, 11:06 am

FYI

----- Forwarded message -----

From: ar medicare <armedicare07@gmail.com>

Date: Sat, Oct 28, 2023 at 9:07 AM

Subject: QUOTATION

To: Sarang KM <sarangcyrix@gmail.com>

Sir,

PFA



AR MEDICARE

Thanks & Regards

AR MEDICARE

Room No:865,2nd floor

AL-RAHA Arcade

Thazhe Chovva, Kannur

Mob: 9446057350, 7994654108, 8138091108

Email: armedicare07@gmail.com

GSTIN :32BDYPV5030P1ZP

—
Thanks & Regards

Sarang.k.m

Divisional Incharge|North Zone