



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 151421

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>Call Thalassemia</u>	Call Registration Date : <u>6-11-23</u>
Address <u>Thalassery</u>	Caller ID : <u>78627</u>
<u>Kannur</u>	Date of Visit : <u>6-11-23</u>
Ph : <u>9495800526</u>	Asset No. : <u>1311919</u>
	EQPT Name : <u>UPS</u>
	Manufacture <u>Energia</u> Model : <u>Energylite 3kVA</u>
	S. No. <u>NA</u> Dept. <u>OT</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not working.

Action Taken : Checked the machine and found that main board and control card are complaint. This machine was installed on 21-7-2018 and aged upto 5 years. Repairing cost and replacement of spares are beyond economic. So we recommend this machine for condemnation.

Completed Date : 6-11-23 Time : 1:30 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasilraj</u>	<u>6-11-23</u>	<u>1:00 PM</u>	<u>1:30 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Rasilraj
Signature : Rasilraj
Date : 6-11-23
Contact Number : 9995521436




Customer Name : Chandni kumar
Signature : Chandni
Date : 6/11/2023
Contact Number : 9495800526
Registration : SAD
Hospital Seal : RESIDENT MEDICAL OFFICER GENERAL HOSPITAL THALASSERY

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kannur
2	Name of Hospital	GH Thalassery
3	Name of Equipment with Make, Model and Serial Number	UPS Make: Energio Model: Energio lite 3kva s/nr: NA
4	Equipment ID & Barcode	1311919
5	Date of purchase / Year of manufacture /Installation Date	21-7-2018
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	6-11-2023
9	Action taken	checked the machine and tried to service main board.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Need to replace main board and control card
12	Cost of spares (specify parts and cost)	—


13	Asset Value	Rs 19000 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	71.73%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix report
16	Reasons for recommending the equipment as BER	checked the equipment and found main board and control card are complaint. this machine was installed on 21-7-2018 and aged upto 5 years repairing cost and replacement of spares are beyond economic. so we recommend this machine for condemnation
17	Name & Signature of CYRIX Authority	Rasitroy 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that main board and control card are in complaint. Aged up to 5 years. The repairing cost economically not feasible as per reestation. So recommended for BER as per BER Protocol.


Signature of JC BM (NHM)
17/11/2023

Chandri. Ramesh
Chand SNO
6/11/2022

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002


RESIDENT MEDICAL OFFICER
GENERAL HOSPITAL
THALASSERY


GENERAL HOSPITAL
THALASSERY - 670 001

Date



S P D POWER LABZ Pvt Ltd.
Securing Power Distribution

GSTIN/UIN: 32ABJCS4412K1ZD

S P D Power Labz Pvt Ltd
Ground Floor,66/1618,Peyoli Road,
Kacheripady,Ernakulam,Pin-682018.

06 NOV 2023
1 WEEK
SPD/UPS/SER/ 231101/3

Customer:

CYRIX HEALTH CARE

SUB: REPAIR CHARGE FOR GOV. HOSPITAL,THALASSERY,3 KVA UPS

SL.NO.	DESCRIPTION	QTY	UNIT PRICE	TOTAL
1	ENERGIO LITE 3 KVA UPS, 1311919 Ups servicing, Battery checking. Main Board servicing, Control card change.	1	11550.00	11550.00

Subtotal	INR	11550.00
Discount	N/A	
GST RATE	18%	
GST		2079.00
Total	INR	13629.00

Please confirm your acceptance of this quote by signing this document

Signature:

Name:

Date:

Thank you for your business!

Should you have any enquiries concerning this quote, please contact

SPD POWER LABZ - COMPLETE SOLUTION FOR UPS & BATTERIES

Tel:0484-4014004

Email: service.ups@spdpowerlabz.in

Mobile:8129998245

GSTIN/UIN: 32ABJCS4412K1ZD

Online U P S , Offline U P S , Servo Stabilizer , Batteries - Sales & Service

RBER

CYRIX
HEALTH CARE SYSTEMS
"Genius In Healthcare"

Ticket No.: 76627

RBER approved on:

Condemn on:

Remark:

enjoy elite

ON / MUTE

SELECT

OFF / ENTER



Uninterruptible Power Supply (UPS)

Model: RESQ3K8SB

Input: 220Vac-240Vac, 50/60Hz, 16A, 1 Φ

Output: 220Vac-240Vac, 50/60Hz, 13.6A, 1 Φ

Capacity: 3000VA/2700W



83321801100938

MADE IN CHINA

IS 16242 (Part 1)/
IEC 62040-1



R-41067318

**Fwd: Service Quotation 1311919 GH
UPS**

From: Sarang KM sarangcyrix@gmail.com
To: Blesson Jose zm2.klbemp@cyrix.in
Sent: Wednesday, 6 March, 11:08 am

FYI

----- Forwarded message -----

From: **Service.UPS**
<service.ups@spdpowerlabz.in>
Date: Mon, Nov 20, 2023 at 5:10 PM
Subject: Service Quotation
To: <sarangcyrix@gmail.com>
Cc: Deepak K N <deepakkn@spdpowerlabz.in>, <praveenk@spdpowerlabz.in>, Shinoj Mohan <shinoj@spdpowerlabz.in>

Dear Sir ,

Greetings of the day,

Please find the attached service quotation for
[Govt.Hospital](#),Thallassery.

**(Neona UPS changed to SPD Power Labz Pvt
Ltd)**

COMPLETE SOLUTION FOR UPS & BATTERIES

(UPS AMC & SERVICE PROVIDER)

With Thanks & Regards

Praveen K



SPD POWER LABZ Pvt Ltd.
Securing Power Distribution

SPD
Power
Labz Pvt
Ltd
Ground

Floor 66/16/18, Peyoli Road
Kacheripady, Ernakulam 682018
Phone: 0484 - 4014004
Mobile : 8129998245
Email: praveenk@spdpowerlabz.in
Website: www.spdpowerlabz.in

--

Thanks & Regards

Sarang.k.m
Divisional Incharge|North Zone
Bemp-Kerala
Mobile : [+91 7593847150](tel:+917593847150)
Email : sarangcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala-
682038
www.cyrixhealthcare.com

P Save Environment, Please do not print this
e-mail unless it's essential