



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 155405

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>CHC IRIKKUR</u>	Call Registration Date : <u>9/11/23</u>
Address <u>IRIKKUR</u>	Caller ID : <u>78162</u>
<u>KANNUR</u>	Date of Visit : <u>9/11/23</u>
Ph : <u>9446447253</u>	Asset No. : <u>1330525</u>
	EQPT Name : <u>Refrigerator</u>
	Manufacture <u>Electrolux</u> Model : <u>na</u>
	S. No. <u>na</u> Dept. <u>OP</u>

Service Classification : Breakdown Call **PMS** **Callibration** **Cust.Training**

Problem Identified : Not working

Action Taken : checked the equipment found equipment not working. Found Compressor faulty and door fully damaged and rusted. equipment installed on 22/9/2010, aged upto 13 years. comparing the life span equipment used. repairing cost and replacment of spare beyond economic and so we recommended to RBER

Completed **Date :** 9/11/23 **Time :** 12:45 **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin</u>	<u>9/11/23</u>	<u>10:30</u>	<u>12:45</u>

Customer Remark **Completed** **Pending**

Service Engineer Name : Aswin
Signature :
Date : 9/11/23
Contact Number : 8801012759



Customer Name : Anunimaim
Signature :
Date : 9/11/23
Contact Number : 9446447253
Designation : SNO
Hospital Seal :


**MEDICAL OFFICER
COMMUNITY HEALTH CENTRE**

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC IRIKKUR
3	Name of Equipment with Make, Model and Serial Number	Refrigerator make: electrolux model: Na Serial number: Na
4	Equipment ID & Barcode	78162 & 1330525
5	Date of purchase / Year of manufacture / Installation Date	22/09/2010
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/ No CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	31/1/2023
9	Action taken	checked and found compressor faulty, door fully damaged and rusted.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No Recommendations
12	Cost of spares (specify parts and cost)	—

13	Asset Value	17,961
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	69.5 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service Report Attached
16	Reasons for recommending the equipment as BER	Checked the equipment found equipment not working. Found compressor faulty, doors fully damaged and rusted. equipment installed on 22/09/2010, aged upto 13 years. Comparing the life span equipment used. Repairing cost and replacement of spare beyond economic and so we are recommended to RBER
17	Name & Signature of CYRIX Authority	Aswin Ramachandran 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & identified that compressor found body fully rusted and gas filters are damaged. Its service cost economically not feasible as per quotation and aged upto 13 years. So recommended for BER as per BER Protocol.

Signature of JC BM (NHM)

CENTRAL
NATIONAL HEALTH MISSION
KANNUR - 670002


14/11/2023.

Arunima.M

Senior Nsg. Officer,
9446447253.
Date 9/11/23.


Seal


NATIONAL HEALTH MISSION
KANNUR - 670002


14/11/2023
MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
KANNUR - 670002
Superintendent / Medical Officer (i/c)

PRIME COOLING SYSTEMS SALES & SERVICE

Panappuzha Road, Mathamangalam
(P.O.) M. M. Bazar, Kannur Dt., Kerala- 670306
Email: primecoolingsystems@gmail.com

Mob: 8156912947
7012820516

Date: 27-11-2025

Quotation

Ekooluge Refrigerator Works.

1	Compressor	3500.00
2	Gross - Filling	2500.00
3	New Door	4000.00
4	Door fixing charge	2500.00
	<u>Total</u>	<u>12500.00</u>



All Type of Splitters & Refrigerators Repairing



VESTAR







Electrolux

POWER SAVINGS GUIDE

Model	Year	Energy Star Rating	Estimated Annual Energy Consumption (kWh)









Sarang KM <sarangcyrix@gmail.com>

Fwd: Quotation RBER CHC IRIKKUR REFRIGERATOR 1330525 #78162

2 messages

Sarang KM <sarangcyrix@gmail.com>
To: "DnD Dept. KLBEMP" <dnd.bempkl@cyrix.in>

Mon, Nov 20, 2023 at 4:53 PM

----- Forwarded message -----

From: **Prakash Kp** <primecoolingsystem@gmail.com>
Date: Wed, Nov 8, 2023 at 3:32 PM
Subject: Quotation
To: sarangcyrix@gmail.com <sarangcyrix@gmail.com>

--
Thanks & Regards

Sarang.k.m
District Incharge | Kannur
Bemp-Kerala
Mobile : **+91 7593847150**
Email : sarangcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com



Save Environment, Please do not print this e-mail unless it's essential



IMG-20231108-WA0044.jpg
45K

Sarang KM <sarangcyrix@gmail.com>
To: BLESSEN JOSE <zm2.klbemp@cyrix.in>

Fri, Mar 15, 2024 at 12:49 PM

FYI
[Quoted text hidden]
Thanks & Regards

Sarang.k.m
Divisional Incharge|North Zone
[Quoted text hidden]



IMG-20231108-WA0044.jpg
45K