



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 144544

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : THQH Chittur	Call Registration Date : 10/10/23
Address : Palakkad	Caller ID : 73859
Ph : 9496238396	Date of Visit : 20/10/23
	Asset No. : 0920181
	EQPT Name : Centrifuge
	Manufacture : Remi Model : R8C-DX
	S. No. CHLC-5218 Dept. Lab

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT WORKING

Action Taken : Checked and found motor, speed regulator, timer unit, antivibrator and control board defective. Equipment installed on 28/05/2014. Repairing cost is above 60% of asset value. machine beyond economical repair & recommended for condemnation.

Completed Date : 20/10/23 Time : 2:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
LIBIL LAL-L	20/10/23	1:00pm	2:00pm

Customer Remark Completed Pending

Service Engineer Name : LIBIL LAL-L
Signature : *Libil*
Date : 20/10/23
Contact Number : 9961933355



Customer Name :
Signature : *AP*
Date : 20/23
Contact Number :
Designation :
Hospital Seal :


SUPERINTENDENT
Palakkad Head Quarters Hospital
Chittur-678 201, Palakkad

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Palakkad
2	Name of Hospital	THQH Chittur
3	Name of Equipment with Make, Model and Serial Number	Centrifuge Remi, R8C DX SN-CHLC-5218
4	Equipment ID & Barcode	0920181
5	Date of purchase / Year of manufacture / Installation Date	28/05/2014
6	Warranty details (Yes/No)	NO warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC / CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	10/10/2023
9	Action taken	Checked & found motor, regulator, Anti vibrator, Control PCB and Timer defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	MOTOR - 9204/- Speed regulator - 2356/- Anti vibrator - 350/- Control PCB - 3128/- Digital Timer - 1294 Total cost = <u>16334RS</u>

13	Asset Value	23430/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	69.71%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	OEM Quotation attached Cyrix Report attached
16	Reasons for recommending the equipment as BER	Equipment aged 9.4 years, Repairing cost is above 65% of asset value. Considering the tender claus 5.3.4.1, we recommended the equipment for condemnation
17	Name & Signature of CYRIX Authority	LIBIL LAL L 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Recommended for BER.




 Consultant (Biomedical Engg)
 NHM (Arogyakeralam), Palakkad
 Signature of JC BM (NHM)

Date




 Signature of
 Superintendent / Medical Officer (i/c)
 SUPERINTENDENT
 Taluk Head Quarters Hospital
 Chittur-678 101, Palakkad

REMI SALES & ENGINEERING LIMITED

REMI GROUP

70/1750, 5th Cross Road, Link Park, Pottakuzhi Road, Pachalam P.O., Kochi-682 012
 Phone: 0484-2401361, SERVICE HELPLINE: 9372773715
 Email: kochi@remigroup.com / service@remilabworld.com

054

Corporate Off: Remi House, 3rd floor, 11 Cama Indl. Estate, Goregaon (East) Mumbai 400 063, INDIA
 ● Board: +91 22 4058 9888 SALES SUPPORT: 8879448682 ● Email: sales@remilabworld.com

QUOTATION CUM PROFORMA INVOICE

Complaint No.		Customer Name & Address	Quotation No	PRINTED
Dated		Cyrin Health Care Pvt. Ltd.	Date	054
Engineer Name	VIPIN K M	Cochin	REF	Phone
Ref.	Phone	Libil	DATE	20/10/23
Email	libilcyrin@gmail.com	8891382686		
Model	RSC DX	CHLC-5218	Sl. No. / Mfg. Dt.	03/14

Sl. No.	Description	Part code	HSN/SAC	Rate	Qty	Amount (A)	GST % (B)	GST Amt (C)	Total Amt (A+C)
1	Electric motor	ML2500	850140	7800	1	7800	18%	1404	9204
2		0143							
3	Digital Timer	L22000	902920	1097	1	1097	18%	197.46	1294.46
4		017							
5	Speed Regulator	L22000	847990	1997	1	1997	18%	359.46	2356.46
6		441							
7	Pcb Rxc	L22000	902780	2651	1	2651	18%	477.18	3128.18
8		572							
9	Anti vibrator	L22000	401699	99	3	297	18%	53.46	350.46
10		595							
								Round off	0.44
								TOTAL (D)	16334
CALIBRATION CHARGES									
SERVICE CHARGES									
								TOTAL (E)	16334

Rupees Amount in Words: Sixteen thousand three hundred thirty Four only

TOTAL Rs. 16334

Terms & Conditions

- Spare parts are suitable only to use with REMI instruments & not for any other make.
- We will accept return of spares only if they are unused and not damaged. At our discretion, the returned spare will be replaced. The difference in prices, if any, will have to be borne by you.
- To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
- 100% Advance along with your order confirmation.
- Payment can be made by Cheque/NEFT in our Bank account. For Cash Payment you must collect Receipt from recipient of cash falling which REMI will not be held responsible for any discrepancy.
- Courier/transportation charges are payable at destination & not included in our price.
- Validity: The quotation is valid for 30 days from the date of this quotation / Proforma Invoice.
- Subject to Ernakulam Jurisdiction only.



BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

● A/C NO: 35699154671	● Bank Name: STATE BANK OF INDIA	● BRANCH: ELAMAKKARA
● IFS CODE: SBIN0013223	● MICR CODE: 682002053	● SWIFT CODE: SBININBB395
● GST NO: 32AAACR0421M1ZS	● PAN NO: AAACR0421M	● CIN NO: U31100MH198PL022314

Certified that this book contains 172
Pages serially numbered 1-172 only

**SUPERINTENDENT
GOVERNMENT HOSPITAL
CHITTUR-678 101.**

CI

Stock Book of Instruments.

Store Keeper,
Govt. Hospital, Chittur.

LAB Vol I
W/e/7 16-9-1998.

Verified
by SVO $\frac{ms}{8/6/99}$

Verified
by SVO $\frac{ms}{22/8/06}$

Tuesday

**SUPERINTENDENT
Taluk Head Quarters Hospital
Chittur-678 101, Palakkad.**



SUPERINTENDENT
Taluk Head Quarters Hospital
Chittur-678 101, Palakkad.

INDEX.

1

SERIAL NO.	PARTICULARS.	PAGE NO.
1.	Microscope	5.
✓ 2.	Chemical Balance.	6.
✓ 3.	Electrical Centrifuge.	7.
✓ 4.	Centrifuge. Hand operated.	8.
5.	Urinometer.	9.
6.	Haemoglobinometer.	10.
7.	Counting Chamber	11.
8.	Spirit Lamp.	12.
✓ 9.	Conical flask. Conical.	13.
✓ 10.	Glass Funnel.	14.
✓ 11.	Measuring Cylinder 100 ml.	15.
12.	Volumetric pipette. 10 ml.	16.
13.	Micro acid glass centrifuge tube.	17.
14.	Syringe. 10 ml.	18.
15.	Syringe. 2 ml.	19.
16.	Test tube. Large.	20.
✓ 17.	Tripoche stand.	21.
✓ 18.	Inise gauge.	22.
✓ 19.	Test tube holder.	23.
20.	KSR. Stand.	24.
21.	Test tube small.	25.
22.	Test tube stand.	26.
23.	VDRL, Slides	27.
24.	Wooden slide Box.	28.
25.	Round bottom flask 500 ml.	29.
26.	Pebry disc.	30.
✓ 27.	Timer.	31.
28.	Revolving stool stool.	32.
29.	Glass, Mortar and pistle.	33.
30.	Dropping Bottle.	34.
31.	KSR. Pipette.	35.
32.	Microscopic Cover glass.	36.
33.	Microscopic Slides	37.



Electrical Centrifuge + bearing out head.

Date	Description	Received	Issue	Balance
16.9.98	b.f. from Page 16 Old book (1 unserviceable)			2
4/7/10	Received from the store.	1		3
4/7/12	Issue to the Lab.		1	
13/9/13	Received from store.	1		4
13.9.13	Issue price 80,770/- 8x 24,109		1	
08.5.14	Received from store	1		5
08.5.14	Issue to lab. 16x15 M/ 2 unserviceable.		1	3
19.10.2015	Centrifuge Model - RSC D x + R - 813 - make REMI. issued to ozhala pathy.		1	2
13/7/18	1 unserviceable			01

Carry over to Page
of unserviceable article
page No: F/8

7/9/18 carried over to page No: 8B of
Substock register of Instruments and furniture II



Handwritten signature and date: 02/10/18
SUPERINTENDENT
Taluk Head Quarters Hospital
Chittur 510 011, P.N. Akkad



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0920181







REMI CENTRIFUGE CE

CAT NO./DATE:	R8C DX	05/14
SERIAL NO	CHLC-8218	
RPM	5250	AMPS: 1.55
SUPPLY	220/230V, 50Hz, 1ϕAC	

REMI ELEKTROTECHNIK LIMITED
(INSTRUMENTS DIVISION)
VASAI - 401 208 (INDIA)



Taluk Head Quarters Hospital
Chittur-678 101, Palakkad



Handwritten: 05/10/14
SUPERINTENDENT
Taluk Head Quarters, Chittur
Chittur-678 101, Palakkad

**Fwd: QUOTATION CUM PROFORMA
INVOICE FOR CYRIX HEALTHCARE
ERNAKLAM-20-10-2023**

From: Jayesh Krishnan jayeshcyrix@gmail.com

To: Blesson Jose zm2.klbemp@cyrix.in

Sent: Wednesday, 6 March, 10:47 am

Dear Blesson,
FYI

----- Forwarded message -----

From: **instservicekochi**

[<instservicekochi@remigroup.com>](mailto:instservicekochi@remigroup.com)

Date: Wed, Oct 25, 2023 at 2:09 PM

Subject: QUOTATION CUM PROFORMA INVOICE
FOR CYRIX HEALTHCARE ERNAKLAM-20-10-
2023

To: [<jayeshcyrix@gmail.com>](mailto:jayeshcyrix@gmail.com),

[<libilcyrix@gmail.com>](mailto:libilcyrix@gmail.com)

Cc: [<raman@remilabworld.com>](mailto:raman@remilabworld.com),

[<kochisales.hcd@remigroup.com>](mailto:kochisales.hcd@remigroup.com),

[<kochi@remigroup.com>](mailto:kochi@remigroup.com)

Dear Sir,

Greetings from REMI GROUP.

As requested attached quote for your reference.

**PLEASE RELEASE THE PAYEMENT IN
ADVANCE(100% ADVANCE) .**

Regards,

Arya Thomas

0484-2401361.

--

Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod, Palakkad
PH:- 7593847156

jayeshcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin

-38 | Kerala | India

**KERALA || KARNATAKA || TAMIL NADU || UTTAR
PRADESH**