



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	FHC PERUVALLUR HOSPITAL
3	Name of Equipment with Make, Model and Serial Number	Equipment: Steriliser MAKE: Life care MODEL: 1500w SI: 002000
4	Equipment ID & Barcode	1031688
5	Date of purchase / Year of manufacture /Installation Date	30-06-2000
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29-01-24. - Trollfree call
9	Action taken	Checked the Steriliser found body fully rusted and leaking, 1.5w coil damaged, coil holder with power cord damaged, coil and holder can replace but Steriliser body is too weak for done sterilisation purpose
10	Present status of the equipment (Fully damaged/ partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	-----



13	Asset Value	RS:3800 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	-----
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending The equipment as RBER	The Steriliser installed on 30/06/2000, with an asset cost of rupees 3800, is in a critical condition. Extensive rusting and leaking compromise its structural integrity, posing risk to safety and reliability during operation. Give its two-decade service, hence we are recommending immediate condemnation to prevent operational hazards and uphold sterilisation efficacy. The equipment continued use jeopardize both safety standards and the effectiveness of sterilisation processes. so we recommend equipment for condemnation
17	Name & Signature of CYRIX Authority	Shameer pe 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**



Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified with full body rust, coil holder with power cord damage & leakage, long term use, Beyond economic repair.

Signature of JC BM (NHM)

Jr. Consultant Biomedical Engineer  
NHM Malappuram

\* Attach Photograph

	<p><b>MEDICAL OFFICER</b> C.H.C Peruvallur Parambilpeedika ( Po) 676 311 Malappuram Dist.</p>	
	<p>Date 12.2.24.</p>	<p>Seal</p>



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 171122

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : FHC</p> <p>Address : PERUVALLUR MADAPPURAM</p> <p>Ph : 9947135303</p>	<p>Call Registration Date : 29.01.24</p> <p>Caller ID : 93131</p> <p>Date of Visit : 30.01.24</p> <p>Asset No. : 10.216.88</p> <p>EQPT Name : Steriliser</p> <p>Manufacture : NA Model : 1500W</p> <p>S. No. : NA Dept. : OP</p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Baby leaking, coil connector and coil damaged

**Action Taken :** checked the steriliser found full body rusted and leaking, 1.5W coil damaged, coil holder and power cord damaged coil and coil holder can replace but steriliser body is too wear for clean sterilisation purpose

Completed  Date : 5/2/24 Time : 12:00pm Spare Required

Spare Replaced  Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.	A	NA		
3.				

Cyrux Engineer	Date	Start Time	End Time
Shameer	5/2/24	10:00am	12:00pm

**Customer Remark** Completed  Pending



<p><b>Service Engineer Name :</b> Shameer</p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> 5/2/24</p> <p><b>Contact Number :</b> 907252055</p>	<p><b>Customer Name :</b></p> <p><b>Signature :</b></p> <p><b>Date :</b></p> <p><b>Contact Number :</b></p> <p><b>Designation :</b></p> <p><b>Hospital Seal :</b></p>
---	---

MEDICAL OFFICER

C.H.C. Peruvallur  
Parambilpeedika (Po) 676 317  
Malappuram Dist.

05/02/24

From

Nursing officer  
CHE Peruvallur.

To the medical officer  
CHE Peruvallur.

(Through proper channel)

Madam ;

30/6/2000 ന്നിൽ KMSCL Supply ൽ -  
ചാനലിലെ ചെറിയ ഒരു steel sterilizer ഉപയോഗിക്കാൻ  
പറ്റാത്ത വിധത്തിൽ ആരുമുപയോഗിച്ചു കണ്ടെടുത്തു.  
അയൽ condemnation ചെയ്യുന്നതിന് ഭയപ്പെട്ടു നടപടി  
എടുക്കണമെന്ന് അപേക്ഷിക്കുന്നു.

നിന്ന്

30/1/2024  
Peruvallur.

MEDICAL OFFICER  
C.H.C Peruvallur  
Parambilpeedika (Po) 676 317  
Malappuram Dist.

21/01/2024  
  
Peruvallur.



STOCK BOOK OF .....

Name of Article. G.R.A.V. CLOTH .....

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward					
3-10-91	228	Recd from MIS/Primary ECMS. 1914 Das Per. 191-116. 208/91-92 dt. 3-10-91			10 NOS	<u>LS</u>	
21/5/92		Issued		10	NIL	<u>LS</u>	
		C/O Page VOL V Instrument Register					
		STERILISER SYRINGE ELECTRIC					
		20.6.2000. 3300-01 Recd from DN S. Tevar (50x200x150 page) R. 2645/2					
31-6/2000		Issued to CP		1	NIL		
19.6.02	12/02-03	Recd from DPO Das. May...		1			RS/3104
6/8/2003		Issued		1	NIL	<u>R</u>	
		C/O Page No 44 VOL V of Instruments Equipments					




*[Signature]*  
Dr. Fousia

*[Signature]*

MEDICAL OFFICER  
C.H.C Peruvallur  
Parambodiya (Po) 676 317  
Malappuram Dist.


Jr. Consultant Biomedical Engineer  
NHM Malappuram



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (JEMP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890815225) 1031688


*[Handwritten signature]*  
21/2/24

Jr. Consultant Biomedical Engineer  
NHM Malappuram



Dr. Fousia.M.V. *[Handwritten signature]*

MEDICAL OFFICER  
C.H.C Peruvallur  
Parambilpeedika (Po) 676 317  
Malappuram Dist.



CYRUS MEDICAL EQUIPMENT MAINTENANCE  
 PROGRAMME (2016)  
 TOLL FREE NO-1800-425-7669  
 BAR CODE (8004-800015225)  
 1031688



*Handwritten signature*  
 2/1/2014

Sr. Consultant Biomedical Engineer  
 NHM Malappuram



Dr. Fousia M.V. *Handwritten signature*

MEDICAL OFFICER  
 C.H.C Peruvallur  
 Parambilpeedica ( Po) 576 317  
 Malappuram Dist.