


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Palakkad
2	Name of Hospital	THQH ALATHUR
3	Name of Equipment with Make , Model and Serial Number	NEBULIZER MAKE : OERTEL GBR MODEL : MCN – S600A S/L: 0131365
4	Equipment ID & Barcode	94082 - 0920917
5	Date of purchase/Year of manufacture /Installation Date	08-07-2014
6	Warranty details(Yes/No)	NO WARRANTY
7	*AMC/CAMC Period agreed at the time of purchase	NO CAMC / AMC
8	Date of breakdown(Date of registration of complaint through email/toll free)	03-02-2024 (Toll Free)
9	Action taken	Checked and found that motor burned and piston rod broken.
10	Present status of the equipment(Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Not Recommending.
12	Cost of spares(specify parts and cost)	NA

13	Asset Value	□1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	NA
15	Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not)	Cyrix Service report attached.
16	Reasons for recommending the equipment as BER	Checked and found that motor burned and piston rod broken. Machine installed on 8-07-2014 and outlive nine plus years. Spare quote not submitted spare are not available in market and since recommending for condemnation.
17	Name & Signature of CYRIX Authority	DIVIN K 


*Not mandatory #Based on the period of life and value as per the BER guidelines

*Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor burned and piston rod broken and the machine installed on 8/7/2014. Spares are not available in the market so recommended for BER.




17/02/24
Signature of JCBM(NHM)

Date




Signature of
Superintendent Medical Officer (i/c)
T.H.O. HOSPITAL
ALATHUR



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 175498

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THH ALATHUR</u></p> <p>Address <u>ALATHUR</u> <u>PALAKKAD</u></p> <p>Ph : <u>9446139336</u></p>	<p>Call Registration Date : <u>03/02/2024</u></p> <p>Caller ID : <u>94082</u></p> <p>Date of Visit : <u>05/02/2024</u></p> <p>Asset No. : <u>0920917</u></p> <p>EQPT Name : <u>NEBUZIZER</u></p> <p>Manufacture <u>ORTEL</u> Model : <u>MCN-5600A</u></p> <p>S. No. <u>0131365</u> Dept. <u>LABOUR WARD</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked the machine. Found that motor burned and piston rod broken. Need to replace compressor motor and piston rod for further checking working condition of equipment

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>DIVIN K</u>	<u>9/02/24</u>	<u>11:30AM</u>	<u>12:30PM</u>

Customer Remark _____ Completed Pending

<p>Service Engineer Name : <u>DIVIN K</u></p> <p>Signature : </p> <p>Date : <u>09/02/24</u></p> <p>Contact Number : <u>7593847155</u></p>	<p>Customer Name : <u>BISI N R</u></p> <p>Signature : <u>SNG</u></p> <p>Date : <u>09/2/24</u></p> <p>Contact Number : <u>9446139336</u></p> <p>Designation : _____</p> <p>Hospital Seal : _____</p>
---	---

SUPERINTENDENT
 T.H.O. HOSPITAL
 ALATHUR

NEBULIZER

	Qty	Unit	M-Stock	Sign
17:4:21	B.F. Gump old clock book B. P.No:212	1	1 NIL	1
	Qty. 8/7/2014			
	Barcode 67091011 - 0 7120917			

UX .



5000
 SUPERINTENDENT
 H.O. HOSPITAL
 ALATHUR

