


459

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	THSH ALATHUR
3	Name of Equipment with Make, Model and Serial Number	DIGITAL BP APPARATUS MAKE: OMRON MODEL: HEM 7120 S. NO.: 20140837278061
4	Equipment ID & Barcode	75809, 0920921
5	Date of purchase / Year of manufacture / Installation Date	21/07/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC / NO CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	20/10/2023, TOLL FREE
9	Action taken	Checked and found that Display, main board, NIBP unit and 4 numbers of batteries faulty
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	2200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICED REPORT
16	Reasons for recommending the equipment as BER	Checked and found that LCD display, main board, BP bulb, BP cuff, BP bladder BP rave and three numbers of batteries faulty. Machine installed on 21/07/2018. and outlive five plus years. Spare quote not submitted spares are not available in market and since recommending for condemnation
17	Name & Signature of CYRIX Authority	DIVINIK 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Main board, LCD Display, BP bulb, BP cuff and bladder damaged. Hence recommended for BER.

Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad

16/07/18
Signature of JC BM (NHM)

Date _____ Seal _____



For Y...

SUPERINTENDENT
J.H.O. HOSPITAL
Signature of _____
Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : **164117**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THQ ALATHUR</u></p> <p>Address <u>ALATHUR</u> <u>PALAKKAD</u></p> <p>Ph : <u>99 4393 36</u></p>	<p>Call Registration Date : <u>20/10/23</u></p> <p>Caller ID : <u>75809</u></p> <p>Date of Visit : <u>22/10/23</u></p> <p>Asset No. : <u>0920921</u></p> <p>EQPT Name : <u>DIGITAL BP APPARATUS</u></p> <p>Manufacture <u>OMRON</u> Model : <u>HEM 7120</u></p> <p>S. No. <u>20140837A(6)</u> Dept. <u>LABOUR ROOM</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Machine not switching ON.

Action Taken : Checked the machine. Found that Display, Main board and pump motor is faulty. Also Found that NIBP cuff leakage. Need to replace these spare for further checking working condition of equipment.

Completed Date : 2/11/23 Time : 1:30 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>DIVINIK</u>	<u>8/11/23</u>	<u>12:30 PM</u>	<u>1:30 PM</u>

Customer Remark Completed Pending

Service Engineer Name : DIVINIK
Signature : [Signature]
Date : 8/11/23
Contact Number : 7593847155

Customer Name : BIJIN.R
Signature : [Signature]
Date : 8/11/23
Contact Number : 9446139336
Designation : 3 NO
Hospital Seal :





B. P. APPARATUS - DIGITAL

		Ad	Issue	Balance	Main stock Balance	St
8/3/21	RD From the store	80093940	2 ✓	2 ✓		2
9/3/21	Issued	Barcode 80093939		2	NIL	2
17/4/21	B.F From Dage No: 132 Old Book Bi		3	3	NIL	5
	(Ad 23/1/2018 - 2 Price Delate Desayathu from					
	Rs 2000 per 1 No, 26/11/2018 - 1					
27/1/23	RD in 6th BOLAION WARD		1			6
	C 1 187.					
9/11/23	Rtd from store		1			7
10/11/23	Issued for use			1	NIL	7



Tampath. K. J. 8/20



Jayanth K Jayanth
SND
For SND Biji-N-R

Handwritten text in blue ink, including a signature "Jayanth K Jayanth", the acronym "SND", and the phrase "For SND Biji-N-R".