



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 174832

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>District hospital</i>	Call Registration Date : <i>01.01.2024</i>
Address <i>Karhangad</i>	Caller ID : <i>88329</i>
<i>Kasaragod</i>	Date of Visit : <i>03.01.2024</i>
Ph : <i>9495039515</i>	Asset No. : <i>1410437</i>
	EQPT Name : <i>CARM</i>
	Manufacture <i>Allengers</i> Model : <i>HF-49</i>
	S. No. <i>241103157-C</i> Dept. <i>OT</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Not working*

Action Taken : *Checked the machine and found machine tube head pulse 3.5kw, Inverter, mother board and memory module with accessories are defective. Need to to replace these spares for further checking of the equipment.*

Completed Date : *27.01.24* Time : *4.30pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Abhinav</i>	<i>27.01.24</i>	<i>3.30pm</i>	<i>4.30pm</i>

Customer Remark Completed Pending


Service Engineer Name : <i>Abhinav</i>	Customer Name : <i>Ramegh P. P.</i>
Signature : <i>[Signature]</i>	Signature : <i>[Signature]</i>
Date : <i>27.01.2024</i>	Date : <i>31.01.24</i>
Contact Number : <i>8943392989</i>	Contact Number : <i>9495039515</i>
	Designation : <i>Nsg officer</i>
	Hospital Seal : <i>[Seal]</i>

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	DH KANHANGAD
3	Name of Equipment with Make, Model and Serial Number	C ARM MACHINE MAKE: ALLENGERS MODEL: HF-49 S/N: 2K1103157-C
4	Equipment ID & Barcode	88329/ 1410437
5	Date of purchase / Year of manufacture / Installation Date	06-05-2011
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/CAMC Period agreed at the time of purchase	NO AMC
8	Date of breakdown (Date of registration of complaint through email/Toll free)	01-01-2024
9	Action taken	Checked the machine, found machine tube head pulse 3.5kw, inverter, mother board and memory module with accessories are defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No Recommendation
12	Cost of spares (specify parts and cost)	Tube head pulse 3.5kw- 184800 Inverter- 201600 Mother board- 44800 Memory module with accessories- 224000 Total- 655200

13	Asset Value	778375
14	#PercentagevalueofthecostofspareswithrespecttoCostofPurchase/Asset Value	84.175%
15	Abstract of Service Report provided bythe OEM/ Authorized Service Provider/CYRIX(AttachedorNot)	Cyrix service report and OEM quotation attached.
16	Reasons for recommending theequipment as BER	Checked the machine, found machine tube head pulse 3.5kw, inverter, mother board and memory module with accessories are defective. Machine installed on 6-05-2011 and aged up to 12+ years. Repairing cost of equipment is 84.17% with respect to asset value. Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation.
17	Name&SignatureofCYRIXAuthority	ABHINAV O 

*Notmandatory #BasedontheperiodoflifeandvalueaspertheBERguidelines


*AttachPhotograph

RemarksandRecommendationsofJuniorConsultant(Biomedical)NHM:
 Checked the machine and found Problem with machine tube head Pulse 3.5kw, inverter, mother board and memory module with accessories. Machine installed on 06/05/2011. OEM quotation attached. Repairing cost is above 80% of asset value.


 Signature of Junior Consultant
 Biomedical Engineer
 Kasargod Health Centre
 Kasargod


 Date 9/2/21
 Seal

 Signature of
 Junior Consultant/Medical Officer(i/c)
 Kasargod Health Centre
 Kasargod



സർവ്വകലാശാസ്ത്രം
 ജില്ലാ ആരോഗ്യപ്പുസ്തക
 കാഞ്ഞിരം
 കാന്തതോൾ ബിജു

Handwritten signature: R. Jeyaraj M.D.



Stock Register of Instruments & Equipments

C - ARM

Name of Article:

Date	Particulars of Receipt / Issue	Receipt	Issue	Balance	Cost	Initial	Name of Receiver	Designation	Signature of Receiver	Sub. Stock Page No.
1	2	3	4	5	6	7	8	9	10	11
3-5-11	Inv No: AM.SAL 00039 dt 14/4/11 KMSC L TVM			1						
					Rs 7,78,375/-					
6/5/11	Transferred to MOT		1	NIL						
				Balance	NIL					

C-ARM mobile image intensifier complete, combination (medical - equipment) comprises of the following:
 ① High frequency C-ARM control sr. no. 2151103147-C/, image intensifier 17" / Display 19" medical, grade-2 non / CCD module / circular orbit B:1, memory system (PS-100) / Trolley unit / C-ARM, Mobile stand / aluminium etc. carriage / cassette holder & accessories. price inclusive of all the charges like transportation, freight, insurance, medical charges & packing etc.

Received by Marykutty Rd user mandt & guaranty card
 dt 6/5/11
 page 135 vol IV stock
 of documentary register
 of MOT



M. Jeyaraj
 മിസ്റ്റർ ജേരാജ്
 കോണ്ടക്ടർ
 കോണ്ടക്ടർ
 കോണ്ടക്ടർ

Date: 15/01/2024

To,
Cyrix Healthcare (P) Ltd
30/641B, Vandipetta Rd, near SBI Bank,,
Poonithura, Maradu, Ernakulam,
Kerala-682038

Sub: Estimate for Tube head-Inverer-Mother Board and Memory unit

Dear Sir,

As per the above mentioned subject please find the below estimate.
Kindly go through the same and we are expecting your work order at the earliest.

SI No.	Item description	QTY	Price	Taxes	Amount in Rs.
1	TUBE HEAD PULSE [S.A] 3.5 KW FOR C-ARM HF WITH CHINES TUBE ASSEMBLED.	1	165000	12%	184800.00
2	INVERTER	2	180000	12%	201600.00
3	MOTHER BOARD	1	40000	12%	44800.00
4	MEMORY MODULE WITH ACCESSORIES	1	200000	12%	224000.00
Grand Total					655200.00
(Rupees six lakh fifty five thousand two hundred only)					

*Above information is not an invoice and only an estimate of services described above.If prices and tax change or additional parts, and labor are required, we will inform you prior to proceeding with the work. Kindly confirm your billing address (as mentioned above) and intimate us if any changes required. This is not included our service charges.

Kindly give an approval for the service in written (Work order) to following address/ Email id, so that we can perform the Job accordingly.

Payment: 100% in advance.

Our GSTIN number: 03AABCA2332H1ZE

For Allengers Medical Systems Ltd



Authorized Signatory
Allengers Medical Systems Ltd, 5B, 5th Floor, Hedge house
Edapally-Palarivattom Road-682025
9946669084/0484-4035626/ker.service@allengers.net

ALLENGERS MEDICAL SYSTEMS LTD.



Fwd: Spare Quotation needed - D H Kanhangad

From: Aneesh P B mklse@cyrix.in
To: Blesson Jose zm2.klbemp@cyrix.in
Cc: Abhijith KP abhijithcyrix@gmail.com
Sent: Monday, 15 January, 3:29 pm

Hi,

FYI

Kind Regards,

Aneesh P B

Deputy Manager - Service

Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah

Jn | Poonithura | Cochin -38 | Kerala | India

Ph: 9744555073 | M 9072522093 | W www.cyrixhealthcare.com

**KERALA || KARNATAKA || UTHAR PRADESH || TAMIL
NADU || KUWAIT || KSA || UAE**

From: RENJITH <ker.service@allengers.net>

Sent: Monday, January 15, 2024 3:23:25 PM

To: Aneesh P B <mklse@cyrix.in>

Subject: RE: Spare Quotation needed - D H

Kanhangad

Dear Aneesh,

Please find the attached quotation.

Thanks and regards

RENJITH RAMESH C | Executive, Office Operations | Cochin Branch

Allengers Medical Systems Ltd

5-B, 5th Floor, Hedge House (Pearl Opulence), Edapally-Palarivattom

Road, Ernakulam – 682 025

Tel (484) 4035626 | mobile +91 7087029374| Toll free 1800 266 8800

www.allengers.com | ker.service@allengers.net

Confidentiality Note: This e-mail and any attachments are confidential and may be protected by legal privilege. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this e-mail or any attachment is prohibited. If you have received this e-mail in error, Please notify us immediately by returning it to the sender and delete this copy from your system.

Thank you for your cooperation.

Please consider the environment before printing this e-mail!

From: Aneesh P B [<mailto:mklse@cyrix.in>]
Sent: Monday, January 15, 2024 11:03 AM
To: RENJITH
Subject: Re: Spare Quotation needed - D H Kanhangad

Dear Team,

Pls share estimate.

Kind Regards,

Aneesh P B

Deputy Manager - Service

Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah Jn |

Poonithura | Cochin -38 | Kerala | India

Ph: 9744555073 | M 9072522093 | W

www.cyrixhealthcare.com

**KERALA || KARNATAKA || UTHAR PRADESH || TAMIL
NADU || KUWAIT || KSA ||UAE**

From: Aneesh P B
Sent: 09 January 2024 12:54
To: RENJITH <ker.service@allengers.net>
Subject: Spare Quotation needed - D H Kanhangad

Dear Team,

Please share spare quotation to DH kanhangad
C-arm with SN: 2K1103157-C.

1. Tube head
2. Inverter module
3. Motherboard
4. Memory Module with lcd monitor.

Kind Regards,

Aneesh P B

Deputy Manager - Service

Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah Jn |

Poonithura | Cochin -38 | Kerala | India

Ph: 9744555073 |M 9072522093 | W

www.cyrixhealthcare.com

**KERALA || KARNATAKA || UTHAR PRADESH || TAMIL
NADU || KUWAIT || KSA ||UAE**