


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	WAYANAD
2	Name of Hospital	THQH VYTHIRI
3	Name of Equipment with Make, Model and Serial Number	Equipment-ESR ANALYSER Make-TRANSASIA Model-VESMATIC 20 Sn-2291
4	Equipment ID/Barcode	1220585//TICKET 83794
5	Date of purchase/ Year of manufacture/ Installation Date	29/11/2013
6	Warranty details (Yes/No)	NO
7	*AMC/CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/Toll free)	24/12/2023/TOLL FEE
9	Action taken	CHECKED FOUND PRINTER MODULE DEFECTIVE ENQUIRED SPARE FROM OEM
10	Present status of the equipment (Fully damaged /partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NO RECOMMENDATION
12	Cost of spares (specify parts and cost)	PRINTER REPLACEMENT KIT FOR VM 20(R30507750)-RS 66764/-


13	Asset Value	84000
14	*Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	79.4%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/CYRIX(Attached or Not)	QUOTATION FROM TRANSASIA CYRIX SERVICE REPORT
16	Reasons for recommending the equipment as BER	Equipment completed 8 years Repairing cost of equipment more than 65%(79.4 % with respect to asset value) Both criteria of RBER Met So we recommend RBER Ss per tender clause 5.3.14.1
17	Name & Signature of CYRIX Authority *Not mandatory #Based on the period of life and value as per the BER guidelines	NIYAS HADI AP  25/01/2024


*Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Printed Module electric. Spare Cost exceeds more than 80% of the total Equipment value.

25/01/2024


25/01/24
Signature of JCBM(NHM)
Bio-Medical Engineer
National Health Mission
(Arogyakeralam)
Wayanad-673122


Date: 27.1.24




Signature of Superintendent Medical Officer (S/O)
Superintendent Hospital
Head Quarters Hospital
Athiri Wayanad



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 153939

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>TIDOLI.</u></p> <p>Address <u>VYTHIRI.</u></p> <p>..... <u>WIDYANAD, KERALA.</u></p> <p>Ph : <u>7034278958.</u></p>	<p>Call Registration Date : <u>4/12/2023.</u></p> <p>Caller ID : <u>83794.</u></p> <p>Date of Visit : <u>6/12/2023.</u></p> <p>Asset No. : <u>1220585.</u></p> <p>EQPT Name : <u>ESR ANALYZER.</u></p> <p>Manufacture <u>TRANEJA</u> Model : <u>VEI MEXIC</u></p> <p>S. No. <u>2291.</u> Dept. <u>LAD.</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Printer module malfunctioning

Action Taken : Inspected the equipment found that the printer module is defective. Need to replace the printer module.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Dijul. C.M. ESO5</u>	<u>6/12/2023</u>	<u>10:00AM</u>	<u>12:00PM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Dijul. C.M.</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>6/12/2023</u></p> <p>Contact Number : <u>8086758545</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>[Signature]</u></p> <p>Contact Number : <u>944609015</u></p> <p>Designation : <u>[Signature]</u></p> <p>Hospital Seal : <u>[Signature]</u></p>
--	---

STOCK REGISTER OF

Name of Article

ESR ANALYSER

Main stock

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward					
13/11/15		Received from the store 01			01	RL	
		1. Date of installation: 13-11-15					
		2. Cost of equipment: Rs. 85,109/-					
		3. Warranty/AMC/ expiry - one year					
		4. Name: VESMATIC 20 (20 PLUS) AUTOMATIC					
		5. Model: 20 PLUS NEW.					
		Main stock					
		PS No: 103 EG 2					
14/11/15		Issued to Lab		INV.		RL	
		Carried over					



(Handwritten signature)

Superintendent
Talk Head Quarters Hospital
Maddur, Bangalore

QUOTATION

Bill To: CYRIX HEALTHCARE PVT LTD
 First Floor,petta Junction 30/641b, Poonithura,Ernakulam, Kerala, Pincode 682038 , , 673576


Ref. No : 2324QUOTE000283
 Date : Jan 6, 2024 11:37 AM
 Validity : 30 days
 TBM GSTIN : 27AAACTP2038C1ZT

Customer GSTIN No : 32AAFCC2499H2ZM

Ship to, Taluk Head Quarters Hospital, Vythiri
 H28V+3CR , , , 673576

SUB : QUOTATION FOR SUPPLY/REPLACEMENT OF SPARES IN INSTRUMENT MODEL: VESMATIC 20 AUTOMATIC , PRODUCT CODE 10334 / CE SR.NO.: 2291

With reference to the above, following Spares / Consumables are required to be replaced / supplied. Price/s of the same is / are given below, for approval:

SR. NO	Part Code	HSN Code	Description of Spares/Consumables	Qty	Unit Price	Discount%	GST%	Amount	
1	166094	90278990	PRINTER REPLACEMENT KIT FOR VM 20(R30507750)	1	56580		18	56580	
Merchant Name : TRANSASIA BIO-MEDICALS LTD								Total (Rs)	56580
SCAN AND PAY WITH UPI QR Merchant Name: TRANSASIA BIO MEDICALS LTD  Payee VPA: transasiabiomedicalsltd@citibank								Discount (Rs.) (-)	
								GST (Rs.)	10184.4
								Rupees in words: Sixty Six thousand Seven hundred Sixty Four rupees and Four paise only	Grand Total (Rs.)

TERMS AND CONDITIONS :

1. **Payment** : Full payment should be made in advance through NEFT / RTGS/ Demand Draft in favour of "Transasia Bio-Medicals Ltd" to **Bank A/C no.: 00600330002538 IFSC code: HDFC0000060, UPI ID : transasiabiomedicalsltd@citibank.** You can also pay us through **UPI apps like BHIM/ Google Pay/ Paytm/PhonePe** etc. by scanning the above mentioned code. Cash Payment is strictly prohibited. No employee of the TBM is authorized to collect cash and or issue receipt against collection of Cash.
2. **GST** : As applicable
3. **Other Cost / Local duties** : Customer's responsibility
4. **Delivery** : Within 1-2 weeks from the receipt of your work order with payment confirmation

Thanking You
 Your faithfully,
 For TRANSASIA BIO-MEDICALS LTD


RECEIVED & ACCEPTED
 CUSTOMER SIGNATURE & STAMP:
 CUSTOMER NAME: _____
 CONTACT NO: _____

TRANSASIA BIO-MEDICALS LTD.

Head Office: Transasia House. 8, Chandivali Studio Road, Andheri (E), Mumbai - 400 072. Tel.: (022) 4030 9000 Fax: (022) 2857 3030 Mumbai : Toll Free No. 1800 103 8226, Whatsapp No.: 8451048434, SMS No.: 9212433444

www.transasia.co.in

DOC NO : SC00-405 / ISS - 7





LUK HEAD QUARTERS

Handwritten signature and stamp in the bottom right corner.



5555
Vos medic 20
 50 Watt 230V 50 Hz
 CE
 2010 11 05 05 05 21 91 1



Handwritten signature in green ink
Taluk Head Quarters Hospital
Mariyati, Mayang



REQUIRED QUOTATION # 1220585 ESR ANALYZER @ THQH VYTHIRI , WAYANAD

4 messages

Akhilesh P <akhileshcyrix@gmail.com>

Tue, 2 Jan, 2024 at 8:49 am

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: Niyas Hadi AP <niyascyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>, transserve <transserve@transasia.co.in>, Rakesh Dhobi <r.dhobi@transasia.co.in>, Pooja Misal <m.pooja@transasia.co.in>

Dear sir,

Please share the price and stock availability of

1. Printer head assy
2. Display

ESR ANALYZER @ THQH VYTHIRI , WAYANAD

MAKE : TRANSASIA
MODEL : VESMATIC 20
S/N : 2291

Thanks & Regards

AKHILESH P
SPECIALIST ENGINEER - LAB NORTH ZONE
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Akhilesh P <akhileshcyrix@gmail.com>

Thu, 4 Jan, 2024 at 1:17 pm

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: Niyas Hadi AP <niyascyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>, transserve <transserve@transasia.co.in>, Rakesh Dhobi <r.dhobi@transasia.co.in>, Pooja Misal <m.pooja@transasia.co.in>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P
SPECIALIST ENGINEER - LAB NORTH ZONE
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>

Sat, 16 Mar, 2024 at 11:18 am

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: Niyas Hadi AP <niyascyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>, transserve <transserve@transasia.co.in>, Rakesh Dhobi <r.dhobi@transasia.co.in>, Pooja Misal <m.pooja@transasia.co.in>

Dear sir
Gently Reminder

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

SANDEEP DAS M <s.das@transasia.co.in> Sat, 16 Mar, 2024 at 12:13 pm
To: Akhilesh P <akhileshcyrix@gmail.com>
Cc: Niyas Hadi AP <niyascyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>, transserve <transserve@transasia.co.in>, Rakesh Dhobi Dhobi <r.dhobi@transasia.co.in>, Pooja Misal Misal <m.pooja@transasia.co.in>

Dear Akhilesh,

Please find the attachment

Get [Outlook for Android](#)

From: Akhilesh P <akhileshcyrix@gmail.com>
Sent: Saturday, March 16, 2024 11:18:54 AM
To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: Niyas Hadi AP <niyascyrix@gmail.com>; Blesson Jose <Zm2.klbemp@cyrix.in>; transserve <transserve@transasia.co.in>; Rakesh Dhobi Dhobi <r.dhobi@transasia.co.in>; Pooja Misal Misal <m.pooja@transasia.co.in>
Subject: Re: REQUIRED QUOTATION # 1220585 ESR ANALYZER @ THQH VYTHIRI , WAYANAD

[Quoted text hidden]