



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 165015

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility Phc Puthussery	Call Registration Date : 8.11.2023
Address Palakkad	Caller ID : 79083
Ph : 9995431894	Date of Visit : 8.11.2023
	Asset No. : 0942870
	EQPT Name : BP Apparatus
	Manufacture nil Model : nil
	S. No. nil Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Error - E4

Action Taken : checked and found that main board component damaged and not working. also found that air pump diaphragm broken. for further checking need to replace main board and pump.

Completed Date : 3.1.2024 Time : 12:50 Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. nil	nil		
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Dileep. A.D	3.1.2024	12:00 pm	12:50 pm

Customer Remark : Completed Pending

Service Engineer Name : Dileep. A. D
Signature : *[Signature]*
Date : 3.01.2024
Contact Number : 9048751999




Customer Name : Dr. Athira
Signature : *[Signature]*
Date : 3.01.2024
Contact Number : മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്
Designation : 9995431894
Hospital Seal : പുതുശ്ശേരി, കബീരക്കോട് വാലക്കാട് - 676 621

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Palakkad
2	Name of Hospital	Phc Pudukkery
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus
4	Equipment ID & Barcode	0942870, ID:79083
5	Date of purchase / Year of manufacture / Installation Date	28.05.2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/ CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	8.11.2023
9	Action taken	Checked and found that main board, air pump faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	need to replace main board air pump.
12	Cost of spares (specify parts and cost)	mainboard - 1200 motor - 600

13	Asset Value	2499
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	72%.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached.
16	Reasons for recommending the equipment as BER	checked the b2 apparatus found equipment completed above 3 years. repairing cost of equipment is 72% of asset value. both criteria [5.3.14] For Rber meet. So we recommending for condemnation
17	Name & Signature of CYRIX Authority	Dileep. 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER because the main barrel and air pump damaged.




3/1/24

Signature of JC BM (NHM)

Date



Signature of
Superintendent / Medical Officer (I/c)

മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്
കുടുംബാരോഗ്യ കേന്ദ്രം
പുതുശ്ശേരി, കണ്ണിങ്ങോട്
ഫോൺ - 678 624
കേരളം

DIGITAL BP apparatus.

28.5.2021

Received from pudussery
panchayat, Kanjikkode

Model no: Bpm-104

Rs: 2499/-

Mfg month: 8/2021

Mfg by: Shenzhen Bst

Technology electronics Co. Ltd

Stock in hand


Received

Issued

Balance

Signature

ബഹിഷ്കരണ ഭരണസമിതി
കുടുംബശ്രീ, കോളിംഗോട്
പാലക്കാട് - 678 621
കെ.എ.എ





AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL: armedicare07@gmail.com
MOB: 7994654108, 9446057350

OUR REF : AR/QTN/136-2023
DATE : 17/12/2023

TO
CYRIX HEALTHCARE
ERNAKULAM

Respected Sir,

SUB: QUOTATION FOR ELKO DIGITAL BP APPARATUS PARTS

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

In case of any further clarifications kindly feel free to contact us.

Thanking you and assuring you of our best attention and engineering services at all times.

Waiting for your earliest reply.

Yours truly,

For AR MEDICARE,
KANNUR.



TRUST US WITH YOUR NEEDS

OUR REF : AR/QTN/136-2023

DATE : 17/12/2023

SLNO:	DESCRIPTION	TOTAL _(INCL.ALL TAXES)
01	ELKO DIGITAL BP APPARATUS MAINBOARD MOTOR	1200/- 600
	TOTAL	1,800/-

TERMS & CONDITIONS OF SALE

Payment Terms : Complete Payment Within 30 Days

Delivery : Within 4 Days from the date of purchase order .



AR MEDICARE

TRUST US WITH YOUR NEEDS

REQUIRED QUOTATION FOR DIGITAL BP APPARATUS PARTS

2 messages

Sarang KM <arangcyrix@gmail.com>
To: ar medicare <armedicare07@gmail.com>

Tue, Mar 19, 2024 at 10:12 AM

Dear sir,

Kindly share the price and availability of ELKO DIGITAL BP APPARATUS main board and motor .

--

Thanks & Regards

Sarang.k.m
Divisional Incharge|North Zone
Bemp-Kerala
Mobile : **+91 7593847150**
Email : [arangcyrix@gmail.com](mailto:sarangcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com



Save Environment, Please do not print this e-mail unless it's essential

ar medicare <armedicare07@gmail.com>
To: Sarang KM <arangcyrix@gmail.com>


Tue, Mar 19, 2024 at 10:47 AM

Sir,
PFA**AR MEDICARE**

Thanks & Regards
AR MEDICARE
Room No:865,2nd floor
AL-RAHA Arcade
Thazhe Chovva, Kannur

Mob: 9446057350, 7994654108, 8138091108
Email: armedicare07@gmail.com
GSTIN :32BDYPV5030P1ZP

[Quoted text hidden]

 **ARQ136 CYRIX.pdf**
481K