


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	PHC KODUR
3	Name of Equipment with Make, Model and Serial Number	BIOCHEMISTRY ANALYZER MAKE = MERIL DIAGNOSTICS MODEL = CLINQUANT MICRO SN = 111710
4	Equipment ID/Barcode	97316 BAR CODE = 1042665
5	Date of purchase/ Year of manufacture/ Installation Date	27-04-2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/CAMC Period agreed at the time of purchase	NO AMC /CAMC
8	Date of breakdown (Date of registration of complaint through email/Toll free)	21-02-2024 TOLL FREE
9	Action taken	CHECKED THE MACHINE AND FOUND PROBLEM WITH RESULT VARIATION, THE PHOTOMETER BLOCK ASSEMBLY IS DEFECTIVE. ENQUIRE OEM FOR QUOTATION
10	Present status of the equipment (Fully damaged /partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NO RECOMMENDATION
12	Cost of spares (specify parts and cost)	1) PHOTOMETER BLOCK ASSEMBLY = RS 30500/- TOTAL = 30500 + GST 18% = 35990/-

J. AE
576 576

13	Asset Value	48000/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	74.9%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/CYRIX(Attached or Not)	CYRIX SERVICE REPORT AND OEM SPARE QUOTATION ATTACHED
16	Reasons for recommending the equipment as BER	CHECKED THE MACHINE AND FOUND PROBLEM WITH RESULT VARIATION, THE PHOTOMETER BLOCK ASSEMBLY IS DEFECTIVE. ENQUIRE OEM FOR QUOTATION .THE EQUIPMENT INSTALLED ON 27-04-2018 AND AGED UP TO 5 YEARS 10 MONTH, REPAIRING COST IS MORE THAN 74.9% OF ASSET VALUE. BOTH CRITERIA FOR RBER MET, AS PER THE TENDER CLAUSE 5.3.14.1, SO WE RECOMENT THIS EQUIPMENT FOR CONDEMNATION.
17	Name & Signature of CYRIX Authority	SHIHABUDHEEN K 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

*Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified with the damage of Photometer ^{Block} Assembly. Spare cost is high and is beyond economic repair.



Jr. Consultant Biomedical Engineer
NHM Malappuram

Signature of JCBM (NHM)



Signature of Superintendent/Medical Officer (i/c)

Date: 01/03/24

01/03/24
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KODUR, KODUR P.O.
MALAPPURAM DISTRICT - 676 504



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 146276

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

<p>Health Facility <u>PHC KODUR</u></p> <p>Address <u>MALAPPURAM</u></p> <p>Ph : <u>9544274638</u></p>	<p>Call Registration Date : <u>21/2/2024</u></p> <p>Caller ID : <u>97316</u></p> <p>Date of Visit : <u>22/2/2024</u></p> <p>Asset No. : <u>1042665</u></p> <p>EQPT Name : <u>Biochemistry Analyzer</u></p> <p>Manufacture <u>Mari Diagnostics</u> Model : <u>Cliniquant</u></p> <p>S. No. <u>111710</u> Dept. <u>Laboratory</u></p>
--------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Biochemistry Analyzer not working.

Action Taken : checked the machine and found problem with result variations photometer black assembly is defective. need to replace the mentioned spare for checking further working condition of equipment

Completed Date : 22/2/2024 Time : 12 PM Spare Required



Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>photometer black assembly.</u>	1		
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Shibubdeen</u>	<u>22/2/2024</u>	<u>10 AM</u>	<u>12 PM</u>

Customer Remark Jr. Consultant Biomedical Engineer Completed Pending
NHM Malappuram

<p>Service Engineer Name : <u>Shibubdeen</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>22/2/2024</u></p> <p>Contact Number : <u>9072522046</u></p>	<p>Customer Name : <u>Fathima Shahanas A</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>22/2/2024</u></p> <p>Contact Number : <u>9544274638</u></p> <p>Designation : <u>Lab technician - gr-II</u></p> <p>Hospital Seal :</p>
---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------





Ref No MD / Mumbai / 2023 / 42012/0319
 Date 29th Feb 2024

To,

Cyrix Healthcare Pvt Ltd
Ernakulam , Kerala

Subject: Quotation For BIO-CHEMISTRY ANALYZER-CLINIQUANT MICRO CM-111710.

Dear Sir/Madam,

We would like to sincerely thank you for giving us the opportunity to collaborate with you through our products and services.

Meril Diagnostics Pvt. Ltd. is head quartered at Vapi, India and the Sales, Application and Service support departments are located in Mumbai. Our team is available across the country with more than 200 highly trained sales, application and service personnel. We firmly believe not only in providing world class products but also a consistent timely support to all our customers.

Attached please find herewith our quotation for your kind perusal. We hope you will find our offer in line with your requirements.

SAP Code	Instrument Name	Spares and Consumables / Contract (AMC & CMC)	Period	UoM	Price (Inc. tax)		
					Discounted Price (Rs.)	GST @ 18%	Total Amount (Rs.)
SISCM-0003	CliniQuant Micro	Photometer Block Assembly	1	Qty..	Rs.30,500/-	Rs.5,490/-	Rs.35,990/-

Request you to please contact us for any further queries.

Terms & Conditions:

- All Payments in advance (Vide DD in Favour of 'Meril Diagnostics Pvt. Ltd.' payable at Vapi).
- Quotation valid up to 30 days from the date of issuance.
- Tax rate as per GST.
- CMC - All spares except consumables (Lamp, Filters, Cuvettes, Sample cups, Tubings) are covered under CMC.
- Non-Functional unit will be not covered under CMC.

Thanks and Regards




Engineer Name: Mr. Gokul
Meril Diagnostics Pvt. Ltd.

Corporate Office: Survey No. 135/ 139, Bilakhia House, Muktanand Marg, Chala, **Vapi – 396191, Gujarat**

Mumbai Office: 602, Midas, Sahar Plaza, JB Nagar, Andheri East, **Mumbai – 400059, Maharashtra**

Call Centre: 18002663745 Direct Line: +91-22-30732545 E mail: diagnostics.csd@merillife.in



T +91 260 3052100
 F +91 260 3052125
www.merillife.com

01/03/24
 MEDICAL OFFICER
 PRIMARY HEALTH CENTRE
 KODUR, KODUR P.O.
 MALAPPURAM DT. - 676 504.
 Meril Diagnostics Pvt. Ltd.
 Survey No. 135/ 139, Bilakhia House
 Muktanand Marg, Chala
 Vapi - 396191, Gujarat, India

Jr. Consultant Biomedical Engineer
NHM Malappuram



[Handwritten Signature]

Jr. Consultant Biomedical Engineer
NHM Malappuram



[Handwritten Signature]
0103124

MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KODUR, P.O.
676 506

BIOCHEMISTRY ANALYSER (semi)
 (clin-quant micro)

3 16-4-19

Rd from store


1

1

"

Taken for lab use

1



[Signature]

Jr. Consultant Biomedical Engineer
 NHM Malappuram



entries of received Qty. should be written in Red Ink and with Batch No, Date of Mfg. & Exp.

[Signature]
 MEDICAL OFFICER
 PRIMARY HEALTH CENTRE
 KODUR, KODUR P.O.
 MALAPPURAM DT - 676 504



Quotation for Photometer assembly phc kodur

1 message

Gokul Das Aniyeri <gokul.aniyeri@merillife.com>
To: akhileshcyrix@gmail.com

Tue, 19 Mar, 2024 at 1:46 pm

Dear Akhilesh
Please find the attachment.

Thanks and Regards
Gokuldas
Thrissur HQ

CONFIDENTIALITY CAUTION

This communication (including any accompanying documents) is intended only for the use of the addressee(s) and contains information that is PRIVILEGED AND CONFIDENTIAL. Unauthorized reading, dissemination, distribution, or copying of this communication is prohibited. If you have received this communication in error, please notify the sender immediately by telephone [+91 260 2408000](tel:+912602408000) or email at IT.Services@merillife.com and promptly destroy the original communication without making a copy. Thank you for your cooperation.