


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Eroorkulam
2	Name of Hospital	CHC vergola
3	Name of Equipment with Make, Model and Serial Number	Centofuge. make:-REM1 model:-REC
4	Equipment ID/ Barcode	0730176
5	Date of purchase/ Year of manufacture	21/01/2008
6	Warranty details	N/A
7	*AMC/ CAMC Period agreed at the time of purchase	N/A
8	Date of breakdown/ Date of registration of complaint	20/10/2023
9	Action taken	checked Machine and Motor damage identified.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Motor Fully damaged. windings, bearing, Armature damage. also speed controller have faulty. and tested. So not recommend for service.
12	Cost of spares (specify parts and cost)	Nil / 2


13	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	9651 a.
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service Report
15	Reasons for recommending the equipment as BER	Electrify motor and major parts fully damaged. And. As per vendor condition 5.3.14.1 minimum year meets and Recommend for BER.
16	Signature of CYRIX Authority	

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

*** Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Submitted after 30 days.


2/12/23

Signature of JC BM (NHM)



Date 26/10/2023


Signature of
Superintendent Medical Officer (NHM)

**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



CYRIX
HEALTHCARE PVT LTD

No. : 153532

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Provider
Tender No. WO-37/2021-2022/698

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>CHK</u> Address : <u>CHK Vengola</u> <u>Ernakulam</u> Ph : <u>9946405722</u>	Call Registration Date : <u>20/10/2023</u> Caller ID : <u>75845</u> Date of Visit : <u>21/10/2023</u> Asset No. : <u>030176</u> EQPT Name : <u>Cadafeye</u> Manufacture : <u>DEMI</u> Model : <u>R8C</u> S. No. : <u>NA</u> Dept. : <u>LAN</u>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : chebel and Motor fix working checked
chebel carbon brushes in good condition checked
and winding issue checked motor stuck also checked
cut rotate motor manually. Totally damaged identified
Recommend for BGR. Installation date 21/10/2023

Completed Date : 21/10/2023 Time : 11:30 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Yedhukashna CSS</u>	<u>21/10/23</u>	<u>10:00 AM</u>	<u>11:10 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Yedhukashna CSS
 Signature : [Signature]
 Date : 21/10/23
 Contact Number : 8086147236

Customer Name : [Signature]
 Signature : [Signature]
 Date : 21/10/23
 Contact Number : 9946405722
 Designation : Lab technician
 Hospital Seal :



REMI CENTRIFUGE

CATALOGUE NO.

SERIAL NO.

220/230V. 50~1Ø AC AMPERE

REMI MOTORS

MUMBAI - INDIA - 400 058

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7869

**BAR CODE - (8004-800615225)
6730178**



38
5/11

b/p from sub stock register vol II page No: 38

Code No: 68275194 Receipt Unused Balance

21-1-08	store	<u>CENTRIFUGE</u>	1	1	2
21-1-08	lab		1	NIL	29

MANUFACTURER - REMI

PRODUCT NAME /

MODEL - REMI MOTORS

SERIAL NO: FH KL 2973

INSTALLATION DATE - 21-01-2008

STOCK LEDGER - P 33/Norm Vol 8

NOT working

Amo
/ RT

UNIQUE ID := CHC/UGLA/KAM/CF-12