


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	Eonakulam
2	Name of Hospital	PHC Choukara
3	Name of Equipment with Make, Model and Serial Number	Bp apparatus OMRON ITEM # 120
4	Equipment ID/ Barcode	0743180
5	Date of purchase/ Year of manufacture	18/8/2018
6	warranty details	NA
7	*AMC/ CAMC Period agreed at the time of purchase	NA
8	Date of breakdown/ Date of registration of complaint	6/11/2023
9	Action taken	checked and Board issue found.
10	Present status of the equipment (Fully damaged / partially damaged)	Partly damaged
11	Recommendations for repair (required service details)	Main board issue identified. need to replace board
12	Cost of spares (specify parts and cost)	Nil


13	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	2200
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service Report
15	Reasons for recommending the equipment as BER	BP apparatus Main Board issue identified sensors not available as per Vendor condition 5.3.19.1 Minimum year of working meets
16	Signature of CYRIX Authority	

*Not mandatory #Based on the period of life and value as per the BER guidelines

• Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


 2/12/23
 Signature of JC BM (NHM)


 Superintendent/Medical Officer (i/c)
 ശ്രീമൂലനാഥൻ പി.ജി., ഫീൽഡ്-693 580

Date 13/11/2023





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 163315

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC</u></p> <p>Address : <u>PHC Changan</u> <u>Ernakulam, Ernakulam</u></p> <p>Ph : <u>9061190013</u></p>	<p>Call Registration Date : <u>06/11/2023</u></p> <p>Caller ID : <u>78585</u></p> <p>Date of Visit : <u>13/11/2023</u></p> <p>Asset No. : <u>0743180</u></p> <p>EQPT Name : <u>Bpappedaly</u></p> <p>Manufacture <u>Dmson</u> Model : <u>HC07120</u></p> <p>S. No. <u>NA</u> Dept. <u>Public health</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

Bpappedaly not working error shown.

Action Taken :

Checked equipment and error found error changes while each time (show default error) And found the equipment board issue identified Board not available recommended for BR

Completed Date : 13/11/2023 Time : 2:00pm Spare Required


Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.	<u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Yeelhubrashe Cy</u>	<u>13/11/23</u>	<u>2:00pm</u>	<u>2:00pm</u>

Customer Remark Completed Pending

Service Engineer Name : Yeelhubrashe Cy
Signature : [Signature]
Date : 13/11/2023
Contact Number : 8080197288



Customer Name : Sini Paul
Signature : [Signature]
Date : 13/11/23
Contact Number : 9400466625
Designation : Jr. PHN
Hospital Seal :



OMRON HEALTHCARE



MAX

OMRON
AUTOMATIC BLOOD PRESSURE MONITOR
MODEL HEM 7120
RATING DC 6V 4W



SERIAL NO

OMRON HEALTHCARE Co., Ltd.
KYOTO JAPAN

OMRON HEALTHCARE (EUROPE) B.V.
Scheidtstr. 132, 1332 CA Amsterdam
THE NETHERLANDS

OMRON HEALTHCARE (INDONESIA) PT.
JAYA PRAKASA
OMRON (INDONESIA) PT.
SINGAPORE PTE LTD



MODEL 51
SERIAL 220

OMRON HEALTHCARE CO. LTD.

ORIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0743180**



HEM-CR24
HEM-RML31



OMRON HEALTHCARE CO. LTD.

OMRON

HEM 7120



SYS
mmHg

DIA
mmHg

PULSE
min



START
STOP

Name of Article 62/CHINA

Date	No and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver
		Brought forward				
		Received from LHI	1	-	1	
		(Digital Ap- Apparatus)		1	NIL	
		Received from LHI	1	-	1	
				1	NIL	