



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 171608

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility P.H.C. Address Cheranelloor Branakulam Ph : 9778779438	Call Registration Date : 4/12/2023 Caller ID : 83895 Date of Visit : 11/12/2023 Asset No. : 0741884 EQPT Name : BP Apparatus Manufacture Omron Model : HRM-7120 S. No. Na Dept. Public Health
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

checked the machine and found that circuit board and motor issue

Action Taken :

checked the machine and found that circuit board and motor issue, Need to replace the motor and circuit board, but spares are not available in market and are not repairable so BP Apparatus Recommended for RBR

Completed Date : 11/12/23. Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Stemiyo Tom	11/12/23	10:00 AM	11:00 AM

Customer Remark Completed Pending

Service Engineer Name : Stemiyo Tom
 Signature :
 Date : 11/12/23
 Contact Number : 9560852484

Customer Name : VISHNUP.B
 Signature :
 Date :
 Contact Number :
 Designation : staff Nurse, 9778779
 Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Eranakulam
2	Name of Hospital	PHC chevanelloor
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus Digital Make: OMRON Model: HRM-7120
4	Equipment ID/ Barcode	0741884
5	Date of purchase/ Year of manufacture	26-3-18
6	Warranty details	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/ CAMC
8	Date of breakdown/ Date of registration of complaint	4/12/2023
9	Action taken	checked the equipment and found that board and motor issue
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	checked the equipment and found that circuit board and motor issue. Need to replace the board and motor But spares not available in market and not repairable
12	Cost of spares (specify parts and cost)	Nil

13	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	2,200
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Report
15	Reasons for recommending the equipment as BER	checked the equipment and found that circuit board and motor issue. Need to replace the motor. But spares are not available in market and not repairable. So Machine recommended for RBER.
16	Signature of CYRIX Authority	Stam:yo Tom <i>Stam:yo Tom</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Dewally
01/01/24
Signature of JC BM (NHM)

Date 4/1/24



Akhila
Signature of For Dr. Sudhakar
Superintendent/Medical Officer (I/c)
ASSISTANT SURGEON
PRIMARY HEALTH CENTRE
C/KANALLOOR
ERNAKULAM

fx-11j



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)

0741884

OMRON

HEM 7120



SYS
mmHg

DIA
mmHg

PULSE
/min



START
STOP

BP Apparatus LCD
(super DLP diamond)

			ed	id	re	il
2054/	27/12/17	R/F Karunya medicine depot, ERM, Aluva IN. NO. 226 dt 21/12/17, B. NO. BPO 6134 E.D. - 12/2025, Rs. 4009.82/1	2104.91/1	2	2	ka
7 ka 3 ka nil ka no.	27/12/17	Issued to OP		2	nil	ka RO Prio: 9

BP Apparatus

4 ka 15/1/21	15/2/18	R/F KMSCL WH ERM [36] dt 15/2/18, SK. NO. 01435, M.O. - 11/17, unit rate - 2407.18/1	1	1	1	ka
6 ka 4 ka	15/2/18	Issued to OP		1	nil	ka Received - Page 9 Ameyan
(2) Anypro P. No. 44 5/18	2/5/18	R/F Karunya community Pharmacy ERM 71851, dt - 23/3/18, B. NO - 1304, IN. NO. Rate - 1069.93/1, xcel E.D. - 6/22	1	1	1	ka
7 ka 3 ka	2/5/18	Issued to Palliative nurse		1	nil	ka NGR 2019

Omron BO monitor HEM 7121

7 ka 3 ka Rd 7)	26/3/18	R/F Trichur Surgicals IN. NO. 8312, dt 26/3/18 Rs - 2688/- outside fund	1	1	1	ka Unit reg. 52
	26/3/18	Issued to m/c JPHN		1	nil	ka
	26/3/18	R/F Trichur surgicals IN. NO. 8311 dt 26/3/18 Rs. 2688/- outside fund, omron BO monitor HEM 7121 outside reg. 52	1	1	1	ka Reg JPHN

26/3/18 Issued to s/c chittoor

BP Apparatus Digital maxepn

27/1/21	27/1/21	R/F sevama surgicals machine drive, IN. NO. 1183 dt 27/1/21 Rs - 3360/-, NHM, model - BP-15 1506/1 -	2	2	2	ka OTM
27/1/21	27/1/21	Issued to OP		2	nil	ka Ameyan