



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 171609

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility U.P.N.C.</p> <p>Address Kadavandhara Ponakulam.</p> <p>Ph : 9745 433446.</p>	<p>Call Registration Date : 07/12/2023</p> <p>Caller ID : 844289</p> <p>Date of Visit : 7/12/23</p> <p>Asset No. : 0747097</p> <p>EQPT Name : BP Apparatus Digital</p> <p>Manufacture Easy Care Model : EC-9000</p> <p>S. No. 002554 Dept. Sub Centre.</p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
checked the machine and found that board and display issue

Action Taken :
checked the machine and found that board and display issue. Need to replace the circuit board and display but spares are not available market, and not repairable. So BP Apparatus recommended for RBER.

Completed **Date :** 7/12/23 **Time :** 10:00 **Spare Required**

Spare Replaced **Requested**


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Stanijo Tom	7/12/23	10:00 AM	11:00 AM

Customer Remark **Completed** **Pending**

Service Engineer Name : Stanijo Tom
Signature : *Stanijo Tom*
Date : 7/12/23
Contact Number : 7560852484

Customer Name : KJIMUL T
Signature : *[Signature]*
Date : 19/12/23
Contact Number : 9745433446
Designation : JPHN.
Hospital Seal :



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	UPHC Kadavanthra
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus Digital Make: Easy Care Model: EC-9000
4	Equipment ID/ Barcode	0747097
5	Date of purchase/ Year of manufacture	29/6/2020
6	Warranty details	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	07/12/2023
9	Action taken	checked the machine and found that board and display issue
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	checked the machine and found that board and display issue. Need to replace the display and board issue. But spares are not available in market and not repairable.
12	Cost of spares (specify parts and cost)	nil

13	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	2200/-
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Report
15	Reasons for recommending the equipment as BER	checked the machine and found that board and display issue. Need to replace the board and display. But spares are not available in market and not repairable so BR Apparatus is recommended for RBER.
16	Signature of CYRIX Authority	Shriyog Tom <i>Shriyog</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

• Attach Photograph

Remarks and Recommendations of Junior Consultant (Binnel) NHM:

Submitted after 30 days

Devathy
01/1/24

Signature of JC BM (NHM)

Date

Signature of
Superintendent/Medical Officer (i/c)

Medical officer in charge
UPHC Kadavanthra



Sales summary

Issue to staff nurse

7478

3400

1 0

1

Nil ~~Send~~

20 local purchase from SBAC

Speciality Drugs

2201303 2240

1 0

1

Nil ~~Send~~

Issue to staff nurse