


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Erodkulam.
2	Name of Hospital	phc Avoly.
3	Name of Equipment with Make, Model and Serial Number	BP apparatus Nucleus, No other Details
4	Equipment ID/ Barcode	0743019 87920
5	Date of purchase/ Year of manufacture	15/9/18
6	Warranty details	NA
7	*AMC/ CAMC Period agreed at the time of purchase	Yes
8	Date of breakdown/ Date of registration of complaint	29/12/23
9	Action taken	check and found solenoid valve compressor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged.
11	Recommendations for repair (required service details)	NAT serviceable.
12	Cost of spares (specify parts and cost)	NA

13	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	2200
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	yes.
15	Reasons for recommending the equipment as BER	check and found Solenoid valve and Compressor not working.
16	Signature of CYRIX Authority	

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

*** Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Solenoid valve Issue. Unable to repair the unit. BER submitted on time.

Dovally Consultant BME
10/01/23
Signature of JC BM (NHM)

Date




Signature of **സൂപ്പർ ഇൻ്റൻ്റ് മെഡിക്കൽ ഓഫീസർ (i/c)**
ആരോഗ്യ



**BIO MEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 157324

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>phc</u>	Call Registration Date : <u>29/12/23</u>
Address : <u>Avoly</u>	Caller ID : <u>87920</u>
<u>Ernakulam</u>	Date of Visit : <u>29/12/23</u>
Ph : <u>9744639446</u>	Asset No. : <u>0743019</u>
	EQPT Name : <u>Bp Apparatus</u>
	Manufacture <u>NA</u> Model : <u>NA</u>
	S. No. <u>NA</u> Dept. <u>observation room</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : unit not working

Action Taken : check the unit and check working, found issue with the solenoid valve and make air pump Mini. Recommended the unit for (RBER)

Completed Date : 29/12/23 Time : 2:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Akhil T Lenin</u>	<u>29/12/23</u>	<u>1:00 PM</u>	<u>2:00 PM</u>

Customer Remark Completed Pending



Service Engineer Name : <u>Akhil T Lenin</u>	Customer Name : <u>Smitha F Thankappan</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>29/12/23</u>	Date : <u>29/12/23</u>
Contact Number : <u>9805474326</u>	Contact Number : <u>9744639446</u>
	Designation : <u>Nursing officer</u>
	Hospital Seal :

15/9/18	Rec'd in stone (From flood cap)	1	3	Smay
	(Digital)			
237.2019	RFS (Mercury free)	2	5	Hub
		2	7	Soliman

nucleus

SYS

DIA

PULSE

kPa

mmHg

/min



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
HEALTHCARE PVT LTD PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0743019**



NTLTD
DER
SERVICE

23

27

CYRIX
HEALTHCARE PVT LTD
• FULADIA
• BP APPARATUS
• 0743019
• 1800-425-7669

C2
P20

