


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Parbhani Dist.
2	Name of Hospital	PHC Chennelkara.
3	Name of Equipment with Make, Model and Serial Number	UPS Make - NA Model - NA SN - NA
4	Equipment ID & Barcode	Id - 89303 ; 0340038
5	Date of purchase / Year of manufacture / Installation Date	17/09/2014
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	NA
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	06/01/2024
9	Action taken	Recommended for RBER
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Inverter PCB 7Ah Battery
12	Cost of spares (specify parts and cost)	24-230 Inverter PCB - 3304/- 7Ah Battery - 1751/- Service charge - 472/- Total - 5527/-


13	Asset Value	4500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	100%.
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Yes, Attached.
16	Reasons for recommending the equipment as BER	Checked the machine found that Inverter PCB and 7Ah Battery damaged. Inverter PCB NOT repairable. Replacement part cost in percentage of Asset Value as per the agreement in a single repair after the minimum year of functioning is way higher than the Limit Recommended for BER.
17	Name & Signature of CYRIX Authority	Rahul C.V 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Physically verified on 31/01/2024. Recommended for BER as per rule.




Signature of JC BM (NHM)



Date

Seal

Signature of
Superintendent / Medical Officer (i/c)


MEDICAL OFFICER
FAMILY HEALTH CENTRE
CHENNARAYANA



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 169409

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>Pte chennerkana</u></p> <p>Address : <u>Chennerkana</u> <u>palhonambhik</u></p> <p>Ph : <u>9747211535</u></p>	<p>Call Registration Date : <u>06/11/24</u></p> <p>Caller ID : <u>89303</u></p> <p>Date of Visit : <u>18/01/2024</u></p> <p>Asset No. : <u>0340038</u></p> <p>EQPT Name : <u>UPS</u></p> <p>Manufacture : <u>NA</u> Model : <u>NA</u></p> <p>S. No. : <u>NA</u> Dept. : <u>LSb.</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Customer reported problem was Machine not switching ON.

Action Taken : checked the machine found that Inverter PCB and 7Ah Battery gone damaged. Inverter PCB not responsible, Replacement part cost in percentage of Asset value as per the agreement in a single repair after the minimum year of functioning is way higher than the limit, Recommended for RBER.

Completed **Date** : 18/01/24 **Time** : 11:15 am **Spare Required**

Spare Replaced **Requested**

#	Description	Qty.	Part Number	PR Number
1.				
2.	<u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Rshul C.V</u>	<u>18/01/24</u>	<u>9.30 am</u>	<u>11.15 am</u>

Customer Remark **Completed** **Pending**

<p>Service Engineer Name : <u>Rshul</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>18/1/24</u></p> <p>Contact Number : <u>7025528912</u></p>	<p>Customer Name : <u>DV BIGN-RM</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>18/1/24</u></p> <p>Contact Number :</p> <p>Designation : <u>Mo/c</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
--	---



MEDICAL OFFICER
FAMILY HEALTH CENTRE
CHENNEERKANA

9747267568

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 16/01/2024

Contact: 8921950217

Email: aylara@ionbackups.com

GSTIN: 32CGUPM9450L1ZN

1KVA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT .LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	24 – 230 INVERTER PCB	1	2800.00	2800.00	18	3,304.00
2	7 AH BATTERY	2	684.00	1368.00	28	1751.00
3	SERVICE CHARGE	1	400.00	400.00	18	472.00
PAYABLE AMOUNT						5,527.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares, 28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN



SENDI AUTOANALYSER

Date	Receipt	Qty	Issue	B.I	Remarks
13/9/14	B/f from page no: 4 of Stock register Lab items	1		1	
	Dt of receipt : 12/9/14				
	Make : ROBONIK				
	Identification mark & sr no of goods 9702W0714RBR				
	Model Name : PRIETEST :				
	R. : 64760				

MEDICAL OFFICER IN CHARGE
FAMILY HEALTH CENTRE
CHENNEERKARA - NALLANIKUNNI
ELAVUMTHITTA P.O. - 689 625

APPE

ANALYZERS



BACKUP



NORMAL



POWER

KMSCL



KTPL

TOLL FREE NO:- 1800-270-0690

(8004) 8907579 - 70033176

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)



TOLL FREE NO:1800 - 425-7669

BAR CODE - (8004-890615225)
0340038

