


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	GH THRISSUR
3	Name of Equipment with Make, Model and Serial Number	Infusion Pump Make : Aspen Model :JSB-1200 Sr.No- 130400935
4	Equipment ID & Barcode	#99291-0810486
5	Date of purchase / Year of manufacture /Installation Date	12-03-2013
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	03-03-2024
9	Action taken	Checked the machine and found main board and drop sensor are complaint. Need to replace the main board and drop sensor for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	N/A



13	Asset Value	24400/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and oem EOL letter attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board and drop sensor complaint. Equipment installed on 12-03-2013 and aged up to 10 Years and 10 months. OEM informed that this machine is discontinued model and spare parts not available, oem letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	Bijo T Joy 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Discontinued model.


Jr. Consultant (Biomedical)
NHM, Thrissur



Signature of JC BM (NHM)



Superintendent / Medical Officer (i/c)



SUPERINTENDENT (Dy. DHO)
GENERAL HOSPITAL
THRISSUR - PIN: 689 001

Date



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 194810

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>GH</u>	Call Registration Date : <u>3-3-2024</u>
Address : <u>THRISSUR</u>	Caller ID : <u>99291</u>
Ph : <u>9656996536</u>	Date of Visit : <u>4-3-2024</u>
	Asset No. : <u>0810486</u>
	EQPT Name : <u>Infusion Pump</u>
	Manufacture <u>Aspen</u> Model : <u>JSB120.0</u>
	S. No. <u>130400935</u> Dept. <u>MICU</u>

Service Classification : Breakdown Call ☒ PMS ☐ Calibration ☐ Cust.Training ☐

Problem Identified : Drop Sensor error

Action Taken : checked the machine and found main board and drop sensor are complaint need to replace the main board and drop sensor for further checking and spare enquired to OEM.

Completed ☐ Date : 4-3-24 Time : 1:30 pm

Spare Required ☐

Spare Replaced ☐ Requested ☐

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer

Date

Start Time

End Time

Abhil Tony

4-3-24

11:30 AM

1:30 PM

Customer Remark

Completed ☐

Pending

Service Engineer Name :

Signature :

Date : 4-3-24

Contact Number : 9745328470

Customer Name : Sindhu Raghavan


Signature :

Date :

Contact Number : 9567169408

Designation : SAIO

Hospital Seal :





Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



Sarang K. M.

Divisional Incharge

Cyrix Healthcare (P) Ltd.
30/641B, Betta, Punithura,
Kochi, Kerala-682038

Date : 09-03-2024

Dear sir,

Subject: Spares for Infusion Pump (**Model:JSB1200**).

Ref: mail date on 04.03.2024.

We are in receipt of your mail dated on 04.03.2024. Please note this model Infusion Pump (**Model:JSB1200**) has been discontinued by our Principal Company.

Hence, the parts are not available.

Thanks & Regards,

For Aspen Diagnostics Pvt. Ltd.

Rajaram Krishnamurthy
National Service Manager