


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	GH THRISSUR
3	Name of Equipment with Make, Model and Serial Number	Infusion Pump Make : Aspen Model :JSB-1200 Sr.No- 13040114S
4	Equipment ID & Barcode	#99289-0810493
5	Date of purchase / Year of manufacture /Installation Date	11-03-2013
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	03-03-2024
9	Action taken	Checked the machine and found main board and drop sensor are complaint. Need to replace the main board and drop sensor for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	N/A



13	Asset Value	24400/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and oem EOL letter attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board and drop sensor complaint. Equipment installed on 11-03-2013 and aged up to 10 Years and 10 months. OEM informed that this machine is discontinued model and spare parts not available, oem letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	Byo T Jay 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Very old machine



[Handwritten Signature]
Signature of JC BM (NHM)



Superintendent / Medical Officer (i/c)



SUPERINTENDENT (Oy DH) GENERAL HOSPITAL
THRISUR - PIN: 680 001

Signature of seal

Date



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 194809

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>UJH</u>	Call Registration Date : <u>3-3-2024</u>
Address <u>THRISSUR</u>	Caller ID : <u>99289</u>
Ph : <u>9656996536</u>	Date of Visit : <u>4-3-2024</u>
	Asset No. : <u>0810493</u>
	EQPT Name : <u>Infusion pump</u>
	Manufacture <u>Aspen</u> Model : <u>JSB1200</u>
	S. No. <u>130401145</u> Dept. <u>MICU</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Drop sensor Error

Action Taken : checked the machine and found main board and drop sensor are complaint. Need to replace the main board and drop sensor for further checking and spare enquired to OEM

Completed Date : 4-3-24 Time : 1:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrix Engineer	Date	Start Time	End Time
<u>ABSIK TONY</u>	<u>4-3-24</u>	<u>11:30 AM</u>	<u>1:30 PM</u>

Customer Remark Completed Pending



Service Engineer Name : ABSIK TONY
Signature :
Date : 4-3-24
Contact Number : 9745328470

Customer Name : Sindhu Raghava
Signature :
Date : 4/3/24
Contact Number : 9567169908
Designation : SNIO
Hospital Seal :





Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



ISO 9001:2015
CERTIFIED COMPANY

Sarang K. M.

Date : 09-03-2024

Divisional Incharge

Cyrix Healthcare (P) Ltd.

30/641B, Betta, Punithura,

Kochi, Kerala-682038

Dear sir,

Subject: Spares for Infusion Pump (**Model:JSB1200**).

Ref: mail date on 04.03.2024.

We are in receipt of your mail dated on 04.03.2024. Please note this model Infusion Pump (**Model:JSB1200**) has been discontinued by our Principal Company.

Hence, the parts are not available.

Thanks & Regards,

For Aspen Diagnostics Pvt. Ltd.

Rajaram Krishnamurthy
National Service Manager