



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

**No. : 188310**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ..... <u>W&amp;C Hospital</u>	Call Registration Date : ..... <u>01/03/2024</u>
Address ..... <u>Palakkad</u>	Caller ID : ..... <u>99013</u>
Ph : ..... <u>9497284990</u>	Date of Visit : ..... <u>02/03/2024</u>
	Asset No. : ..... <u>0912486</u>
	EQPT Name : ..... <u>Nebulizer</u>
	Manufacture <u>INSTAPRO</u> Model : <u>PRONEB</u>
	S. No. ..... <u>NA</u> Dept. ..... <u>PICU</u>

**Service Classification : Breakdown Call**  **PMS**  **Callibration**  **Cust.Training**

**Problem Identified :** ..... NOT working

**Action Taken :** ..... checked the nebulizer and found nebulizer motor burned, piston broken. Need to replace nebulizer motor with piston for further checking of complaint.

**Completed**  **Date :** 02/03/24 **Time :** 11:00AM **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>LIBIL LAL-L</u>	<u>02/03/24</u>	<u>10:00AM</u>	<u>11:00AM</u>

**Customer Remark**  **Completed**   **Pending**

<b>Service Engineer Name :</b> <u>LIBIL LAL-L</u>	<b>Customer Name :</b> <u>Shreba Sebastian</u>
<b>Signature :</b> <u>[Signature]</u>	<b>Signature :</b> <u>[Signature]</u>
<b>Date :</b> <u>02/03/2024</u>	<b>Date :</b> <u>2/3/2024</u>
<b>Contact Number :</b> <u>8891382686</u>	<b>Contact Number :</b> <u>9497284990</u>
	<b>Designation :</b> <u>Shreba SNO</u>
	<b>Hospital Seal :</b>



**SUPERINTENDENT**  
Govt. W&C Hospital  
Palakkad




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	PALAKKAD
2	Name of Hospital	WOMEN AND CHILD HOSPITAL
3	Name of Equipment with Make, Model and Serial Number	NEBULISER MAKE : INSTA PRO MODEL : PRONEB SN : NA
4	Equipment ID/ Barcode	#99013 / 0912486
5	Date of purchase/ Year of manufacture/Installation Date	10/12/2019
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	01/03/2024 (Toll free)
9	Action taken	Checked the machine found nebuliser motor burned , piston brokened. Need to replace nebuliser motor with piston for further checking of machine.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS.1353/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the equipment found nebuliser motor and piston damaged. Equipment installed on 10/12/2019 and aged upto 4 years 3months. Qutation not submitted since Spare parts are not available on market. So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 

\*Not mandatory      #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and piston damaged here recommended for BER.

Consultant (Biomedical Engg)  
NHM (Arogyakeralam), Palakkad



Signature of JC BM (NHM)

Date: 22/3/2024



22/3/2024  
SUPERINTENDENT  
Govt. W & C Hospital  
Palakkad  
Superintendent/ Medical Officer (i/c)



# STOCK REGISTER FOR INSTRUMENT AND FURNITURE .

WEC HOSPITAL . PALAKKAD

PICU . From . 1-1-19

Reg. CODE . B . VOL . II

This is to certified that this Register contains  
1 to 264 pages (two sixty four) . which is serially numbered .

Verified up to 31/12/19

*[Signature]*  
NISHA

Store Verification Office  
Dist. Medical Office (Health)  
Palakkad

*[Signature]*  
22/01/2020

Checked up to 31/3/22

Store Verification  
Dist. Medical Office  
Palakkad

*[Signature]*  
13/5/22



SUPERINTENDENT  
Govt. W&C Hospital  
Palakkad



INDEX

S.No.	Name of Items	Page No.	S.No.	Name of Items	Page No.
1	ARTERY FORCEPS STRAIGHT	9	30	NEBULIZER	67
2	ARTERY FORCEPS (MOSQUITO)	11	31	NEEDLE HOLDER MED	69
3	AMBU BAG WITH MASK [medium]	13	32	NEEDLE BURNER READY	71
4	AMBU BAG WITH MASK [ADULT]	15	33	OXYGEN CYLINDER	73
5	B.P. APPARATUS.	17	34	OXYGEN STAND	75
6	C. PAP VENTILATOR.	19	35	OXYGEN FLOWMETER.	77
7	CRASH CART [SS]	21	36	O <sub>2</sub> KEY.	79
8	CURVED SCISSORS.	23	37	O <sub>2</sub> CHAMBER	81
9	CHESTLE FORCEPS.	25	38	PERFORATED VISITORS	83
10	BED SIDE LOCKER	27		CHAIR	85
11	BED PAN	29	39	PULSE OXYMETER (PEDIATRIC)	87
12	COT & MATTRESS.	31	40	PEDASTAL FAN.	89
13	DEFEBRILATOR & CARDIAC-MONITOR & PACING.	33	41	PULSE OXYMETER & ACCESSORIES.	91
14	FIRE EXTINGUISHER.	35	42	SIDE TROLLEY [SS]	93
15	INFUSION PUMP	37	43	S.S. BIN [15x11"]	95
16	IV STAND.	39	44	S.S. BIN [11x9"]	97
17	INJECTION TROLLEY.	41	45	SCISSOR STRAIGHT [M]	99
18	SALUOMETER. (C...)	43	46	S.S. TRAY (BIN)	101
19	INDUCTION COOKER.	45	47	S.S. TRAY SMALL	103
20	DIGITAL Thermometer.	47	48	STETHESCOPE.	105
21	KNEE HAMER.	49	49	SUCTION APPARATUS YHP	107
22	LARYNOSCOPE & BLADE ADULT.	51	50	SPOT LIGHT.	109
23	LARYNOSCOPE & BLADE PEDIATRIC.	53	51	STERILIZER [MCH]	111
24	HOSPITAL COT [MCH]	55	52	STEEL RACK.	113
25	MEDICINE TROLLEY.	57	53	SMALL TRAY (2x6).	115
26	MULTIPARA MONITOR & CAPNOURAM.	59	54	STRUCTURE TROLLEY	117
27	MATTRESS [MCH]	61	55	SABOL	119
28	MULTIPARA MONITOR & M... [MCH]	63	56	SHOE RACK.	121
29	M... [MCH]	65	57	ROOM AIRCONDITIONER	123
			58	3 BUCKET SYSTEM TROLLEY	





NEBULIZER.

67

date	From whom to whom to issued	Description	Cost	Received	Issued	Balance	main Sheet Balance	Total
	B/F Beg B	VO I page no: 21						
21/8/18.	From store.		7392	1		1		
31/8/18.	issued to use.				1	Nil	1	
22/10/19.	am servicable and unservicable Register	and Bought forwarded to Page no: 15		1	1	Nil	Nil	
31/10/19	Received from store.	(KIP)		1		1	1	
10/12/19	issue to use			1	1	Nil	1	
20/1/22	Received from store			1	1	Nil	2	2 ✓
30/1/22	issue to use			1	1	Nil	2	2 ✓
15/3/22	Received from store			2	2	Nil	4	4 ✓
20/3/22	issue to use			2	2	Nil	4	4 ✓



SUPERINTENDENT  
Govt. W & C Hospital  
Palakkad





SUPERINTENDENT  
Govt. W&C Hospital  
Palakkad

A handwritten signature in blue ink is written over the text 'SUPERINTENDENT Govt. W&C Hospital Palakkad' at the bottom right of the page.