


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	DISTRICT TB CENTRE MANJERI
3	Name of Equipment with Make, Model and Serial Number	Make: ROBONIK Model: PRIETEST TOUCH Serial Number :AT1400312RBK
4	Equipment ID/ Barcode	#100712 - 1012135
5	Date of purchase/ Year of manufacture/Installation Date	13/03/2012
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of Purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	12/03/2024
9	Action taken	Checked the machine & Found keypad & Primary Board defective. Need to replace These spare for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	NOT WORKING
11	Recommendations for repair (required service details)	No recommendation
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	64760/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX Service Report & OEM EOL Letter Attached
16	Reasons for recommending the equipment as BER	Checked the machine and found Key pad and Primary Board defective. The equipment was installed on 13-3-2012 Aged up to 12 Years and 11 months. Oem informed that this machine is obsolete model, end of life letter issued.As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	GOKUL KT 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified with the damage of primary board & keypad. Beyond economic repair. And the said Model of Machine is End of Life as per OEM.

Jr. Consultant Biomedical Engineer
NHM Malappuram



Signature of JC BM (NHM)


23/03/2024

Date: 27/03/24

Signature of Superintendent/Medical Officer (i/c)


DISTRICT TB OFFICER
DISTRICT TB CENTRE
MALAPPURAM - 676123



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 164759

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>DISTRICT TB</i>	Call Registration Date : <i>12/03/2024</i>
Address <i>CENTRE MANJERI</i>	Caller ID : <i>100712</i>
<i>MALAPPURAM</i>	Date of Visit : <i>13/03/2024</i>
Ph : <i>9656671287</i>	Asset No. : <i>1012135</i>
	EQPT Name : <i>Biochemistry Analyser</i>
	Manufacture <i>Robonik</i> Model : <i>Prisetest Touch</i>
	S. No. <i>AT1400312RB</i> Dept. <i>Laboratory</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Biochemistry Analyser key pad not functioning.*

Action Taken : *Checked the machine and found key pad, primary board defective. need to replace these spare for further checking.*

Completed Date : *13/3/2024* Time : *12:10pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <i>Key pad</i>	1		
2. <i>primary board</i>	1		
3.			

Cyrix Engineer

shihabudheen

Date

13/3/2024

Start Time

10 AM

End Time

12:10pm

Customer Remark

[Signature]
23/3/24

Completed

Jr. Consultant Biomedical Engineer
NHM Malappuram

Pending

[Signature]
22/03/24

Service Engineer Name : *shihabudheen*

Signature : *[Signature]*

Date : *13/3/2024*

Contact Number : *9072522046*

Customer Name : *Rajitha P*


Signature : *[Signature]*

Date : *13/3/24*

Contact Number : *9446766096*

Designation : *STLS*

Hospital Seal :



STOCK REGISTER

Name: AUTOMATIC BIOCHEMISTRY ANALYSER (PIPETEST TOUCH)

Date	Received from or Issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs	Ps				
13-3-2012	ROBNIK (INDIA) PVT LTD	143	Rs					
	KODAMBAKKA M-9-B DIRE	93-12	88200		1 NOS		(NOS)	R
	KTOPYCOCONY - CHENNAI							
	600024-TEC-+9(44)2481520							
	<u>ACCESSORIES</u>							
	1) SINE WAVE UPS - 1 NOS							
	2) HALOGEN LAMP - 1 NOS							
	3) THERMAL PRINT PAPER - 2 NOS							
	4) DUST COVER - 1 NOS							
	5) VARIABLE VOLUME PIPETTE 10-100 ML - 1 NOS							
	6) FIXED VOLUME PIPETTE 500 ML - 1 NOS							
	<u>REAGENTS</u>							
	1) UREA MOD - 500 ml x 1							
	2) CREATININE - 500 ml x 1							
	3) BILIRUBIN TCD - 500 ml x 1							
	4) GLUCOSE - 500 ml x 1							
	5) CHOLESTROL - 500 ml x 1							
	6) DC ABNORMAL - 05 ml							
	7) DC NORMAL - 05 ml							
16-3-12	TO LAB TECHNICIAN	INDENT						
	MAJIDEVAN W/CD					1 NOS	NIL	R
	ACCESSORIES, REAGENTS							
16-4-13	HEALTH MEDICAL SERVICE CORPORATION LIMITED, TN	KNSCL/ BME/35 (100) 10 16.4.13	Rs 84521		4 NOS		4 NOS	R
20-5-13	PHC NENBARA	04/13-14				1	3 NOS	R
20-5-13	PHC THUNQATARA	5/13-14				1	2	R
20-5-13	PHC VAKKAN CHERI	6/13-14				1	1	R
20-5-13	TBC POKKANI	7/13-14				1	NIL	R

SN NO - AT1400312 RBK

[Signature]
 23/3/24
 Jr. Consultant Biomedical Engineer
 NHM Malappuram



[Signature]
 27/03/24
 DISTRICT TB OFFICER
 DISTRICT TB CENTRE
 MANJERI
 MALAPPURAM - 676123

Date : 18th March 2024

Dear Valued Customer,


Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT1400312RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT1400312RBK** as **End of Life (EOL) Notified.** (During from installation is 10 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**




Authorised Signatory



23/3/24
Jr. Consultant Biomedical Engineer
NHM Malappuram



27/3/24
DISTRICT TB OFFICER
DISTRICT TB CENTRE
MANJERI
MALAPPURAM - 676123



Jr. Consultant Biomedical Engineer
 NHM Malappuram
 NATIONAL HEALTH MISSION
 CIVIL STATION, MALAPPURAM-676 505

Manu
 27/3/24
 DISTRICT TB OFFICER
 DISTRICT TB CENTRE
 MANJERI
 MALAPPURAM - 676123