


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

461

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL
3	Name of Equipment with Make, Model and Serial Number	Equipment : SUCTION APPARATUS Make: SUPREME Model : NA SN : NA
4	Equipment ID & Barcode	99874 & 0110845
5	Date of purchase / Year of manufacture / Installation Date	20/05/2014
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	06-03-2024 (Toll free)
9	Action taken	Checked the machine and found motor coil burned, vacuum pressure gauge broken, suction bottle damaged.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Suction motor – Rs.4650/- Suction jar – Rs.2500/- Total value including GST – Rs.7150/-

13	Asset Value	Rs. 10200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	70%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found that motor has been burned out, pressure gauge broken off from the machine, suction bottle damaged. Machine installed on 20.05.2014 and covered up to 9+ years. Repair cost is 70%. Quotation submitted. Both criteria met and hence recommending the equipment for condemnation as per clause 5.3.14.1.
17	Name & Signature of CYRIX Authority	ADITHYA J S 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor burned and pressure gauge broken.
So RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram



20-03-2024

Signature of JC BM (NHM)




Date 23-03-2024

Seal



Signature of
Superintendent / Medical Officer (i/c)





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 193005

461

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Trivandrum</u></p> <p>Ph : <u>8138064143</u></p>	<p>Call Registration Date : <u>06-03-2024</u></p> <p>Caller ID : <u>99874</u></p> <p>Date of Visit : <u>08-03-2024</u></p> <p>Asset No. : <u>0110845</u></p> <p>EQPT Name : <u>Suction Apparatus</u></p> <p>Manufacture <u>Supreme</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>OT</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment not turning on.
Not working

Action Taken : Checked the machine and found suction motor defective. Motor coil burned down. Vacuum pressure gauge broken suction bottle damaged. Machine body rusted. Need to replace these spare parts for further checking the working condition of equipment

Completed Date : 08-03-24 Time : 4:30pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	NA	-	-	-
2.				
3.				



Cyrix Engineer	Date	Start Time	End Time
<u>ADITHYA J.S</u>	<u>08-03-24</u>	<u>3:00pm</u>	<u>4:00pm</u>

Customer Remark Completed Pending

Service Engineer Name : ADITHYA J.S
Signature : [Signature]
Date : 08-03-2024
Contact Number : 9745054034

Customer Name : [Signature]
Signature : [Signature]
Date : 8/3/2024
Contact Number : 8138064143
Designation : SNO OT
Hospital Seal : 8138064143

60 Suction apparatus

DATE	NAME of items	Qty	Issued Balance	Signature	Remarks
1-2-2014	Brought forward from page NO 16.3 Volume NO 1 Balance (18)				
19.3.93 (1)	16.9.98 (1)	10.5.99 (1)	7.4.2001 (1)	18	18'
5.11.2003 (2)	21.3.2005 (3)	15.6.2007 (3)			
23.5.2011 (5)	19.11.2013 (3)				
20/5/2014	Suction apparatus (Model Ship come plus .8.8.)	2	20		Rs 10,200/each (400 thousand and two hundred/each only)
	Stock of Reg. of Tax page NO 121 Volume 8A.				
20/5/2014	Suction apparatus (P/S) from G.W. page NO 78	1		21	
	As per order no. C1-5229/2014/				
	GAT dated 3/8/2015 Condemed (6)		15		
	As per order no C1, 5608/2016 dated 17/6/2016				
	Condemed (10)				
	21/3/2005 (3)	15/6/2007 (3)	23/5/2011 (5)	5	
25/7/18	Renewal form stove	1	6		um Stock Regri Page NO 84
4/3/2020	Renewal form stove	2	8		um Issue No. K10087/LP/00432/19-20 MSR.V0883/27
	Rs = 11,088 x 2 = 22,176/-				
17/4/2021	Received from stove NO 1400037/100920/21-22	1	9		um
14.7.21	Enter cl in Unservicable register page NO - 47	2	07		
23.9.21	Transferred to Stock register Vill. 11-A Page. 63		07		NIL
	As per order NO. C2-2161/26.7.2022/G.H.T dated 26.7.2022. Condemedation done				





9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No 22
CSI Golden Jubilee Complex
Baker Hill
Shashn Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0

State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

Estimate

Estimate For:

MS. TANIMA

CYRIX HEALTH CARE PVT LTD
1ST FLOOR, 30/641B
KOCHI DANUSHKODI ROAD
POONITHURA, ERNAKULAM
682038

Contact No.: 9072522291

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: 2023240137
Date: 15/03/2024
Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	□ 3,940.68	□ 709.32 (18%)	□ 4,650.00
2	SUCTION JAR BLACK LID	9018	2	Nos	□ 1,116.07	□ 267.86 (12%)	□ 2,500.00
Total			3			977.18	7,150.00

Sub Total	□ 6,172.82
SGST@6%	□ 133.93
CGST@6%	□ 133.93
SGST@9%	□ 354.66
CGST@9%	□ 354.66
Total	□ 7,150.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH

Description

GH TRIVANDRUM

OT DEPARTMENT

TICKET NO: 99874

BAR CODE 0110845

MAKE: SUPREME

Estimate Amount In Words

Seven Thousand One Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory