

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL
3	Name of Equipment with Make, Model and Serial Number	Equipment : BP APPARATUS Make: DIAMOND Model : DELUXE SN : 07538
4	Equipment ID & Barcode	100342 & 0110931
5	Date of purchase / Year of manufacture /Installation Date	06.10.2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	09-03-2024 (Toll free)
9	Action taken	Checked the machine and found mercury column broken and spilled out, tank damaged.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 1323
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found that the mercury column has been broken and mercury spillage occurred. Mercury tank also damaged. Mercury not available in the market, recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	MOHAMMED SIYADH M

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Mercury column broken and mercury spillage occurred. So RBER.

MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram


Handwritten Signature
 20-03-2024

Signature of JC BM (NHM)

Date



Handwritten Signature
 Signature of
 Superintendent / Medical Officer (i/c)
Superintendent
General Hospital
Thiruvananthapuram



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 192804

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY / AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com E-mail : bengal@cyrix.in

Health Facility <u>AM Thuvananthapuram</u> Address <u>AM Junction</u> <u>Thuvananthapuram Kerala</u> Ph : <u>99 61 89 13 92</u>	Call Registration Date : <u>09/03/24</u> Caller ID : <u>100342</u> Date of Visit : <u>16/03/24</u> Asset No. : <u>010931</u> EQPT Name : <u>BP Apparatus Mercury</u> Manufacture <u>Diamond</u> Model : <u>Regulge</u> S. No. <u>07538</u> Dept. <u>PMR Island</u>
--	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Checked the machine and found mercury column and mercury tank broken. Mercury spilled out completely.

Action Taken : Mercury column and mercury tank broken. Mercury spilled out completely. Need to replace the mercury tank and mercury column and should refill the mercury for further checking and working condition of the equipment.

Completed Date : 16/03/24 Time : 04:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>

Cyrix Engineer <u>Mohamed Sijad</u>	Date <u>16/03/24</u>	Start Time <u>03:30 pm</u>	End Time <u>04:00 pm</u>
--	-------------------------	-------------------------------	-----------------------------

Customer Remark Completed Pending

Service Engineer Name : Mohamed Sijad M
 Signature : [Signature]
 Date : 16/03/24
 Contact Number : 7293499129

Customer Name : Sajitha. M. N
 Signature : [Signature]
 Date : 16/03/24
 Contact Number : 7306427483
 Designation : Nursing officer
 Hospital Seal :



(New BP apparatus)

[Handwritten notes and signatures]

B.P. APPARATUS

	Received	Balance	Remarks
20-11-2013	Physically found & taken from stock	5	4 not working
20-11-13	one BP apparatus received from store cost of Rs: 17.14/- page no. 168 HDC local purchase volume 4.	6	
14/1/17	Received from store (Digitized BP apparatus) ms volume 8 & page no. 61 ids 1200/- (Entered in page no 117)	1	5 not work Fin into
06-10-18	Tr. in from spl. OP Barcode No (8004) 890757955700 12577	7	2 CS Sub
25/6/19	Point of PMA OP no. 1	6	
22/12/21	Received from store (LP) 1 (Digital) (Entered in Page No. 117)	7 Cancelled	8
9/4/23	Tr. in from Wd 3 & 4 (Page No. 46) (5 not working)	1	7 SND
13/2/24	Condemnation done as per Order No. Ce-5549/2023, G.H.T Dated 13/2/2024 (3) (2 not working) (3 not working)	4	SND
1/3/24	3 not working		
9/3/2024	Tr in from Cardiology Jew (old BP Stock book No. Apparatus) 1	5	4 not working
9/3/2024	Tr in from CD OP Stock book Vol 2 Page No. 13	1	6
9/3/24	Tr in from Wg Stock book page no. (New BP apparatus)	1	7



Roshni
n/o

Handwritten signature and initials in blue ink.

Diamond

Regular

B.P. APPARATUS

112 MAY.17 07538

IS 3390




CM/L-0196043

IN No. of S38
Call No. 04119
M/S. Engineering & Services
M/S. ALABASTON (PVT) LTD.
M/S. ALABASTON (PVT) LTD.

TANK LOCK OPERATING INSTRUCTION

To operate the instrument.
Turn the handle of the lock by 90° towards right (lock when) as shown in the diagram. The mercury will rise in the glass tube. Now the instrument is ready for use.
To close the instrument.
Turn the handle of the lock by 90° towards left (unlock when) as shown in the diagram. The mercury will fall in the glass tube. Now the instrument is ready for use.



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0110931



Superintendent
General Hospital
Thiruvananthapuram-35

Roshmi
N/A