


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

474

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL
3	Name of Equipment with Make, Model and Serial Number	Equipment : SUCTION APPARATUS Make: MONARCH Model : MSI- RG02 SN : 2020029
4	Equipment ID & Barcode	102531 & 0111869
5	Date of purchase / Year of manufacture /Installation Date	19/11/2013
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	21-03-2024 (Toll free)
9	Action taken	Checked the machine and found motor and suction bottles damaged.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Suction motor – Rs.5103/- Suction jar with lid– Rs.3790/- Total value including GST – Rs.8893/-

13	Asset Value	Rs. 9900
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	89%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and quotation attached
16	Reasons for recommending the equipment as BER	Checked and found that motor and suction bottles damaged. Machine installed on 19.11.2013 and covered up to 10+years. Repair cost is above 85%. Quotation submitted. Both criteria met and hence recommending the equipment for condemnation as per clause 5.3.14.1.
17	Name & Signature of CYRIX Authority	ADITHYA J S 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Motor damaged and suction bottles need replacement
So cannot be repaired economically.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


4-04-2024
Signature of JC BM (NHM)

Date




Signature of
Superintendent / Medical Officer (i/c)

ndent
ospital
Thiruvananthapuram



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Order No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 192896

474

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>General Hospital</u>	Call Registration Date : <u>21/03/2024</u>
Address <u>Trivandrum</u>	Caller ID : <u>102531</u>
Ph : <u>9766173198</u>	Date of Visit : <u>22/03/2024</u>
	Asset No. : <u>0111869</u>
	EQPT Name : <u>Suction Apparatus</u>
	Manufacture <u>Monarch</u> Model : <u>MS1-R602</u>
	S. No. <u>2020029</u> Dept. <u>01</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not switching ON


Action Taken : Checked the machine and found that the motor and suction bottles damaged. Need to replace these parts for further checking of the machine. Enquired with OEM for parts.

Completed Date : 22/03/24 Time : 1:00pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ADITHYA. J.S</u>	<u>22/03/24</u>	<u>12:30 PM</u>	<u>1:00 PM</u>
Customer Remark	<input checked="" type="checkbox"/> Completed <input type="checkbox"/> Pending		



Service Engineer Name : <u>ADITHYA. J.S</u>	Customer Name : <u>Sreelatha</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>22/03/24</u>	Date : <u>22/3/2024</u>
Contact Number : <u>9765056344</u>	Contact Number : <u>800 9495895103</u>
	Hospital Seal : <u>[Seal]</u>



CLOUD99 BIOTECH

Office : 97444 7304

Service : 97444 7324

Sales and service of all kinds of medical equipments
Room No: 22, First Floor, CSI Golden Jubilee Complex,
Behind Shastri Road Waiting Shed, Kottayam-686001
Email: cloud99biotech@gmail.com

QUOTATION

Ref No: C99B/0148/2023-24

Date: 28/03/2024

To,

GH

TRIVANDRUM

Dear Sir

We are pleased to submit a quotation for **GH, TRIVANDRUM** for Suction Apparatus. Barcode is **0111869, Make- MONARCH, Model-MSI RG02.**

SL No	ITEM NAME	PRICE PER UNIT	GST	PRICE INCLUDING GST
1	MOTOR	4325	18%	5103
2	BOTTLE WITH LID – 2 NOS	3212	18%	3790
	TOTAL			8893

Terms & Conditions

1. Taxes: GST @ 18 % inclusive
2. Payment Terms: 100% advance payment
3. Quotation validity: 15 days from date of quotation

We hope you will find our rates are competitive and wait for your valued order.

Thanking you,

Yours Faithfully


FOR CLOUD99 BIOTECH

AUTHORIZED SIGNATORY

We deal with sales and service of all kinds of hospital equipments, neonatal equipments, laboratory equipments, dental equipments, suction, hospital furniture, radiology equipments, physiotherapy equipments, all type of autoclaves, ICU equipments, etc.


Date	Name of items	Qty	Unit	Balance	Signature	Remarks
12-2014	Brought forward page No 163 Volume No 1 Balance (18)					
19.3.93 (1)	16.9.58 (1)	10.5.99 (1)	7.4.2001 (1)	18	18	
5.11.2003 (2)	21.3.2005 (3)	15.6.2007 (2)	23.5.2011 (5)			
20/5/2014	Suction apparatus (Model Spicene plus .85) Stock of Reg. of the page No 121 Volume 84.	2		20		Rs 10,000 each (Ten thousand and two hundred/each only)
20/5/2014	Suction apparatus from G.N. page No 79	1		21		
	As per order no. C ₄ -5229/2014/ GNT dated 3/8/2015	Condensed (6)		15		
	As per order no C ₄ 5608/2016 dated 17/6/2016	Condensed (10)		5		
25/7/18	Received from store	1		6		cm Stock Regi Page No 84
4/3/2020	Received from store Rs-11,088 x 2 = 22,176/-	2		8		cm Issue No. K.C.0087/LP/00632/19-20 NSR.VRESB 27
17/4/2021	Received from store No 1400037/acc/00920/21-22	1		9		cm
14-7-21	Enter of in Unexmolla register Page No-47	2		07		
23.9.21	Transferred to stock register Vol. 11-A Page. 13			07		NIL
	As per order no. C ₂ -2161/26.7.2022/C.H.T dated 26.7.2022					Condensation done





LOW  HIGH

INPUT 220 / 230 VOLTS
A.C. MAINS 5 AMPS 50Hz
SINGLE PHASE
DUTY : CONTINUOUS

WATT	
TYPE	1
S.NO	


S 4533
 810098395
Model: MSI-RG02
 Type: 1
 S No: 2020029

Suction Apparatus

POWER  OFF
 ON

Toll FREE - 1800-2700-690
 4791 3486233 - 34861476 10012486

VACUUM GAUGE

LOW  HIGH

INPUT 220 / 230 VOLTS
A.C. MAINS 5 AMPS 50Hz
SINGLE PHASE
DUTY : CONTINUOUS

WATT	
TYPE	1
S.NO	

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
 TOLL FREE NO: 1800- 425-7669
 BAR CODE - 0111869 

CYRIX
 ELEC. TRIC. & S.C. TRIC. UNIT
 CN: 0011869
 ID: 0111869
 I.A.S. ON: 22-04-2023
 I.A.S. BY: 23-04-2023