


471

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL
3	Name of Equipment with Make, Model and Serial Number	Equipment : PHACO EMULSIFICATION UNIT Make: APPASAMY ASSOCIATES Model : GALAXY PRO SN : NA
4	Equipment ID & Barcode	102462 & 0112023
5	Date of purchase / Year of manufacture /Installation Date	04.05.2012
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	21-03-2024 (Toll free)
9	Action taken	Checked the machine and found the touch screen, display, main control board and foot switch defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs.7,88,849
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and Beyond economical repair certificate from manufacturer
16	Reasons for recommending the equipment as BER	Checked the machine and found the touch screen, display and main control board and foot switch defective. Equipment installed on 04.05.2012 and has covered up to 11+ years. On enquiring for replacement of spare parts, it was identified and communicated by the manufacturer that the particular series is now outdated and no longer spares are available and hence cannot be serviced. Considering the circumstances and in adherence to the tender clause 5.3.14.2, which states that in case an equipment is declared as end of support by manufacturer, the equipment should be evaluated for condemnation and hence recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	ADITHYA J S 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine installed on 2012 and also major parts are damaged. So as per conditions it can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


4-4-2024


Signature of JC BM (NHM)



Date

Seal

Signature of
Superintendent / Medical Officer (i/c)


Superintendent
General Hospital
Thiruvananthapuram



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 192892

471

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Trivandrum</u></p> <p>Ph : <u>9744773198</u></p>	<p>Call Registration Date : <u>21-03-2024</u></p> <p>Caller ID : <u>102462</u></p> <p>Date of Visit : <u>21-03-2024</u></p> <p>Asset No. : <u>0112023</u></p> <p>EQPT Name : <u>Phacoemulsification Unit</u></p> <p>Manufacture <u>Appasamy</u> Model : <u>Galaxy pro</u></p> <p>S. No. <u>NA</u> Dept. <u>OT</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Touch screen not working

Action Taken : Checked the machine and found touch screen, display, main control board and footswitch defective. Checked with OEM for spare parts.

Completed Date : 21/03/24 Time : 1:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	-	-	-
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Adithya J.S</u> <u>Tanima Ehinan</u>	<u>21/03/24</u>	<u>1:00pm</u>	<u>1:30pm</u>
Customer Remark	Pending		






<p>Service Engineer Name : <u>Adithya J.S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>21-03-2024</u></p> <p>Contact Number : <u>9745054344</u></p>	<p>Customer Name : <u>Sreelekha S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>9/9/5895103</u></p> <p>Contact Number : <u>[Number]</u></p> <p>Designation : <u>Senior nursing officer</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--

Superintendent
 General Hospital
 Thiruvananthapuram
[Signature]
 2024

PHACO MACHINE

330

Date	Name of equipment	RD	Credit Balance	Signature	Remarks
1-2-2014	Brought Forward from page No 204 Volume 1 Below ①				
14-5-2012	(Rs 7,88,849 (Rs Seven lakhs, eighty eight thousand eight hundred and eighty nine only) Page No 290 Volume 8)				
20/9/2011	RT out to Stockhall VOT eye page no: HT				Nil



CURE HEALTH CARE PVT LTD
REPAIR SERVICE PROVIDER
REGISTRATION NUMBER: 15783
HOSPITAL NAME: ASM
APPROVED BY: ASM

CURE HEALTH CARE PVT LTD
TOLL FREE NO: 1800-425-788
844 0000 0112833

APPASAMY ASSOCIATES
CHENNAI, INDIA.
Model : GALAXY PRO (P)
Power cons : 1.1 A
Power Supply : 150 V / 50 Hz
Sl. No



APPASAMY ASSOCIATES (P) LTD.

Regd. Office : 20, SBI Officers' Colony, 1st Street,
Arumbakkam, Chennai - 600 106, INDIA

☎ (91-44) 30101400, 30101401

✉ info@appasamy.com

CIN : U33201TN2019PTC129614

GSTIN : 33AASCA3660R1Z0

BEYOND ECONOMICAL REPAIR CERTIFICATE

Date: 28.03.2024

To
The Superintendent,
General Hospital,
Trivandrum.

Sub: BER Certificate- Regarding
Equipment: Phaco Emulsification
Model: Galaxy Pro - P
Sl. No: 08-14-0811

Dear Sir/Madam,

We would like to inform you that our service engineer has visited your hospital on 21.03.2024 and inspected above equipment and found that the equipment was 10 years old and used for a considerable long period and cannot be repaired/ serviced and also spare parts are not available. Since the above Equipment is outdated, we suggest you to purchase the new equipment.

Looking forward to your continued support.

Thanking you,

Sincerely yours,

FOR APPASAMY ASSOCIATES PVT.LTD.

