


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	TH MALAYINKEEZH
3	Name of Equipment with Make, Model and Serial Number	Equipment : BP APPARATUS Make : DR.ODIN Model : BSX516 SN : 516CAL1117614
4	Equipment ID & Barcode	98926 & 0121883
5	Date of purchase / Year of manufacture /Installation Date	27-07-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01-03-2024 (Toll free)
9	Action taken	Checked the machine and found solenoid valve and motor defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	Rs. 2200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found solenoid valve and motor is defective. Quotation not attached as spare not available in the market. Hence recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	PRANAV P S 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Solenoid valve and motor is defective.
 So BER.

MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram



 20-3-2024

Signature of JC BM (NHM)

Date



Seal


 20-3-24

Signature of
 Superintendent / Medical Officer (i/c)
Superintendent
 Taluk Headquarters Hospital
 Talavinkil



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 178025

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>THQV</u>	Call Registration Date : <u>1/2/24</u>
Address : <u>Maly. n. k. resher</u>	Caller ID : <u>98926</u>
<u>Tirunelveli</u>	Date of Visit : <u>2/3/24</u>
Ph : <u>7570235972</u>	Asset No. : <u>0121883</u>
	EQPT Name : <u>BP apparatus</u>
	Manufacture <u>Dr. med.</u> Model : <u>BSX</u>
	S. No. <u>5166911 784</u> Dept. <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Inflation and deflation issue

Action Taken : Checked the machine and found that the motor and sensor value is defective. need to replace these spare for further checking and working condition of the equipment.

Completed Date : 2/3/24 Time : 10:30am Spare Required

Spare Replaced Requested




Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Pranav</u>	<u>2/3/24</u>	<u>10am</u>	<u>10:30am</u>

Customer Remark Completed Pending

Service Engineer Name : Pranav P.S.
Signature : [Signature]
Date : 2/3/24
Contact Number : 9094520652

Customer Name : Ashetha HS
Signature : [Signature]
Date : 2/3/24
Contact Number : 8921014485
Designation : Nursing Officer
Hospital Seal : [Seal]



Date	Received	Issued	Balance	
10/12/19	B/P fm Subscriber Pg: Pr: do	2 do.	2	
5/7/2021	Physically found Ur. I.	2 do.	4 do.	
27/7/2021	Received from store [LP] Main store book Pg no: 60 Lot No. 516W1032021 ₹. 1385/- Di. Odia. workings of net	2 do.	5 do.	
16/8/2021	one machine mt. Keprirchha PBR report got.			
4/12/2021	Received from store Value ₹ 3,050.50 (Donation → CAMARA BANK) Issued to OPF	1 do.	7 do.	