



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 194076

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THQH</u></p> <p>Address <u>Cherthala</u> <u>Alappuzha</u></p> <p>Ph : <u>9846553583</u></p>	<p>Call Registration Date : <u>2/3/2024</u></p> <p>Caller ID : <u>99165</u></p> <p>Date of Visit : <u>4/3/2024</u></p> <p>Asset No. : <u>0420341</u></p> <p>EQPT Name : <u>ECG machine</u></p> <p>Manufacture <u>BPL</u> Model : <u>10&T-Digi</u></p> <p>S. No. <u>CHML4922894</u> Dept. <u>ECG Dept</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : ECG machine is not proper working and main Board is faulty

Action Taken : Checked the machine and found that ECG main board, pointer's Head Assy & Battery are complaint. Enquired the OEM for quotation

Completed Date : 27/3/24 Time : 3:00PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nasut</u>	<u>27/3/24</u>	<u>12:00PM</u>	<u>3:00PM</u>

Customer Remark Completed Pending

Aaryg ECG Theekun Not working.

Service Engineer Name : Nasut

Signature : Nasut

Date : 27/3/24

Contact Number : 9526744524



Customer Name : _____

Signature : [Signature]

Date : _____

Contact Number : _____

Designation : SUPERINTENDENT

Hospital Seal : TALUK HEAD QUARTERS HOSPITAL, CHERTHALA

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH CHERTHALA
3	Name of Equipment with Make, Model and Serial Number	ECG MACHINE SINGLE CHANNEL MAKE :- BPL MODEL :- CARDIART 108T DIGI SN :- CGML4G22896
4	Equipment ID & Barcode	99165 & 0420341
5	Date of purchase / Year of manufacture /Installation Date	03/02/2015
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	02/03/2024
9	Action taken	Checked the machine and found that ECG Mainboard, Printer Head Assy & Battery are complaint. Enquired the OEM for Quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommending For Repair.
12	Cost of spares (specify parts and cost)	108T DIGI MAINBOARD :- 10620 9.6V BATTERY :- 2048 PRINT HEAD ASSY 108T :- 2206.6 TOTAL :-14874.6

13	Asset Value	RS:- 18656/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Percentage Value Of The Cost Of Spare:- 79%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report & OEM Quotation
16	Reasons for recommending the equipment as BER	Checked the machine and found that ECG Mainboard, Printer Head Assay & Battery are complaint. Enquired the OEM for Quotation. The equipment installed on 02/03/2015, aged up to 9 years. Percentage of service cost is high (79%) and economically not viable. Both criteria met as per the tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NASEEF K N <i>Naseef</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM ALAPPUZHA

[Signature] 04/04/24

Signature of JC BM (NHM)

For condemnation

Date

Seal



[Signature]
Signature of SUPERINTENDENT
Superintendent, Medical Officers,
ALUR HEAD QUARTERS HOSPITAL
CHERTHALA

Date	Item	Receipt No.	8	9 Nos. 5
19.04.2014	BPL ECG machine Model No. 6108-T 2 Accessories (Single Channel) SL No. AVM/EG/40459 From Elixix Medi Systems 3 yrs warranty from 19.04.14 to 18.04.17 24622.50, Page No. 135 of stock book B-1 NR 410	1 No.	E+1	9 Nos. 5
03.02.2015	BPL ECG machine Model No. 108T DIGI SL No. COM14G 22896 (SING FROM Elixix Medi System Warranty period from 02.02.2015 - 01.02.2016 Order No. C3/2013/14/7.H.C dated 19/01/2015	1 No.	9+1-10	10 No. 6
24/10/17	Received from store Register	1 No.	Page No. 176	Instrument
24/10/17	Received from the Store [B.P.L ECG machine] page no: 176 Model No: Cassidial 9108 Serial No. DUTBFE.5851 Warranty period from 23/10/2017 to 27/10/2019 K.M.S.C. Order No: 50/406/2017-18/38 dated 07/09/2017	1 No.	R.J.	Instrument Register



Handwritten signature and initials in blue ink, with the number '3/4' written below.

QTN/BISO/270324

27/03/2024

QUOTATION

To,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

Dear Sir,

Please find below our lowest quotation for the following items:-

SL NO.	ITEM DESCRIPTION	COST PER UNIT	UNIT PRICE
1	108T MAINBOARD	9000	9000+18%TAX
2	9.6V BATTERY	1600	1600+28%TAX
3	PRINT HEAD ASSY 108T DIGI	1870	1870+18%TAX

Terms and conditions:

1. Payment 100% in advance along with Purchase order
2. Delivery within 15 days

Thanking you and assuring you our best attention always.

For Biomedix Solutions



Authorized Signatory



[Handwritten Signature]
3/4

Mobile: 9846969564 | Email: biomedixservice@gmail.com

Branohes: Kozhikode, Kannur, Kottayam-Kollam.





For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA



BIOMEDIX SOLUTIONS

No: 7/437- B, Mannarakayam P.O,
Kathilankalpadi, Kanjirapally, Kottayam - 686506,
Mob:9846969564,



CHERTHALA
 SUPERINTENDENT
 TALUK HEAD QUARTERS HOSPITAL
 CHERTHALA

3/11

Handwritten signature