



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 193763

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC, CHERUMANA</u></p> <p>Address <u>CHERUMANA</u> <u>ALAPPUZHA (D)</u></p> <p>Ph : <u>9544391279</u></p>	<p>Call Registration Date : <u>14.03.2024</u></p> <p>Caller ID : <u>101239</u></p> <p>Date of Visit : <u>15.03.2024</u></p> <p>Asset No. : <u>0443718</u></p> <p>EQPT Name : <u>NEBULISER</u></p> <p>Manufacture <u>OWISEL OXY MED</u> Model : <u>MO-NB201</u></p> <p>S. No. <u>000</u> Dept. <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT WORKING

Action Taken : ON CHECKING FOUND THAT MACHINE MOTOR AND PISTON ARE DAMAGED CONDITION. ENQUIRED SPARE NEED TO REPLACE SPARE FOR FURTHER CHECKING.

Completed Date : 20.3.24 Time : 11:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer

Date

Start Time

End Time

SAUBAN BIN YASEEN

20.3.24

9:30am

11:00am

Customer Remark

Completed

Pending

Service Engineer Name : SAUBAN

Signature :

Date : 20.3.24

Contact Number : 9745859470



Customer Name : Cyrix and P

Signature :

Date : 22/3/2024

Contact Number :

Designation : Nursing Officer

Hospital Seal :


(Handwritten notes and signatures in the bottom right corner)

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	PHC CHERUTHANA
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :- OWGELS OXY-MED MODE :- MONBZ01 SN :- NA
4	Equipment ID & Barcode	101239 & 0443718
5	Date of purchase / Year of manufacture / Installation Date	27/03/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	14/03/2024
9	Action taken	Checked the machine and found that the nebulizer motor and piston are damaged condition.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	RS:- 1353/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached.
16	Reasons for recommending the equipment as BER	Checked the machine and found that the nebulizer motor and piston are damaged condition. Equipment is installed on 27/03/2018 and aged up to 6 years. Quotation not submitted since spares are not available in the market. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	SAUBAN BIN YASEEN 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


 JYOTHISH THOMAS
 Consultant (Biomedical Engineering)
 NHM ALAPPUZHA

Signature of JC BM (NHM)

Date




 Signature of the Superintendent / Medical Officer (I/c)
 NHM Alappuzha
 0471-251117

NEBULIZER

Date	Receipt	Issue	Balance	DA
4-6-12	Received from pharmacy	1	Nil not working for 7115 (got certificate for repairable) Nil from (pharmacy)	4-6
18/9/17	Compact Nebulizer (HPE) Received from pharmacy (Dorset) OXY med (OWGELS)	1	working. Working.	①
27/9/18	Nebulizer	1		②
12/4/18	Nebulizer (Truair checks)	1		③
17/7/19	Nebulizer (Life care / compressor)	1		④
26/8/19	Nebulizer (Life care) Exp. dat - 1/28 Rate 1232.13/-	1		⑤
pharmacy stock page 70 - 70				
29/7/21	Nebulizer (Dr. Morepen) Date & Year of import 4/2020 Rate - 2250/-	1		⑥



സെക്രട്ടറി
 കേരള സർക്കാർ
 മെഡിക്കൽ സർവ്വീസ്
 തൃശ്ശൂർ



[Handwritten Signature]

ആരോഗ്യ വകുപ്പ്
കേരള സർക്കാർ
ആരോഗ്യ വകുപ്പ്
ആരോഗ്യ വകുപ്പ് എ.ഐ. - 680517.