


REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	GENERAL HOSPITAL, KOTTAYAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : B P APPARATUS Make: MICO SURGICALS Model: MICO SN: 70716
4	Equipment ID & Barcode	98975 & 0510869
5	Date of purchase / Year of manufacture / Installation Date	19-02-2011
6	Warranty details (Yes/No)	No warranty
7	*AMC/CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/Toll free)	01-03-2024
9	Action taken	Check and Found the mercury tank leaked and spilled out the mercury. screws of hinges broken with in the case due to rusted
10	Present status of the equipment (Fully damaged/partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Mercury leak and equipment was wear and tared. Not recommending for repair.
12	Cost of spares (specify parts and cost)	Equal to Asset value

13	Asset Value	Rs.1323
14	#Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	Equal to Asset value
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/CYRIX(Attached or Not)	Cyrix service report attached.
16	Reasons for recommending the equipment as BER	Check and Found the mercury tank leaked and spilled out the mercury. Screws of hinges broken with in the case due to rusted, need to replace all spare and mercury not available in the market. The machine installed on 19/02/2011 and covered up to 12+years .so recommending to condemnation
17	Name & Signature of CYRIX Authority	SANISH SIVAN 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

*Attach Photograph

Remarks and Recommendations of Consultant (Biomedical) NHM:


Asset value: 1323

Cost of spares: ~~equal~~ Equal to Asset value

Life period - 13 year


As per the guideline recommended as R.B.E.R

CONSULTANT BIOMEDICAL ENGINEER
NATIONAL HEALTH MISSION
KOTTAYAM
Signature of JCBM (NHM)
15/12/24



Date

Seal


5/3/2017
Superintendent / Medical Officer (A/C)
General Hospital
Kottayam



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 172602

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>G.H</u>	Call Registration Date : <u>01/03/2024</u>
Address : <u>General Hospital</u>	Caller ID : <u>98975</u>
<u>Kottayam</u>	Date of Visit : <u>11/03/2024</u>
Ph : <u>9495443913</u>	Asset No. : <u>0510869</u>
	EQPT Name : <u>BP Apparatus</u>
	Manufacture <u>Mico Suresh</u> Model : <u>Mico</u>
	S. No. <u>70716</u> Dept. <u>Ward-6</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Found the mercury tank have leak and spilled out the mercury. Screws of hinges broken within the case due to rust.

Action Taken : Checked and understand that need to replace all spares and mercury to make the equipment functional. The equipment installed on 19/2/2011. Hence recommending the equipment as B.E.R and for condemnation.

Completed Date : Time : Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sarish Swan</u>	<u>11/3/24</u>	<u>11:00am</u>	<u>11:50am</u>

Customer Remark Completed Pending
old Bp apparatus not able to use in ward sup senior Nsg officer.

Service Engineer Name : <u>Sarish Swan</u>	Customer Name : <u>Geey. VV</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>11/3/24</u>	Date : <u>[Date]</u>
Contact Number : <u>9966646571</u>	Contact Number : <u>[Number]</u>
	Designation : <u>S.No.</u>
	Hospital Seal : <u>[Seal]</u>



Jan

