


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

445

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Trivandrum
2	Name of Hospital	GH Trivandrum
3	Name of Equipment with Make, Model and Serial Number	Name of Equipment : Pulse oximeter Make : L & T Model : Comet SN : C12P3002
4	Equipment ID & Barcode	0110234
5	Date of purchase / Year of manufacture / Installation Date	30/6/2016 (transferred in)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	10/2/2024
9	Action taken	check the machine Main board and display board not working
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Major part of the machine (main board and display board) are not working.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	45013
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report and photos, OEM end of service support report.
16	Reasons for recommending the equipment as BER	Major part of the machine (main board and display board) not working. This machine that has been out of production and manufacturer discontinues this machine. parts are not available for repair.
17	Name & Signature of CYRIX Authority	Josekutty George 


\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified the EOL from OEM.  
So can be RBER.


MANEESHA MOHAN M.R.  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
14-02-2024

Signature of JC BM (NHM)

Date



  
Signature of Superintendent / Medical Officer (i/c)  
General Hospital  
Thiruvananthapuram



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 176348

445

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>GH Trivandrum</i>	Call Registration Date : <i>10/2/2024</i>
Address <i>Trivandrum</i>	Caller ID : <i>95362</i>
Ph : <i>9496196776</i>	Date of Visit : <i>12/2/2024</i>
	Asset No. : <i>0110234</i>
	EQPT Name : <i>Pulse Oximeter</i>
	Manufacture <i>L&amp;T</i> Model : <i>Comet</i>
	S. No. <i>C12P3002</i> Dept. <i>Male Ward</i>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : *Not working*

Action Taken : *Check the machine Main board and display not working. This machine that has been out of production and manufacturer discontinues this machine. parts are not available for repair.*

Completed  Date : *13/2/2024* Time : *10:30 am* Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. RBER			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Josekutty George</i>	<i>13/2/2024</i>	<i>10am</i>	<i>10:30am</i>

Customer Remark *Completed*  Pending



Service Engineer Name : <i>Josekutty George</i>	Customer Name : <i>SUREJA.P</i>
Signature : <i>[Signature]</i>	Signature : <i>[Signature]</i>
Date : <i>13/2/2024</i>	Date : <i>13/2/24</i>
Contact Number : <i>8848230407</i>	Contact Number : <i>9448843558</i>
	Designation : <i>SK6</i>
	Hospital Seal :

Dear Customer,

Date: 24.01.2023

**Subject: End of service support for Comet model**

It is to inform you that the above said product is out of production and it is **obsoleted.**

There is no service support available for it, hence it is fit for condemnation.

Thanking you,

For Skanray Technologies Limited





*H M.*

Avinash H M

Service Operations Co-ordinator



PART NO.:  
F3-85-390-0208-06  
SER. NO  
C12P3002



**MEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)**  
**FREE NO: 1800- 425-7666**  
BAR CODE - 0110234

