



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 168674

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>WAC Thayaland</u>	Call Registration Date : <u>27/11/23</u>
Address : <u>Irirandam</u>	Caller ID : <u>82510</u>
Ph : <u>8547305724</u>	Date of Visit : <u>28/11/23</u>
	Asset No. : <u>0114845</u>
	EQPT Name : <u>Pharmacy Refrigerator</u>
	Manufacture <u>Electrolux</u> Model : <u>ERP 285 F</u>
	S. No. <u>63400458</u> Dept. <u>Blood Bank</u>

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : Checked and found machine is rusted and door clamp broken due to rust. cannot fix the door back.

Completed Date : 28/11/23 Time : 3pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.	—	—	—
2.	—	—	—
3.	—	—	—

Cyrix Engineer	Date	Start Time	End Time
<u>Salika S</u>	<u>28/11/23</u>	<u>2pm</u>	<u>3pm</u>

Customer Remark Completed Pending



Service Engineer Name : Salika S
Signature : [Signature]
Date : 28/11/23
Contact Number : 6282656760


Customer Name : N. USHA
Signature : [Signature]
Date : for M/O
Contact Number : 98472 99500
Designation : for M/O
Hospital Seal : [Signature]

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	TRIVANDRUM
2	Name of Hospital	W & C THAYCAUD
3	Name of Equipment with Make, Model and Serial Number	Refrigerator Electrolux ERP 285FF 63400458
4	Equipment ID & Barcode	0114845
5	Date of purchase / Year of manufacture /Installation Date	2/4/2007
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	NO
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	27/11/23
9	Action taken	Checked and found machine is rusted and door holding clamp broken
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Checked and found machine is rusted. Door holding clamp broken due to rust. Cannot fix the door.
12	Cost of spares (specify parts and cost)	Not repairable

13	Asset Value	17,960/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached
16	Reasons for recommending the equipment as BER	Checked and found refrigerator is rusted. Door clamp broken due to rust. Cannot fix the door.
17	Name & Signature of CYRIX Authority	


*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Fridge leg fully rusted, door cannot be fixed. Can be RBER.

MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram


 6-12-2023

Signature of JC BM (NHM)



Date

PI SUPERINTENDENT
 Govt. WI & O Hospital Thiruvananthapuram
 Thiruvananthapuram 14
 6/12/23

Signature of Superintendent / Medical Officer (i/c)

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0114845**



MODEL

ERP 285FF

CAPACITY

275 Lts

SUPPLY

220V, 10, 50Hz

VOLTAGE RANGE

170 - 250 V

RATED CURRENT

1.7 A Max

POWER INPUT

180 W Max

REFRIGERANT

R134a, 76gm

PNC No - 925024754

SR No - 63400458

 **Electrolux**®

WIRING DIAGRAM

