


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

436

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Thiruvananthapuram
2	Name of Hospital	Urban Primary Health Center Chacka.
3	Name of Equipment with Make, Model and Serial Number	Centrifuge, Make: Remi Model: C-854/8 Serial No: NA
4	Equipment ID & Barcode	89895 - 0140722
5	Date of purchase / Year of manufacture / Installation Date	17-09-2015
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/12/2023
9	Action taken	Checked the machine and found motor winding, Carbon brush, armature damaged, fully tested.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Machine cannot be repaired because motor winding and armature damaged Carbon brush checked and machine totally tested.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	9657 Rs. 5450
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached
16	Reasons for recommending the equipment as BER	Motor winding and armature damaged. Carbon brush damaged. Machine fully rusted. Cannot be repaired economically. So recommending for condemnation.
17	Name & Signature of CYRIX Authority	Mohammed Siradh M 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine fully rusted, Motor winding and armature damaged. So RBER.


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


4.01.2024
Signature of JC BM (NHM)

Date 5/01/24



Signature of
Superintendent / Medical Officer (i/c)


Medical Officer
UPHC-CHACKAL
Thiruvananthapuram



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 178221

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

436

Health Facility : UPHC CHACKAI
Address : CHACKA
TRIVANORUM
Ph : 9539375448

Call Registration Date : 29/12/2023
Caller ID : 87895
Date of Visit : 5/01/2024
Asset No. : 0140722
EQPT Name : Centrifuge
Manufacture : REMI Model : C-854/8
S. No. : NA Dept. : LAB

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Motor winding damage, Carbon brush dried out, armature damage, fully rusted

Action Taken : Checked the machine, centrifuge motor winding damage, carbon brush dried out, armature damage, fully rusted. Machine in this condition cannot be repaired.

Completed Date : 5/01/2024 Time : 10 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>RBER</u>			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Siyadh</u>	<u>5/01/2024</u>	<u>9 am</u>	<u>10 am</u>

Customer Remark Completed Pending

Service Engineer Name : Siyadh
Signature : Siyadh
Date : 5/01/2024
Contact Number : 7293499127

Customer Name : [Signature]
Signature : [Signature]
Date : 5/01/24
Contact Number : [Signature]
Designation : Medical Officer
Hospital Seal : UPHC-CHACKAI
Thiruvananthapuram



