

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

438
437

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Thuvananthapuram
2	Name of Hospital	Fite Parakkal
3	Name of Equipment with Make, Model and Serial Number	BP Apparatus LED ELKO - EL520. NR.
4	Equipment ID & Barcode	0144415.
5	Date of purchase / Year of manufacture / Installation Date	15/03/18
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	NR
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	04/01/24
9	Action taken	Checked the machine and found board damaged, Battery case damaged
10	Present status of the equipment (Fully damaged / partially damaged)	Partially Damaged
11	Recommendations for repair (required service details)	Replacing board and replacing Battery case is not economically possible.
12	Cost of spares (specify parts and cost)	NR

13	Asset Value	2200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NR
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Find the attached copy of Cyrix Service report along with the RBER Protocols.
16	Reasons for recommending the equipment as BER	Board damaged and battery case damaged cannot be certified
17	Name & Signature of CYRIX Authority	Mohamed Sujath M

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Board and Battery case damaged.
So can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

[Signature]
8.01.2024

Signature of JC BM (NHM)

Date



MEDICAL OFFICER IN CHARGE
PH CENTRE PARANIYAM
POOVAR P.O., TVM, Pin : 695525

[Signature]
Signature of
Superintendent / Medical Officer (NHM)

MEDICAL OFFICER IN CHARGE
PH CENTRE PARANIYAM
POOVAR P.O., TVM, Pin : 695525



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 115055

438

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>FHC Peraaniam</u>	Call Registration Date : <u>04/01/24</u>
Address <u>Peraaniam</u>	Caller ID : <u>88909</u>
<u>Thuramantapuram, Kerala</u>	Date of Visit : <u>08/01/24</u>
Ph : <u>8113807029</u>	Asset No. : <u>01444115</u>
	EQPT Name : <u>Bp apparatus LED</u>
	Manufacture <u>Elko</u> Model : <u>82520</u>
	S. No. <u>N/A</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
checked at machine and found main board damaged and battery case damaged

Action Taken :
Main board and battery case damaged. cannot be repaired

Completed Date : 08/01/24 Time : 02:00 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>RBER</u>	<u>N/A</u>	<u>N/A</u>	<u>N/A</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Mohamed Sujath M</u>	<u>08/01/24</u>	<u>01:00 pm</u>	<u>02:00 pm</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Mohamed Sujath M</u>	Customer Name : <u>Asaram R-N</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>08/01/24</u>	Date : <u>8/1/24</u>
Contact Number : <u>7293499123</u>	Contact Number : <u>Nursing officer</u>
	Designation : <u>[Signature]</u>
	Hospital Seal : <u>8113807029</u>



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
0144415**



കുടുംബ ആരോഗ്യ കേന്ദ്രം പരണിയം

STOCK REGISTER OF

63

Name of Article BP apparatus (~~Electronics~~) (DIGITAL)

Date	Name and date of Voucher or Invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Batch No	Exp.	Cost	Remarks
15/3/18		from I	A D B						
			2:	1 to	Doctors' room				
				1			3/2022	2276.36/-	
				1	NIL				to Pre check up area)
21-2-21			1	1					Precheck up area -
									- Stock key for instrument general II B - Page 18 of 20 Pharmacy
									See page no: 100
22/3/21			1	1	Nil				Nurses Room. Model - LT 520. from Pharmacist Viji