



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 169002

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>GH Pathanamthitta</u></p> <p>Address <u>Pathanamthitta</u> <u>689645</u></p> <p>Ph : <u>9747211535</u></p>	<p>Call Registration Date : <u>28/2/2024</u></p> <p>Caller ID : <u>98678</u></p> <p>Date of Visit : <u>29/2/24</u></p> <p>Asset No. : <u>0310110</u></p> <p>EQPT Name : <u>UPS</u></p> <p>Manufacture <u>SOCOMEK</u> Model : <u>NRT2U5000C</u></p> <p>S. No. <u>5115240014</u> Dept. <u>Radiology</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : No battery backup.

Action Taken : Checked and found that the inverter Pcb and battery was faulty. Enquired quotation with vendors for spare parts

Completed Date : 29/2/24 Time : 12:00pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	/			
2.	/			
3.	/			

Cyrux Engineer	Date	Start Time	End Time
<u>Shiju John Chacko</u>	<u>29/2/24</u>	<u>10:00am</u>	<u>12:00pm</u>

Customer Remark Completed Pending



Sms

SUPERINTENDENT
GENERAL HOSPITAL
PATHANAMTHITTA


<p>Service Engineer Name : <u>Shiju John Chacko</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>29/2/24</u></p> <p>Contact Number : <u>9605381275</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>29-2-2024</u></p> <p>Contact Number : <u>9712251551</u></p> <p>Designation : <u>[Signature]</u></p> <p>Hospital Seal : <u>[Signature]</u></p>
--	--

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair(BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PATHANAMTHITTA
2	Name of Hospital	GH PATHANAMTHITTA
3	Name of Equipment with Make, Model and Serial Number	UPS FORCR (6KVA) MAKE-SOCOMEK MODEL-NRT2U5000C SINo - 5I15240014
4	Equipment ID&Barcode	98678 & 0310110
	Date of purchase/Year of manufacture /Installation Date	04/01/2016
6	Warranty details(Yes/No)	NO WARRANTY
7	*AMC/CAMC Period agreed at the time Of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/Tollfree)	28/02/2024
9	Actiontaken	Checked and found that inverter Pcb and battery was defective. Quotation enquired with vendor for spare parts.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair(required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Inverter PCB- 14514/- 65AH Exide SMF Battery-16 nos- 108544/- TOTAL COST- 123058/-


13	Asset Value	19000
14	#Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	647 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/CYRIX(Attached or Not)	CYRIX REPORT AND SPARE QUOTATION ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found that inverter Pcb and battery was faulty. The Machine was aged up to 8 yrs. and repairing cost of the machine is 647%, both criteria met. So, machine recommended for BER
17	Name &Signature of CYRIX Authority	Shiju John Chacko 

*Not mandatory #Based on the period of life and value as per the BER guidelines

*Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical)NHM:

Physically verified on 19/3/2024. Recommended for BER as per rule



Signature of JCBM (NHM)

CONSULTANT (BIOMEDICAL)
NATIONAL HEALTH MISSION
PATHANAMTHITTA

Date 15-3-2024





Signature of
Superintendent/Medical Officer(I/c)

SUPERINTENDENT
GENERAL HOSPITAL
PATHANAMTHITTA

YP/VIII/320 AYLARA, KERALA KOLLAM

Date:04/03/2024

Contact : 8921950217

Email : aionbackups12@gmail.com

GSTIN : 32CGUPM9450L1ZN

6KVA SOCOMEK UPS SERVICE PROPOSALTO,
CYRIX HEALTHCARE PVT. LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	SOCOMEK 6KVA INVERTER PCB	1	12300.00	12300.00	18	14,514.00
2	65AH EXIDE SMF BATTERY	16	5300.00	84,800.00	28	108,544.00
PAYABLE AMOUNT						123058.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares, 28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN

21/1/16

Computerized Radiography - model H
Purchase order no. 2013-20018
Invoice no. 2013-20018
Invoice date: 12/1/2013
Invoice to: Building no 5028, Health colony P.O. Pathanamthil
Invoice CS 3583

Purchase order no. 113-5062 dated 12/1/15
Invoice no. KMSCL/EP/TS4/411/2013 compute model H
Radiography - model H. warranty - 3 years
model PR Perfect CS c Drypix plus. Serial no. 56270107
Equipment type - Reader

Fuji computed Radiography system c
Accessories - 1 nos Rs 2,024,787.83/-
Trade discount 31,487.84/-
vat 99,665.01

Total Rs 2,092,965/-

Issued to X-ray unit x

Received Radiography
Remy

P.No: 105

X ray Stock Code

(Signature)
A. V. ...
B. ...

(Signature)

809
SUPERINTENDENT
GENERAL HOSPITAL
PATHANAMTHIL



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800- 425-7669

BAR CODE - 0310110



socomec

Via Sita 13 Zona Industriale Sciozzze
36033 Vidor (Venezia) Italy
Innovative Power Solutions Tel: +39 0444 598611 FAX: +39 0444 598622

MODEL NRT2-U5000C
INPUT 200/208/220/230/240V~ 50/60Hz
29/7/28/5/27/0/25/8/24/7A
OUTPUT 200/208/220/230/240V~ 50/60Hz
25/0/24/0/22/7/21/7/20/8A
5000VA 4500W

NUMBER OF PHASES 1φ
MADE IN CHINA



S/N: 5116240014

REV 08

CM-2

Ambrone & Breda & Lem...

Superior General Health Path...