


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Pathanamthitta
2	Name of Hospital	General Hospital, Pathanamthitta
3	Name of Equipment with Make, Model and Serial Number	Blood Bank Tube Sealer Make : Hemopharm Model : BT200 SN : NA
4	Equipment ID & Barcode	88390 & 0310189
5	Date of purchase/Year of manufacture /Installation Date	31/03/1998
6	Warranty details(Yes/No)	NA
7	*AMC/CAMC Period agreed at the time Of purchase	NA
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	02/01/2024
9	Action taken	Tried to Service the Main Board
10	Present status of the equipment(Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Needs to Replace the main board
12	Cost of spares(specify parts and cost)	NA


13	Asset Value	40000/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX
16	Reasons for recommending the equipment as BER	Machine was aged upto 26 years. OEM and Spare parts support are not available. As per the clause 5.3.14.2, This Machine recommended for BER.
17	Name & Signature of CYRIX Authority	Shiju John Chacko 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**

\*Attach Photograph

Remarks and Recommendations of Junior Consultant(Biomedical)NHM:

Physically verified on 27/2/2024. Recommended a BER as per rule.

  
Signature of JCBM (NHM)



Date 22-02-2024

Seal

SUPERINTENDENT  
GENERAL HOSPITAL  
PATHANAMTHITTA



Signature of  
Superintendent/Medical Officer (i/c)



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 185203

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>CCH Pathanamthitta</u>	Call Registration Date : <u>02/01/2024</u>
Address : <u>Pathanamthitta</u>	Caller ID : <u>88390</u>
<u>689645</u>	Date of Visit : <u>22/02/2024</u>
Ph : <u>7907816877</u>	Asset No. : <u>0310189</u>
	EQPT Name : <u>Blood Bank Tube Sealer</u>
	Manufacture <u>HEMOPHARM</u> Model : <u>BT-200</u>
	S. No. <u>NA</u> Dept. <u>Blood Bank</u>

**Service Classification : Breakdown Call**  **PMS**  **Calibration**  **Cust.Training**

**Problem Identified :** Not working.

**Action Taken :** Checked the machine. Found that machine main board fully damaged. not serviceable

This machine recommended for BER.  
Completed  Date : 22/2/24 Time : 12:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

**Cyrux Engineer**  
Shiju John Chacko


Date	Start Time	End Time
<u>22/2/24</u>	<u>10:00am</u>	<u>12:00pm</u>

**Customer Remark**  Completed  Pending

*(Signature)*  
SUPERINTENDENT  
GENERAL HOSPITAL  
PATHANAMTHITTA

<b>Service Engineer Name</b> : <u>Shiju John Chacko</u>	<b>Customer Name</b> : <u>Aju</u>
<b>Signature</b> : <i>(Signature)</i>	<b>Signature</b> : <i>(Signature)</i>
<b>Date</b> : <u>22/2/24</u>	<b>Date</b> : <u>22/2/24</u>
<b>Contact Number</b> : <u>9605381275</u>	<b>Contact Number</b> : <u>9846696221</u>
	<b>Designation</b> : <u>Blood bank STC</u>
	<b>Hospital Seal</b> :





*Machine Not working.*

*Afrin Bhandari D/o*

*3/10/2017*

SL.NO.	DATE	PARTICULARS OF ASSETS	PARTICULARS OF SUPPLIER	
			NAME AND ADDRESS	BILL NO. AND DATE
34	31/9/2011	cryofuge	Model- 60001	
35	23/7/2011 15/3/14	Air Conditioners - 3 [split] Cover to unservisable Register Page NO:15	SANSINO	One condemned as per the order no P32968 Balance - Nil 02
36	11/05/2012 10/11/2014	Portable Tube Sealer Electre Tube Sealer Invoice No.: 6128/97-98 Humphreys	Model NO: LM100 SLNO 1004769 - BT 200 - expt - mg	Price - 95,000

dated: 31-03-98  
 Supply Order No: SAC/HO/9732/97-98  
 Dts dated 30-03-98  
 Code No: E/62.21400

Recd on 10/11/2014  
 10/11/2014

Agri Blood Bank etc.

22/2/24

*[Handwritten signature]*

SUPERINTENDENT  
 GENERAL HOSPITAL  
 PATHANAMTHITTA





10-11-2014 Tube sealer given to Blood Bank  
GIA, PIA on temporary basis

Saravallu  
10/11/14

NIL

Tube Sealer BT 200 Receipt

Invoice NO. 6128/97-98/EQPT-MB

Penpol dt. 31-3-98 As per

Supply order No. SAC/HCI

92737/97-98/DHS dt. 30-3-98

Code No. F/62.214000

Price Rs 95000