



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 183507

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>DH. MAVELIKKARA</u></p> <p>Address <u>MAVELIKKARA</u> <u>ALAPPUZHA (D)</u></p> <p>Ph : <u>9447961500</u></p>	<p>Call Registration Date : <u>04.02.2024</u></p> <p>Caller ID : <u>94279</u></p> <p>Date of Visit : <u>05.02.2024</u></p> <p>Asset No. : <u>0411127</u></p> <p>EQPT Name : <u>NEBULISER</u></p> <p>Manufacture <u>MEDITECH</u> Model : <u>HANDY MEG</u></p> <p>S. No. <u>NGC-119184</u> Dept. <u>MEDICAL WARD</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT WORKING

Action Taken : CHECKED THE MACHINE AND FOUND THAT THE NEBULISER MOTOR AND PISTON ARE DAMAGED. CONDITION AILED TO REPAIR. MOTOR UNIT. ENQUIRED SPARES.

Completed Date : 03.24 Time : 12:00 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>MOTOR UNIT (NEBULISER)</u>	<u>1</u>	<u>NA</u>	<u>NA</u>
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SAUBAN BIN YASEEN</u>	<u>2.3.24</u>	<u>10:00am</u>	<u>12:00 pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>SAUBAN</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>02.03.2024</u></p> <p>Contact Number : <u>9745809490</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>23/03/2024</u></p> <p>Contact Number : <u>956252416</u></p> <p>Designation : <u>N/O</u></p> <p>Hospital Seal :</p>
--	---




Mavel-KA
Superintendent
District Hospital
Mavelikkara

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	DH MAVELIKKARA
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER MAKE :MEDTECH MODEL :- HANDYNEB SMART SN :- NGC-119184
4	Equipment ID & Barcode	94279 & 0411127
5	Date of purchase / Year of manufacture /Installation Date	09/03/2019
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	04/02/2024
9	Action taken	Checked the machine and found that the Nebulizer motor and piston are damaged condition. Enquired spares in the market.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs:- 1353 /-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that the nebulizer motor and piston are defective, enquired for spares in the market but spare are not available. Equipment is installed on 09/03/2019 and aged up to 4 years and 11 months. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	SAUBAN BIN YASEEN 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


JYOTHIBA THOMAS 12/04/20
 Consultant (Biomedical Engineering)
 NHM ALAPPUZHA

Signature of JC BM (NHM)

Date




Superintendent
District Hospital
Mavelikkara
 Superintendent / Medical Officer (i/c)


NEBULIZER

3	18.03.09	NEBULIZER CBT from old Register Pg. No. 104 Recd from the store 22.8.2009 amt 18.03.2009		3.
1	14/6/12	Recd from the store Received only	0.24	04 2 HWA P/B
1	22/5/12	Received from post-operative ward Received from store	1 1	5 6
5		Conditioned & Analyzed as per Conclusion: B1- 179/2013 Supt 17/1/13	2	4
4				
3	6/7/16	Received from store (As per order) (Store page no - 247)	1	5
4	1.11.16	Conditioned as per order NO D4-18785/16/G.M.A(A) dt 18.11.2016	1	4
3				
2	11/1/17	As per order no B1-1768/17/1/17 order no D4-9640/17/1/17	1	3
4	3/1/19	Recd from store (model no: 102) ^{ppc} (253) Rate A 21590192090 1800 105 105 <u>2010</u>	1	7 3
2	9/3/19	Recd from store (model no: 102) (253) (med sub) ready stock available	1	4
2	25/3/2020	Recd from store worth 50500 (checked) Rs 1151 Recd -	1	5
	10/12/17	To cut for use from Reg. II		12 ✓
	10/1/21	To cut for use from Reg. II		13 ✓



Final for
Superintendent
District Hospital
Mavelikkara

(C.O page no 149)



DISTRIB...
PUNJ...
MAY 18 2016

Shree. Ka
Superintendent
District Hospital
Mavelikkara