

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Alappuzha
2	Name of Hospital	Taluk Hospital Harippad
3	Name of Equipment with Make, Model and Serial Number	Pulse oximeter skanray cornet plus
4	Equipment ID/ Barcode	0420119
5	Date of purchase/ Year of manufacture	2013
6	Warranty details	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	24/09/2023
9	Action taken	checked and found it is not possible to repair.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair. It is in unrepa- reable condition.
12	Cost of spares (specify parts and cost)	Na

13	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
14	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Attached
15	Reasons for recommending the equipment as BER	on checking found that display, battery and spce probe is defective. This product is out of production and it is obsolete by OEM so spare for this product is not available, hence it is suggesting for consideration.
16	Signature of CYRIX Authority	M. Sub. C. Sub. C.

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

*** Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Recommended for BER


 28/10/23
 JYOTHISH THOMAS
 Consultant (Biomedical Engineering)
 TALUK HEAD QUARTERS HOSPITAL ALAPPUZHA

Signature of JC BM (NHM)

Date




 Signature of
 Superintendent/Medical Officer (i/c)

Superintendent
 Taluk Head Quarters Hospital
 Haripad- 690 514



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 140862

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>THQH, Harippad</i>	Call Registration Date : <i>24/09/23</i>
Address <i>Harippad</i>	Caller ID : <i>71145</i>
<i>Alappuzha</i>	Date of Visit : <i>27/9/23</i>
Ph : <i>9745301546</i>	Asset No. : <i>0480179</i>
	EQPT Name : <i>Pulse Oximeter</i>
	Manufacture <i>Skandray</i> Model : <i>Comet Plus</i>
	S. No. <i>Na</i> Dept. <i>Paediatric ward</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Not working*

Action Taken : *On checking found that display is defective, Battery is defective and SPO2 probe is defective. This product is out of production and it is obsolete by the manufacture. So spare for this product is not available. Hence it is suggesting for condemnation.*

Completed Date : *27/9/23* Time : *3:00* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Maidula G</i>	<i>27/9/23</i>	<i>18:00</i>	<i>3:00</i>

Customer Remark Completed Pending

Service Engineer Name : *Maidula G*
Signature : *[Signature]*
Date : *27/9/23*
Contact Number : *9061199471*

Customer Name : *[Signature]*
Signature : *[Signature]*
Date : *[Date]*
Contact Number : *[Number]*
Designation : *Superintendent*
Hospital Seal : *Taluk Head Quarters Hospital Harippad - 680 514*

Dear Customer,

Date: 24.01.2023

Subject: End of service support for Comet model

It is to inform you that the above said product is out of production and it is obsoleted.
There is no service support available for it, hence it is fit for condemnation.

Thanking you.

For Skanray Technologies Limited



Avinash H M.

Avinash H M

Service Operations Co-ordinator


pulse oxymetre

		Receipt
10/12	Received from store (TCU) Tr. in from TCU page No: 27 (Stock Register) (BF from volume 11, page no. 55)	1 7
2-13	Rel from store (COMET-Plus) (cost Rs 46200/- S.No. C13 HAB315) Warranty - 2 yr. 16) (Page no. volume. 14)	1 7

CYRIX MEDICAL EQUIPMENT MAINTENANCE PROGRAMME (MEP)

TOLL FREE NO: 1800 - 425-7868

BAR CODE - (8004-000018228)
0428779



Cancel



Handwritten text: *Cham*