


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	K.R. Narayanan Memorial Super Speciality Hospital Ozhavoor.
3	Name of Equipment with Make, Model and Serial Number	Nebulizer Make: Lifecare Model: EN1369009 Sn: NA.
4	Equipment ID & Barcode	0511991
5	Date of purchase / Year of manufacture /Installation Date	19-08-2021
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20-08-2024, #97161
9	Action taken	Serviced.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Motor coil damaged. Rewinding not possible. Required new motor unit with all parts.
12	Cost of spares (specify parts and cost)	Equal to Asset Value

13	Asset Value	NA
14	"Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Equal to Asset Value
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix
16	Reasons for recommending the equipment as BER	Motor is main part of nebulizer. Spares not available. Machine not serviceable without spares. Also machine aged upto 3 years.
17	Name & Signature of CYRIX Authority	Nidhish Roy 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - not available.
 Installation date as per stockbook 17/2/2021
 cost of spares equal to Asset value
 Life period = 3 years
 Cyrix reported that spares not available.


 Signature of JC BM (NHM)


KOTTAYAM

Machin not working

Rems:
 Rems. M'V
 S No.

Date




 Signature of Superintendent / Medical Officer (i/c)

SUPERINTENDENT

K. B. Narayana Memorial Hospital



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 166322

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>K.R. Narayanan</u>	Call Registration Date : <u>20-02-2024</u>
Address <u>Memorial Hospital</u>	Caller ID : <u>97161</u>
<u>Uzhavoor, Kattayam</u>	Date of Visit : <u>21-02-2024</u>
Ph : <u>9526752336</u>	Asset No. : <u>0511991</u>
	EQPT Name : <u>Nebulizer</u>
	Manufacture <u>Lifecare</u> Model : <u>CNB69009</u>
	S. No. <u>N/A</u> Dept. <u>Female ward</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Motor Complaint

Action Taken : Checked. Scaviced. Motor coil break and piston needs service. Required new motor with all parts for solving issues. Spare parts not available. So that machine unserviceable condition.

Completed Date : Time : Spare Required




Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nidhish Roy</u>	<u>21-02-2024</u>	<u>11:10am</u>	<u>11:50am</u>

Customer Remark Machine not working Rem. m.v Remf 370. Completed Pending

Service Engineer Name : <u>Nidhish Roy</u>	Customer Name : <u>[Signature]</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>21-02-2024</u>	Date : <u>[Signature]</u>
Contact Number : <u>9072575296</u>	Contact Number : <u>K.R. Narayanan Memorial</u>
	Designation : <u>Speciality Hospital</u>
	Hospital Seal : <u>Uzhavoor</u>



NEBULISER

<p>12/02/2021 Covid Supply through DDMS</p>	<p>R/f Kmscl on No - KL00529 / 02039/20-21 dtd 17/02/2021 (1) ^{UK 100} DDMS Entry - KL00529 / 100094/20-21 dtd 17/02/2021 Nebuliser Batch 115 KL00529 ^{supplied} 54P00407 ITEM CODE E00021 Issued all items to Head Nurse Head Nurse Through DDMS DDMS for Covid ward DDMS KL00529 / TR 100428/20-21 ?</p>	<p>7</p>	<p>7</p>	<p>NIL</p>
<p>12/10/2021</p>	<p>Returned from ward (H/N) for used to transferred to THQ+1 Ruavi langad - KL00529/TR/00039/21-22</p>	<p>3</p>	<p>3</p>	<p>NIL</p>