


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kottayam
2	Name of Hospital	PHC Parampuzha, Kottayam
3	Name of Equipment with Make, Model and Serial Number	Eg: BP Apparatus - (Digital) Make: Inji Model: Mdi Sl no: 18060 2683
4	Equipment ID & Barcode	0543918
5	Date of purchase / Year of manufacture Installation Date	19-2-2019
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	07/02/2024 Id: 94766
9	Action taken	checked card fund equipment is showing error card showing wrong values.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Need to replace the main board card motor.
12	Cost of spares (specify parts and cost)	Equal to Asset Value.

13	Asset Value	1500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Equal to Asset Value
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Leyrix report attached.
16	Reasons for recommending the equipment as BER	Need to replace the main board and motor. But quotation not getting from OEM/ supplier. Hence recommending the equipment as RBER
17	Name & Signature of CYRIX Authority	 Sanish Swain

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM

Asset value - 1500
 cost of spare - equal to Asset value
 Life period - 5 years. (2019 to 2024)
 The machine is around 5 years old

CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH MISSION

 27/2/24
 Signature of JC BM (NHM)


 Superintendent / Medical Officer (i/c)

Signature of
 Seal

Date

2024-02-27 10:00:00
 2024-02-27 10:00:00



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 172556

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC</u></p> <p>Address : <u>PHC Palsampuzha</u> <u>Kottayam</u></p> <p>Ph : <u>9496720871</u></p>	<p>Call Registration Date : <u>07/2/2024</u></p> <p>Caller ID : <u>94766</u></p> <p>Date of Visit : <u>07/2/2024</u></p> <p>Asset No. : <u>0543918</u></p> <p>EQPT Name : <u>B.P Apparatus, Digital</u></p> <p>Manufacture <u>Inf.</u> Model : <u>mdi</u></p> <p>S. No. <u>180602683</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Showing error and Wrong Values

Action Taken : checked and found equipment showing Wrong Values. Repairing of the equipment is not possible and is not feasible.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sanus's Swan</u>	<u>07/2/24</u>	<u>3:00pm</u>	<u>3:30pm</u>

Customer Remark Completed Pending

Not working properly Bmp

Service Engineer Name : Sanus's Swan

Signature : [Signature]

Date : 7/2/24

Contact Number : 9946646571

Customer Name : [Signature]

Signature : [Signature]

Date : [Signature]

Contact Number : [Signature]

Designation : [Signature]

Hospital Seal : [Signature]

Digital BP monitor

0.37

Date		Received	Issued	Balance	Remarks
12/2/19	Stock received from the store Li' Reg. Page no. 12.	2		2	
12/2/2020	Stock received from the store ...: Flood registers Page 14 (Price/unit Rate Rs 1119/19 Total Rs 5595/95)	5		7	

