


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC MALAPPATTAM
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE Make : SAMSO Model: BRIGHT SN:N/A
4	Equipment ID/ Barcode	#104590-1341991
5	Date of purchase/ Year of manufacture/Installation Date	26-08-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 03-04-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found main board, Load cells damaged. Need to replace the these spares for further checking .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	715/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 26-08-2019 aged up to 4 years. Quotation not submitted since spares are not available from the market, So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that the machine not switching on. r/s main board & load cells are damaged. Aged up to 4 years. So recommended for BER as per BER Protocol.



28/04/2024
Signature of JC BM (NHM)
NATIONAL HEALTH MISSION
KANNUR - 670002

MINIMOL JOSIE


RBSK NURSE

Date: 24/4/2024




Signature of
Superintendent Medical Officer (S/O)
Primary Health Centre
Kannur



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 185640

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC MALAPPATTAM</u></p> <p>Address <u>MALAPPATTAM</u> <u>KANNUR</u></p> <p>Ph : <u>7025528483</u></p>	<p>Call Registration Date : <u>3/4/24</u></p> <p>Caller ID : <u>104590</u></p> <p>Date of Visit : <u>4/4/24</u></p> <p>Asset No. : <u>1341991</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>sanso</u> Model : <u>bright</u></p> <p>S. No. <u>na</u> Dept. <u>main centre</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine is not switching ON.

Action Taken : checked the machine and found that machine is not switching ON, found main board, load cells damaged. Need to replace the these spares for further checking.

Completed Date : 4/4/24 Time : 12:00 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin</u>	<u>4/4/24</u>	<u>11:00 am</u>	<u>12:00 pm</u>

Customer Remark Completed Pending



<p>Service Engineer Name : <u>Aswin</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>4/4/24</u></p> <p>Contact Number : <u>8301012759</u></p>	<p>Customer Name : <u>MINIMOL JOSE</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>4/4/24</u></p> <p>Contact Number : <u>9400106110</u></p> <p>Designation : <u>RBSK nurse</u></p> <p>Hospital Seal :</p>
---	--

MEDICAL OFFICE
Primary Health Centre
Malappattam, Kannur



SAMSUNG



Malabar, Kozhikode
Private Health Centre (PHC)
21/10/2018

PHC
KANNUR DIST
MALAYALAM
Date:

KEEP
TOLL FREE 800-1000 - 654-7890
10/1/2004
Unit No. 101-20

TOLL FREE 800-1000 - 654-7890
Unit No. 101-20

CYRIX
10/1/2004
Unit No. 101-20

CE
OC
ELECTRONIC COMPONENTS

MARY HEALTH CENTRE
MALAPPATTAM
Date:
KANNUR DT. *

2/12
MEDICAL OFFICES
Primary Health Centre
Malappattam, Kannur