


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	UPHC MUTHATHI
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make : INFI Model:BSX515 SN:20200402309
4	Equipment ID/ Barcode	#106387 -1344703
5	Date of purchase/ Year of manufacture/Installation Date	03-11-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 15.04.2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found main board, Pressure sensor and pump defective. Need to replace the these spares for further checking .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found main board, Pressure sensor and pump are defective. Equipment installed on 03.11.2020 aged up to 3 years and 5 months. Quotation not submitted since spares are not available from the market, So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that machine not switching on. found main board, pressure sensor and pump are defective. Aged up to 3 years & above. So recommended for BER as per BER Protocol.


23/09/2024

Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION

KANNUR - 670002

Roshni. T.T



Nursing officer.

Date: 27/04/24





Signature of MO UPHC
Superintendent/Medical Officer (i/c)

Dr. Ankeeta Bham
MUTHARU

MEDICAL OFFICER

UPHC Primary Health Centre
National Health Mission
Kannur Municipality



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
202552

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility UPHC Muthathi.....</p> <p>Address Kannur, Kerala.....</p> <p>Ph : 8848668115.....</p>	<p>Call Registration Date : .. 15-04-2024</p> <p>Caller ID : .. 106387</p> <p>Date of Visit : .. 16-04-2024</p> <p>Asset No. : .. 1344703</p> <p>EQPT Name : .. BP apparatus</p> <p>Manufacture .. Inf Model : .. BSX.515</p> <p>S. No. 20200402309 Dept. OP</p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Found machine is not switching ON.....

Action Taken : Checked the machine and found that machine is not switching ON, found mainboard, pressure sensor and pump defective. Need to replace these spares for further checking.....

Completed Date : .. 16-04-2024 Time : .. 11:00AM .. Spare Required

Spare Replaced Requested


No.	Description	Qty.	Part Number	PR Number
1.	NA	NA	NA	NA
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
Nitheesh P	16-04-2024	10:00AM	11:00AM

Customer Remark Completed Pending

Service Engineer Name : Nitheesh P
Signature :
Date : 16-04-2024
Contact Number : 8301871531

Customer Name : Rosmo T T
Signature :
Date : 16-04-24
Contact Number : 8848668115
Designation : Nursing officer
Hospital Seal :



Handwritten notes and signatures in the bottom right corner.


CYRIX
HEALTH CARE PVT LTD

**BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)**



TOLL FREE NO: 1800 - 425-7669

**BAR CODE - (8004-890615225)
1344703**



Dr. Anthe Bhanu
Dr. Anthe Bhanu
MEDICAL OFFICER
Urban Primary Health Centre
National Health Mission
Mangaluru Municipality