


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	WAYANAD
2	Name of Hospital	THQH VYTHIRI
3	Name of Equipment with Make, Model and Serial Number	Equipment-Centrifuge Make-Rotek Model-M8E Sn-0825
4	Equipment ID/ Barcode	1220528//TICKET 95965
5	Date of purchase/ Year of manufacture/Installation Date	15/05/2008
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	14/02/2024/Toll free
9	Action taken	Checked, found motor and display board full set defective Enquired spare from OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Centrifuge motor-Rs 5310/- Display board full set-Rs 5900/- Total -Rs 11210/-

13	Asset Value	RS 9651/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	116%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Quotation from Rotek (OEM) Cyrix service report
16	Reasons for recommending the equipment as BER	Checked, found motor and display full board set defective Equipment completed 15 years Repairing cost of equipment 116 % with respect to asset value Both criteria of RBER Met So we recommended for condemnation as per tender clause 5.3.14.1
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 

Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Centrifuge Display board, Complaint: 15yrs already completed. Recommended Condemnation.



28/3/24.

Signature of JC BM (NHM)
CONSULTANT
Bio-Medical Engineer
National Health Mission
(Arogyakeralam)
Wayanad-673122

Mechin is unrepairable



Date:



Signature of
Superintendent/Medical Officer

Taluk Head Quarters Hospital
Vythiri, Wayanad.



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 181813

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THKOH.</u></p> <p>Address <u>VYTHIRI.</u></p> <p>..... <u>WAYANAD, KERALA.</u></p> <p>Ph : <u>9747621775</u></p>	<p>Call Registration Date : <u>14/2/2024.</u></p> <p>Caller ID : <u>95965</u></p> <p>Date of Visit : <u>16/2/2024.</u></p> <p>Asset No. : <u>1220528</u></p> <p>EQPT Name : <u>CENTRIFUGE.</u></p> <p>Manufacture <u>ROTEK</u> Model : <u>M8E</u></p> <p>S. No. <u>0825</u> Dept. <u>ICU LAB.</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not rotating.

Action Taken : Inspected the equipment found that the motor and display board malfunction. Need to replace motor and display board.

Completed Date : 16/2/2024 Time : 12:00PM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>DISPLAY BOARD.</u>	1		
2. <u>MOTOR.</u>	1.		
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Dijulim ER05</u>	<u>16/2/2024</u>	<u>10:00AM</u>	<u>12:00PM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Dijulim</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/2/2024</u></p> <p>Contact Number : <u>8086753545</u></p>	<p>Customer Name : <u>Taluk Head Quarters Hospital</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/2/2024</u></p> <p>Contact Number : <u>9747621775</u></p> <p>Designation : <u>Lab Technician</u></p> <p>Hospital Seal :</p>
--	---



**B&C INDUSTRIES, West Vengola P.O, Perumbavoor ,
Ernakulam , Kerala - 683556 Mail ID – mail@rotekinstruments.com**

Ph: 9446139563, 8593077000, 8593000606

GST : 32AAEFB4140N1ZO

QTN -B&C- 418 /23 - 24

29/02/2024

To

Cynix Healthcare Pvt Ltd

Ernakulam

Sub: Quotation for Centrifuge Spares

Sir

We are sending the rate of centrifuge spares. The rate is given below for your kind consideration.

Centrifuge at THQH - Vythiri

Centrifuge (MSE) Motor - Rs 4500 -00

Display board full set - Rs 5000-00

9500 -00

Add GST @ 18%

1710-00

11210-00
=====

Thanking You

Yours Faithfully

For B&C Industries




Superintendent
Talar Head Quarters Hospital
Vythiri, Wayanad.



COST OF ASSET (TAX+LANDING CHARGES)	IDENTIFICATION NUMBER	LOCATION OF THE ASSET	REMARKS
4,800.00	KSACS /ICTC /04	Lab technician's Room	From Hospital Store Local purchase by KSACS Fund
1,926.66	KSACS /ICTC /05	Lab technician's Room	From Hospital Store (KSACS)
1,686.00	KSACS /ICTC /06	Lab technician's Room	From Hospital Store (KSACS)

SL.NO.	DATE	PARTICULARS OF ASSETS	NAME AND ADDRESS	PARTICULARS OF SUPPLIER	BILL NO. AND DATE
04	15/05/08	Micro pipette	Habbay Diagnostics Kozhikode		Bill No. 15238 15/05/08
05	15/05/08	Needle Destroyer	Habbay Diagnostics Kozhikode		Bill No. 15238 15/05/08
06	15/05/08	Centrifuge	Habbay Diagnostics Kozhikode		Bill No. 15238 15/05/08



[Handwritten signature]

**Superintendent
Tahir Read Quarters Hospital
Wayanad, Wayanad.**



Superintendent
Head Quarters Hospital
Vythiri, Wayanad.



Quotation

3 messages

Rotek Instruments <mail@rotekinstruments.com>
To: niyascyrix@gmail.com <niyascyrix@gmail.com>

Thu, 29 Feb 2024 at 12:45 pm

Sir
PFA

Thanks & Regards



ROTEK INSTRUMENTS

B&C Industries, West Vengola P.O., Perumbavoor- 683556, Ernakulum Dist., Kerala,

PH: 8593070000/8593000606 / 9446139563 web: www.rotekinstruments.com



Niyas Hadi AP <niyascyrix@gmail.com>
To: Rotek Instruments <mail@rotekinstruments.com>
Cc: Sarang KM <sarangcyrix@gmail.com>

Tue, 5 Mar 2024 at 9:33 am

Dear Sir
Thanks for giving quotation
Kindly please quote with your sign and seal

Thanks & regards

NIYAS HADI
District incharge
BEMP project
Wayanad
kerala
mob -7593847171

email -niyascyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

[Quoted text hidden]

Rotek Instruments <mail@rotekinstruments.com>
To: Niyas Hadi AP <niyascyrix@gmail.com>

Tue, 5 Mar 2024 at 11:25 am

Thanks & Regards



ROTEK INSTRUMENTS

B&C Industries, West Vengola P.O.,Perumbavoor- 683556, Ernakulum Dist., Kerala,

PH: 8593070000/8593000606 / 9446139563 web: www.rotekinstruments.com



[Quoted text hidden]