


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KOZHIKODE
2	Name of Hospital	THQH THAMARASSERY
3	Name of Equipment with Make, Model and Serial Number	Equipment : Digital Weighing Machine Pediatric Make: Pragathi Medical System Model: 2034 SN: NA
4	Equipment ID & Barcode	95979 & 1120220
5	Date of purchase / Year of manufacture /Installation Date	12-03-2016
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-02-2024
9	Action taken	Checked and found load cell, main board, display, machine body and transformer defective. Enquired the spare with vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Total spare cost =4484/-



13	Asset Value	6725 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.6%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found load cell, main board, display, machine body and transformer defective. Equipment installed on 12-03-2016 and aged up to 7 years 11 months. Repairing cost of the equipment is 66.6 % both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	PRABHATH K 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked and found that load cell, main board, display, machine body and transformer are defective. Equipment installed on 12/3/16. So recommended for condemnation.

Signature of Consultant Bio-Medical (NHM)  
National Health Mission  
Kozhikode 22-02-2024



  
DR: BHATHIDA  
JUNIOR CONSULTANT  
GOVT. TALUK HOSPITAL  
THAMARASSERY - 673573

Date

Seal

Signature of  
Superintendent / Medical Officer (i/c)





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 180998

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p><b>Health Facility</b> T.H. Thamarassery</p> <p><b>Address</b> Thamarassery Kozhikode</p> <p><b>Ph :</b> 7736240442</p>	<p><b>Call Registration Date :</b> 14-02-2024</p> <p><b>Caller ID :</b> 95979</p> <p><b>Date of Visit :</b> 15-02-2024</p> <p><b>Asset No. :</b> 1120220</p> <p><b>EQPT Name :</b> Digital Weighing Machine</p> <p><b>Manufacture :</b> Pragathi Medical Model : 2039</p> <p><b>S. No. :</b> NA <b>Dept. :</b> NLCU</p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Machine not switching ON.

**Action Taken :** Checked and found bad cell, main board and transformer are defective, body completely rusted. Need to replace those spares for further checking working condition of the equipment. Enquired OEM for spares.

**Completed**  **Date :** 15/2/24 **Time :** 11:15 AM **Spare Required**

**Spare Replaced**  **Requested**

No.	Description	Qty.	Part Number	PR Number
1.	←—————			
2.			←—————	
3.				

Cyxix Engineer	Date	Start Time	End Time
Athul Raju	15/2/24	11:00 AM	11:15 AM

**Customer Remark**  Completed  Pending

Installed on 12/12/2016, due to prolonged use, machine rusted and damaged, not repairable.

Dr. S. SURESH P. POOY

<p><b>Service Engineer Name :</b> Athul Raju</p> <p><b>Signature :</b> [Signature]</p> <p><b>Date :</b> 15/2/24</p> <p><b>Contact Number :</b> 9022522054</p>	<p><b>Customer Name :</b> Dr. S. SURESH P. POOY</p> <p><b>Signature :</b> [Signature]</p> <p><b>Date :</b></p> <p><b>Contact Number :</b></p> <p><b>Designation :</b> MEDICAL OFFICER</p> <p><b>Hospital Seal :</b> NATIONAL HEALTH MISSION Taluk Hospital, Thamarassery</p>
---	--





**Required Quotation - 95979-1120220-Digital Weighing Machine-Thqh Thamarassery, Kozhikode**

3 messages

Tue, Mar 5, 2024 at 7:42 AM

Jineesh P P <jineeshcyrix@gmail.com>  
To: Sales@pragathineo-natal.com  
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir

Please share price and spare availability of load cell,main board , display,transformer and rusted top body part for digital weighing machine

Make : Pragathi Medical Systems  
Model: 2034  
Sn:NA

Attached photos for your reference.

Thanks & Regards,

Jineesh PP  
Sr.BME  
Mob-7593847163  
jineeshcyrix@gmail.com  
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

Jineesh P P <jineeshcyrix@gmail.com>  
To: Sales@pragathineo-natal.com  
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Wed, Mar 6, 2024 at 11:25 AM

Dear Sir,

Please update the status,  
[Quoted text hidden]

Sales Pragathi <sales@pragathineo-natal.com>  
To: Jineesh P P <jineeshcyrix@gmail.com>  
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Wed, Mar 6, 2024 at 5:02 PM

Baby Weighing Scale  
Rs. 3,800 + 18% tax = 4,484

transport  
200 + 18% tax = 236

Total = 4,720 /-

Bank: CANARA BANK. Branch: Tumkur Road  
A/C NO : 0882261000480 IFSC CODE : CNRB0000882  
[Quoted text hidden]

Thanks and regards,

Panchakshari B.





**Pragathi Medical System.**

49/1, Beside Ullas Theatre,

Yeshwantpur Industrial Suburb,

Bangalore-560022.

Ph: 9844111792, 9449764764

[www.pragathineo-natal.com](http://www.pragathineo-natal.com)











*[Signature]*  
SNO INCHARGE LR

*[Signature]*  
Dr. SUHADA  
NHM MO.

MEDICAL OFFICER  
NATIONAL HEALTH MISSION  
Taluk Hospital, Thamarassery

